

Handheld Pulse Oximeter

VM-2160

Service manual



 VIAMED	Service manual	Rev. 3
	VM-2160	Page 2 of 15

VM-2160 - Service Manual

REV: 3

Released: 11/2017

The information in this document is subject to change without notice.

Copyright © 2008 Viamed Ltd. All Rights Reserved.

This document contains proprietary information that is protected by copyright. All rights are reserved. No part of this document may be photocopied, reproduced or translated to another language without prior written consent of Viamed Ltd.

Contact information

Viamed Ltd
15 Station Road
Cross Hills
Keighley
West Yorkshire
BD20 7DT
United Kingdom

Phone: +44 (0)1535 634542
Fax: +44 (0)1535 635582
E-mail: info@viamed.co.uk
www.viamed.co.uk

Table of Contents

1	Introduction	4
2	Service menu	4
2.1	Service PIN	4
2.2	Service menu structure	4
3	Maintenance	6
4	Service and repairs	6
4.1	General	6
4.1.1	Device disassembly	6
4.1.2	Device assembly	6
4.1.3	VM-2160 components and HW compatibility	7
4.2	Replace Components	8
4.2.1	Main-PCB	8
4.2.2	USB-PCB	8
4.2.3	USB-PCB cable	9
4.2.4	Flat ribbon cable	10
4.2.5	SpO ₂ Module	10
4.2.6	SpO ₂ sensor socket	11
4.2.7	Keypad	12
4.2.8	Display window	12
4.2.9	Housing	13
5	Troubleshooting Guide	14
6	Spare parts	15

 VIAMED	Service manual	Rev. 3
	VM-2160	Page 4 of 15

1 Introduction

This service manual contains information for servicing the VM-2160 - handheld pulse oximeter. Only authorized service personnel may service this product.

The Handheld Pulse Oximeter VM-2160 - User Manual (the operator's manual) is an integral part of the service procedures. Before servicing the monitor, read the operator's manual carefully for a thorough understanding of how to operate the unit.

Only use the tools and test equipment as specified in this manual. Only use original spare parts available from your local distributor.



NOTE: This manual is for VM-2160 monitor. A separate manual is available for the VM-2160 monitor with high performance SMARTsat® technology.



Warning:

Incorrect procedures may harm the patient, or damage the monitor.

2 Service menu

The Service menu is accessible to authorized personnel only. The PIN Code specified below shall not be given to any third parties. Changes of settings are only allowed with profound knowledge about each setting.



Warning:

Wrong settings may lead to misinterpretation and may affect patient safety!

2.1 Service PIN

The PIN Code for VM-2160 can be accessed by menu item "Service".

By entering the Service PIN CODE you confirm that you have read and understood the service manual and take full responsibility for the consequences that might result from the changed alarm functionality.

PIN CODE : 1 9 8 5

2.2 Service menu structure

Within the service menu point START UP SETTINGS selective default start up parameters can be adjusted or the default start up parameters can be reset to the factory defaults.

The service menu point SYSTEM INFORMATION is for factory internal use only; there are no adjustable parameters in this menu.



Warning:

Changing the default start up settings of the device can seriously affect the functionality of the device and the alarms.

⚠ Warning:

Only authorized personnel with the required technical and medical knowledge are allowed to change the default start up settings.

Start-up setting:

Parameter	Possible settings	Factory default	Description
Volume	1/ 2/ 3/ 4	2	Set default loudness. 1 = off
Alarm SpO2 ↑	100 to Alarm SpO2 ↓+1/ Off	100	Set upper alarm limit SpO ₂
Alarm SpO2 ↓	45 to Alarm SpO2 ↑-1/ Off	85	Set lower alarm limit SpO ₂
Alarm Pulse ↑	300 to Alarm Pulse ↓+1/ Off	140	Set upper alarm limit pulse rate
Alarm Pulse ↓	20 to Alarm Pulse ↑-1/ Off	45	Set lower alarm limit pulse rate
High Alarms	On/Off	On	Turn alarms of high priority on/ off
Medium Alarms	On/Off	On	Turn alarms of medium priority on/ off
Low Alarms	On/Off	On	Turn alarms of low priority on/ off
SpO ₂ Averaging	Stable/Standard/Sensitive	Standard	SpO ₂ -Averaging
Memory full	On/Off	On	Enables display of „memory-full“ message each start-up, if the device memory is full
Auto-off	On/Off	On	Turns device off after 2 minutes, if no measuring takes place
Factory settings	No/Yes	No	Reset device to factory settings
Exit			Exits menu

System Information:

Parameter	Displayed status (read only)	Description
System failure code	0x00	System error code displayed as Hexadecimal value
Led damaged num	0x00	Number of LED errors displayed as Hexadecimal value
Work time	0h 0min	Display of total working hours
Flash error num	0	Total number of Flash error
Module version	SWxxxx Vx.xx.xxxx	Version number of SpO ₂ Module
Module SN	xxxxWMxxxx	Serial number of SpO ₂ Module
LOT number	yyyy.mm	Year.Month of production
Exit		Exits menu

3 Maintenance

The VM-2160 monitor is permanently factory calibrated. The maintenance and calibration-free SpO₂ measurement technology ensures a robust measurement function throughout the lifetime of the monitor. However, to meet any local or specific customer requirements, a basic maintenance plan every 2 years is nevertheless highly recommended.

Recommended test steps are described in detail in the separate document *VM-2160 Maintenance*. All test steps involved in the maintenance check are to be documented in the separate document *VM-2160 Maintenance Checklist*. Both documents are available at the manufacturer upon request.

4 Service and repairs

4.1 General

Opening of the housing (particularly removal of screw-securing) is only allowed by authorised personnel trained by the manufacturer.

 **Warning:**

After every manipulation inside the device (removal of screw-securing) the complete Service Protocol needs to be successfully completed before clinical use of the device to ensure proper operation. Contact the Viamed service department (info@viamed.co.uk) for the latest Service Protocol Template.

 **Warning:**

Remove batteries before each intervention to avoid damage to the circuits!

4.1.1 Device disassembly

Remove the screw-lock with a screw driver. Remove all 6 screws and carefully remove the bottom housing from the front housing by lifting the backside slightly and shift left with low force to not damage the USB connector. Remove the white flat ribbon cable from the USB-PCB.



4.1.2 Device assembly

Prior to inserting the screws, pay attention that no cables (especially the USB-PCB cable) are pinched by the screw holes. Do not tighten the screws with electric screwdriver but only with hand tools to avoid over winding. After the device is set back up together and bolted, insert the screw-lock in the upper right hole. Push gently inside without damaging the housing.

4.1.3 VM-2160 components and HW compatibility




Table 1: VM-2160 Hardware Compatibility List

OxyTrue®A Main-PCB Version	SpO ₂ connector (Mounted on Main-PCB)	Compatible USB-PCB Version
SMC20110215033(OLED) (discontinued)	6 pin-Wire (REF: 0039017)	USI20080807014 (discontinued)
SMC20131031033 (TFT) (discontinued)		USI20131031014 (discontinued)
SMC20141021033(TFT) (REF: 0039012)	6 pin-Flex (REF: 0039023)	USI20141021014 (REF: 0039013)


4.2 Replace Components

4.2.1 Main-PCB

⚠ Warning:

Before replacing the Main-PCB, check the HW compatibility of the Main-PCB and the USB-PCB in Table 1 (**section 4.1.3**). The use of Main-PCB with incompatible USB-PCB may lead to malfunction and may affect patient safety!

To replace the Main-PCB, open the housing as described in section **4.1.1**. Disconnect all cables from the circuit board. Remove the SpO₂ module and all 6 board screws. The Main-PCB can now be replaced.



Before placing the new PCB, remove protective foil from the display. Remove dust at the display and display window with compressed air (clean and deionized, without oil). Place the new PCB into the upper housing and assure correct fitting.




Mount the PCB back on the housing with 6 screws. Insert the SpO₂ module and connect all cables. Perform a function-test before fixing the housing parts with screws.

4.2.2 USB-PCB

⚠ Warning:

Before replacing the Main-PCB, check the HW compatibility of the Main-PCB and the USB-PCB in Table 1 (**section 4.1.3**). The use of Main-PCB with incompatible USB-PCB may lead to malfunction and may affect patient safety!

Remove the defective USB-PCB by opening the housing as described in section **4.1.1**. Disconnect all cables and remove the battery leads from the bottom housing with a soldering iron. Remove the three USB-PCB screws, take out the PCB and remove the buzzer cables at the USB-PCB with a soldering iron.




Take the new USB-PCB and solder the buzzer cables to the USB-PCB. Place the USB-PCB on the guiding pins at the bottom housing without pinching the buzzer cable. Screw the PCB to the bottom housing and solder the battery leads to the terminals at the back housing. Perform a function-test before fixing the housing parts with screws.

4.2.3 USB-PCB cable

Replace the USB-PCB cable if no USB connection can be established. Open the housing (section **4.1.1**) and disconnect the cable from both plugs, replace the cable with a new cable (REF: 0039022). Both cable ends should be inserted in the plug (see image below).



Pay attention that no cable is clamped inside the screw-holes while mounting the housing back together.



4.2.4 Flat ribbon cable

Replace the Flat ribbon cable if it is mechanically damaged or if the power supply of the device shows malfunction.


Open the housing (section **4.1.1**) and disconnect the cable from both plugs, replace the cable with a new cable (REF: 0039018). Both cable ends should be inserted in the plug (see image below). Fold the cable to the direction of the USB plug while closing the device (not in direction of the SpO₂ module) to prevent clamping.



4.2.5 SpO₂ Module

If the module is defective or not connected correctly, the message “Device defective!” will be displayed. Open the housing (section **4.1.1**) to verify correct fitting or replacement of the module.

First check correct fitting to avoid unnecessary replacements. The SpO₂ module must be connected to the Main-PCB in alignment with the white frame marking on the Main-PCB. All pins must be connected to the socket (no pin may be visible).



For repositioning remove the module and check the soldering of the socket at the module. Replace the module if the socket solder points are damaged/loose. In case of no damage at the socket, reposition the module correctly and perform a functional check. Replace the SpO₂ module if the error persists.

After replacing the module, test the SpO₂ measurement function. Replace the Main-PCB if the malfunction continues (chapter **4.2.1**).

4.2.6 SpO₂ sensor socket

If the device is showing “Sensor fault”, or the measurement is interrupted from time to time, the SpO₂ sensor socket may be faulty.

If a sensor failure can be excluded, the SpO₂ sensor socket has to be replaced. To exclude a sensor failure, perform a measurement with a known working sensor. If the failure persists, replace the SpO₂ sensor socket.

Open the housing (section 4.1.1) and disconnect the flat ribbon cable. The device can be opened completely. Remove the SpO₂ sensor socket and replace it with a new one.



Select the SpO₂ sensor socket depending on the HW version of the main PCB. See Table 1 and 2 (section 4.1.3 and 4.1.4) to order the correct spare part.

	
6 pin-Wire (REF: 0039017)	6 pin-Flex (REF: 0039023)





NOTE: The dark square at the flex of the SpO₂ sensor socket must face the SpO₂ module, else the message “Sensor fault” will continue.

The keyway at the SpO₂ sensor socket has to point to the display side after inserting it into the front housing.

	
The dark square at the flex must face the SpO ₂ module.	The keyway at the SpO ₂ sensor socket has to point to the display side after inserting.

After replacing the SpO₂ sensor socket, test the SpO₂ measurement function by connecting a sensor to the finger. Wiggle the sensor connector in the plug and check that the plethysmogram is not influenced by the movement of the connector. The measurement has to be continuous with no error message.

	
The Plethysmogram may not be influenced by wiggling the plug.	No error messages may occur while wiggling the plug.

4.2.7 Keypad

Open the housing as described in section **4.1.1** and remove the Main-PCB. Be careful not to touch the inside of the display window. Place a scalpel or similar tool carefully under the lower left edge of the keypad. Lift the keypad and remove it carefully from the top housing (with finger or pliers).

Glue remains can be removed with fingers or suitable tool without scratching the front housing. Clean the surface of the housing with 70% isopropanol.

Remove the supporting foil from the new keypad and place it on the lower rim of the front housing. Stick it carefully around the display window and fix it by pressing with the finger.

		
Attach tool and remove the keypad.	Remove the supporting foil	Fix the new keypad to the housing

Remove dust at the display and display window with compressed air (clean and deionized, without oil). Place the Main-PCB into the upper housing and assure correct fitting. Close the device according to section **4.1.2**.




NOTE: It is possible to replace the keypad without opening the device.

4.2.8 Display window

Open the housing as described in section **4.1.1** and remove the Main-PCB.

Place a flat screw driver or similar tool between the display window and the housing to remove the display window. Remove dirt and glue remains from the housing using 70% isopropanol.

Remove the backside protecting foil from the new display window. Do not touch or damage the inner surface of the display. Fingerprints at the inner surface of the display window cannot be removed.

	
Remove the display window with a flat screw driver	New display window



NOTE: The display should now be opaque. If not, remove the middle protection foil carefully without removing the outer silvery foil. Insert new window in an appropriate way. Remove dust at the

display and display window with compressed air (clean and deionized, without oil). Place the Main-PCB into the upper housing and assure correct fitting. Close the device according to section **4.1.2**.


4.2.9 Housing

Bottom housing

A spare serial number label with transparent protection sheet is required during replacement of the back housing. Order the label by contacting the Viamed service department (info@viamed.co.uk) State your device name and serial number.

Replace the bottom housing by open the device and removing the USB-PCB (refer to section **4.2.2**).

Place the label on the housing. Ensure that the buzzer hole is not covered by the label to avoid reduction of alarm sound volume. Close the housing according to section **4.1.2**. Connect a new screw-lock.



Front housing

Replace the front housing by open the device and removing the Main-PCB (refer to section **4.2.1**).

If no window or keypad is fitted, place keypad and window according to section **4.2.7** and **4.2.8**.

Close the housing according to section **4.1.2**.

5 Troubleshooting Guide

ID	Failure mode	Possible causes	Corrective action	see Chapter
#1	No display function (stripes, dark)	display defective, control of display defective	Replace Main-PCB	4.2.1
#2	Keys without function	Main-PCB or buttons on Main-PCB defective	Replace Main-PCB	4.2.1
#3	No USB connection	PC to USB cable defective, PCB-connector-cable faulty, USB-Connector on USB-PCB or Main-PCB defective	Replace cable(s) or Main-PCB	4.2.3 / 4.2.1
#4	Data download not possible despite recognized USB connection	Incompatibility of PC configuration, USB-PCB cable faulty, USB-Connector on USB-PCB or Main-PCB defective	Check PC configuration and USB drivers. Replace cable(s) or Main-PCB	4.2.3 / 4.2.1
#5	Keypad mechanically damaged	Wear of Keypad due to mechanical stress	Replace Keypad	4.2.7
#6	Clock is set to zero after batteries were changed	Buffer battery on Main-PCB flat or Main-PCB defective	Replace buffer battery on main PCB if voltage below 3.1V. Replace Main-PCB	4.2.1
#7	Display window defective (scratched, broken)	Wear of display window due to mechanical stress (fall damage)	Replace display window	4.2.8
#8	Device displays „Sensor fault“	SpO ₂ sensor defective or SpO ₂ sensor socket faulty	Replace SpO ₂ sensor or SpO ₂ sensor socket	4.2.6
#9	Interrupted measurement			
#10	Device displays „Device defective!“	SpO ₂ -Module faulty, wrong/defective positioned SpO ₂ Module or Main-PCB defective	Correct/replace SpO ₂ -Module or replace Main-PCB	4.2.5 / 4.2.1
#11	Strong fluctuation of measured values	- Patient has low perfusion / Motion artefacts - Sensor faulty, SpO ₂ Module or sensor socket defective	- Warm hands of patient, ensure no motion artefacts - Replace sensor, replace SpO ₂ Module	4.2.6 / 4.2.5
#12	ERROR code 01	Refer to ID #10 (SpO ₂ -Module faulty, wrong/defective positioned		
#13	ERROR code 02	Error in flash for stored data	Delete all data. If the error continues, replace Main-PCB	4.2.1
#14	ERROR code 03	Buffer Battery for Real Time Clock flat (should be 3.1V)	Set time in the main menu, replace main PCB if the error continues.	4.2.1

6 Spare parts

All available spare parts are listed in the table below with separate part numbers (REF) for VM-2160.




Fig. 1: Exploded View – VM-2160

For hardware compatibility of parts see section 4.1.3.

Part number (REF)	Description	Identifier (Fig.1)
VM-2160 REF: 0012160		
0039024	Front housing without keypad and display window.	3, 4
0039025	Bottom housing with battery cover and buzzer.	6, 7, 8, 9, 13, 14, 15
0002002	Screw set housing (2 long, 4 short screws)	n/a
0039003	Battery cover	9
0039001	Display window	2
0039000	Keypad	1
0039017	6 pin-Wire SpO ₂ sensor socket (mounted on Main-PCB version SMC20110215033 and SMC20131031033)	11
0039023	6 pin-Flex SpO ₂ sensor socket (mounted on Main-PCB version SMC20141021033)	11
0039018	Flat ribbon cable (power connection at Main-PCB to USB-PCB)	n/a
0039022	USB-PCB cable (USB connection at Main-PCB to USB-PCB)	n/a
0039013	USB-PCB version USI20141021014	12
0039012	Main-PCB version SMC20141021033 (with TFT display)	5
0002006	SpO ₂ module	n/a