

CERTIFICATE OF CALIBRATION

Issued By Britannia Test & Measurement Ltd

Date of Issue 08 March 2019

Certificate Number
BTM1669

Page 1 of 3 Pages



Britannia

Test & Measurement

Britannia Test & Measurement Ltd
3K Brighthouse Brighthouse Village
Brighthouse Road
Middlesbrough
TS2 1RT

Approved Signatory



☐ J. Lillystone ☒ P Trattles

Customer : ViaMed
15 Station Road, Keighley
West Yorkshire BD20 7DT

Date Received :

Instrument -	System ID :	ID2750	
	Description :	Electrical Safety Analyzer	
	Manufacturer :	Fluke	
	Model Number :	ESA615	
	Serial Number :	2433026	Last Certificate Number : STD6503
	Procedure Version :	1.00	Last Calibration Date : 07/03/2018

Environmental Conditions

Temperature : 21°C +/- 1°C
Relative Humidity : 50% +/- 10%

Mains Voltage : 240V +/- 10V
Mains Frequency : 50Hz +/- 1Hz

Comments

Instrument was placed in lab and allowed to stabilise before calibration.

Traceability Information

Instrument description	Serial number	Certificate number	Cal. Date	Cal. Period
3200A Electrical Test Calibrator (STD)	M1384F15	36462	24/07/2018	52
8840A Multimeter	4016001	36467	19/07/2018	52

Calibrated By : P Trattles

Date of Calibration : 08 March 2019

This certificate provides traceability of measurement to recognised National Standards, and to the units of measurement realised at the National Physical Laboratory or other recognised National Standards laboratories.
Copyright of this certificate is owned by the issuing laboratory and may not be reproduced except with the prior written approval of the issuing laboratory.
This certificate complies with the requirements of BS EN ISO 10012:2003.

CERTIFICATE OF CALIBRATION

Certificate Number

BTM1669

Page 2 of 3 Pages

Test Title	Tolerance	Applied Value	Reading	Pass/Fail
General Operation Tests				
Controls & Display	---	---	Pass	
200mA Protective Earth Resistance				
Resistance	16mΩ	0.09Ω	0.096Ω	Pass
Resistance	17mΩ	0.14Ω	0.153Ω	Pass
Resistance	19.4mΩ	0.26Ω	0.269Ω	Pass
Resistance	21.6mΩ	0.38Ω	0.383Ω	Pass
Resistance	25mΩ	0.55Ω	0.568Ω	Pass
Resistance	35mΩ	1.05Ω	1.071Ω	Pass
Insulation Resistance @ 250V				
1MΩ	220kΩ	1.00MΩ	1.0MΩ	Pass
5MΩ	300kΩ	5.00MΩ	5.0MΩ	Pass
10MΩ	400kΩ	10.00MΩ	9.9MΩ	Pass
15MΩ	500kΩ	15.00MΩ	14.9MΩ	Pass
19MΩ	580kΩ	19.00MΩ	18.9MΩ	Pass
50MΩ	4MΩ	50.00MΩ	49.2MΩ	Pass
90MΩ	7MΩ	90.00MΩ	87.9MΩ	Pass
Insulation Resistance @ 500V				
1MΩ	220kΩ	1.00MΩ	1.0MΩ	Pass
5MΩ	300kΩ	5.00MΩ	5.0MΩ	Pass
10MΩ	400kΩ	10.00MΩ	9.9MΩ	Pass
15MΩ	500kΩ	15.00MΩ	14.9MΩ	Pass
19MΩ	580kΩ	19.00MΩ	18.9MΩ	Pass
50MΩ	4MΩ	50.00MΩ	49.9MΩ	Pass
90MΩ	7MΩ	90.00MΩ	88.6MΩ	Pass
Leakage Current (AC+DC)				
2mA	30uA	2.04mA	2.06mA	Pass
4.7mA	57uA	4.88mA	4.93mA	Pass
Equipment Current				
O/C Test	---	---	Pass	
0.5A Test	225mA	0.50A	0.5A	Pass
Point To Point Voltage				
Mains	2.3V	102.90V	102.9V	Pass
Mains	4.3V	205.90V	205.6V	Pass

Uncertainties

Earth Bond Ohms	0.5% + 4mohms
Bond Current	0 to 30 Amps: 1.5% + 6counts.
PAT Leakage	1.5% + 0.3mA
Insulation Ohms	10kohms to 5Mohms: 0.1%, 5Mohms to 10Gohms: 1% +/- 1 Digit.
Frequency	0.1ppm ± 1digit
Resistance	0 to 10MOhms 0.005% ± 1digit: 10Mohms to 1Gohm 0.4% ± 1 digit
A.C. Current	0 to 10A: 0.02% ± 2digit
A.C. Voltage	0 to 1000V: 0.01% ± 1digit

CERTIFICATE OF CALIBRATION

Certificate Number
BTM1669

Page 3 of 3 Pages

Test Title	Tolerance	Applied Value	Reading	Pass/Fail
------------	-----------	---------------	---------	-----------

End of results

Uncertainties

Earth Bond Ohms	0.5% + 4mohms
Bond Current	0 to 30 Amps: 1.5% + 6counts.
PAT Leakage	1.5% + 0.3mA
Insulation Ohms	10kohms to 5Mohms: 0.1%, 5Mohms to 10Gohms: 1% +/- 1 Digit.
Frequency	0.1ppM ± 1digit
Resistance	0 to 10MOhms 0.005% ± 1digit: 10Mohms to 1Gohm 0.4% ± 1 digit
A.C. Current	0 to 10A: 0.02% ± 2digit
A.C. Voltage	0 to 1000V: 0.01% ± 1digit