

EC Declaration of Conformity

This declaration of conformity is issued under the sole responsibility of the manufacturer in accordance with the Council Directive 93/42/EEC in its currently valid version (Annex I) concerning medical devices.

Product: **Oxygen Monitor**

Type: **MySign® O**
incl. Accessories

Classification:
(RL 93/42/EEC,
Annex IX) **Class IIa**

CE mark: 

Notified Body: TÜV SÜD Product Service GmbH, Ridlerstr. 65, 80339 Munich, Germany

Conformity
Assessment Process: Annex II, section 3 of the Directive 93/42/EEC

Particular product
standard applied: ISO 80601-2-55


Issued by: **ENVITEC-WISMAR GMBH**
Alter Holzhafen 18
D-23966 Wismar
Germany

Valid from: **2018-12-04**
Valid until: **2019-06-30**

Place, Date: **Wismar, 2018-12-04**

Authorized Signature:


ppa. Marko Sprössel


ppa. Sandra Brieger