

DVM Calibration record

MICROCAL CHILL.

Equipment description if new	
Serial Number if new	8 58080
DVM Calibration Index Number	CE100
DVM Bar code	
Equipment Used	Time 1030 MicroCal + DVM. CE 169
Serial number Test Equipment	682926 683916
Calibration Certificate Number	F098426
Test Equipment Calibration Index Number	CE076 (Barcode 102060)
CE100	DVM CE 169
Micro Cal @	DVM Reading
0mV	
5mV	
10mV	10.013
15mV	15.196
20mV	20.087
25mV	25.089
30mV	30.061
35mV	35.096
40mv	40.008
Pass / Fail	
Tested By	JSL
Tested On	10/6/08
Signed	<i>[Signature]</i>

100mV	100.01
500mV	500.00
1.0V	1.0009

NB Set up is by visually setting a Verrier
 These results are accurate enough for use in testing Oxygen
 monitoring Products

[Signature]