



Steve Nixon <steve.nixon.viamed@googlemail.com>

Oxygen sensors returns review

1 message

Steve Nixon <steve.nixon@vandagraphst.com>
 Reply-To: steve.nixon@vandagraphst.com
 To: Achats M3S <achats@multi3s.com>

30 September 2019 at 15:29

Hi Patrice

Just a quick update concerning the returns rates for your sensors:

		First sold	Total returns	Faults	No Fault	Out of Warranty
8010018	Triton T22D Oxygen Sensor	02/05/2014	0.28%	0.14%	0.14%	0.00%
801xxxx	For comparison – VST whole range mean values	06/01/2011	0.61%	0.22%	0.07%	0.32%

I believe that the low incident of issues is pleasing and the fault percentage is within acceptable limits; taking into account the usage, application and environments that the sensors are used in.

Thank you for not returning any out of warranty sensors.

Kind regards

Steve

Steve Nixon
 Director - Vandagraph Sensor Technologies Ltd.

Office: +44 (0)1535 634900
 Mobile: +44 (0)7850 252267