

QA
0310200 Low-Flow Blender air & oxygen.
Barcode ID: 958128
Serial Number BX1651896

Carried out by: Philip Crossley
16/Sep/2020 1:05 pm

Performance evaluation: Pass

Signed: _____

Name: _____

Position: _____

Date: _____

QA 0310200 Serial Number BX1651896 Barcode ID: 958128
Carried out by: Philip Crossley 16/Sep/2020 1:05 pm

[1]2303 Asset Number

[2]2206 21% Output between 20.8 - 23
21.7

[3]2205 30% Output between 27 - 33
30.3

[4]2204 60% Output between 57 - 63
60.0

[5]2203 90% Output between 87 - 93
89.0

[6]2202 100% Output between 98-100
100.0

[7]2201 Check Alarm for no Oxygen
Yes

[8]2199 Check Alarm for no Air
Yes

[9]2198 Check O2/Air Balance Alarms before output change
Yes

[10]2197 Check Duck Bill Valves
Yes

[11]2196 Check AUX
Yes

[12]2195 Check for Leaks
Yes