

Gmail

Compose

Back Archive Spam Delete | Mark as unread Snooze | Move to Inbox Labels More

Inbox 364 **MHRA Adverse Incident Report - Maxtec EyeMax 2 (Premie)** Inbox x

Snoozed

Sent

Categories

Social 717 **Steve Nixon** 29 Jan 2020, 16:16 (12 days ago) **Reply**
to Bob, Derek, Steve
MHRA Adverse Incident Report - Maxtec EyeMax 2 (Premie)
Hi Bob
Can you please provide feedback and advice from Maxtec regarding the attached adverse incident report received from the MHRA. We have just learned of this and we need to construct a response ASAP either from Maxtec or from Viamed (on behalf of Maxtec).

Updates 1,185 Initial quick thoughts:
a) Did they affix the EyeMax 2 in accordance with the instructions?
b) Did they use the correct size EyeMax 2, did they measure the occipital-frontal circumference? What was the documented occipital-frontal circumference?
c) Was the EyeMax 2 overstretched at any time?
d) Was the EyeMax 2 used on a single patient? How long had the EyeMax 2 been in use on the patient concerned?
e) At the time of the reported incident, how often was the position of the EyeMax 2 checked?

Forums

Promotions 3,119

Envitec Invoices

Inspiration

 Derek +

 **Lee Ping** Ok dude no worries.. cheers for the sleeping around
You: Actually just had tea and practice

 Helen Lamb No one has joined your call

Regards >

<https://mail.google.com/mail/u/0/?tab=rm&ogbl#search/maxtec+eyemax/WhctKJVjVlcKJLpnFtkGHRmpjpMJJTvVMksbtWlgTGDxQqpDrIKkTWX...> 1/1