

Freigabe Label – Release for Label

☐ Label neu / new

☒ Label geändert / changed DCR-Nr.

Die Freigabe wird erteilt für – Released for:

☐ Artikelnummer – Part No.

E1002737

☐ Produkt – Product

☒ Product Label

Sensor Label 90 x 15

☒ Label Primär-Verpackung
/primary label

Blister Label 63,5 x 46,6

☐ Label Sekundär-Verpackung
/secondary label

☐ Sonstiges – Other


Label:

DIVESOFT Oxygen Sensor type 22D #3092 www.divesoft.com +420 416 857 057     S/N D123456	DIVESOFT Hájkova 2495 413 01 Roudnice nad Labem Czech Republic +420 416 857 057 #3092 www.divesoft.com info@divesoft.com Oxygen Sensor type 22D Fore use with Divesoft Liberty CCR  2020-02 S/N D123456      This oxygen sensor contains caustic liquid, in case of leakage avoid contact with eyes and skin. Do not puncture or staple this bag. Do not remove the oxygen sensor from this bag until ready to use. Depending upon the ambient conditions the oxygen sensor may require 30 minutes to reach signal stability after installation.
Sensor Label 90 x 15	Blister Label 63,5 x 46,6

Die Freigabe wird erteilt durch – Approved by:

	Funktion Function	Freigabe für Approval for	Unterschrift Signature	Initialen Initial	Datum Date
☒	R&D Engineer	Normenkonformität – Compliance with product standards			
☒	Produktmanager Product manager	Layout, Corporate Identity			
☒	QMB – Quality Manager	Übergeordnete Normen und Richtlinien – QMS standards & Directives			
☒	Vertrieb Sales	Kundenanforderungen – Customer requirements			
☒	FTV Production engineering	Probedruck, Stammdatenpflege Navision – Test print, maintenance of Navision material master data			
☒	Fertigungsleiter Production manager	Überführung in die Fertigung / Eintrag ins FKB – Transfer into production, entry into DHR			

Kundenfreigabe / Customer Approval

	Kunde/ Customer	Signature	Name	Date
☒	VANDAEER SENSOR TECHNOLOGIES		S. Nijon	10/02/20