

Medical Oxygen Sensors



**Highest quality and reliability
to meet the high standards in
medical applications**

ENVITEC-Wismar GmbH

Alter Holzhafen 18
23966 Wismar, Germany

Tel. +49-(0)-3841-360-1
Fax. +49-(0)-3841-360-222
e-mail: info@envitec.com
<http://www.envitec.com>



A member of the **Biotest** Group

Medical Oxygen Sensors

	Type	Compatible with *
	OOM101	Hudson 5556; MSA 655264; Teledyne R15; Ohmeda 0237-2034-700
	OOM103	Teledyne R22CC
	OOM102-1	Hudson 5567; MSA 406931; Teledyne R17
	OOM104	Designed for Dräger Incubator I 8000 only
	OOM105	Hudson 5558; MSA 804674; Teledyne T7
	OOM107	Hudson 5569; MSA 485905; Teledyne T1, T2
	OOM107-2	Hudson 5568; MSA 478841; Teledyne C1R, C2R
	OOM201	Dräger 6850645; MSA 655263; Teledyne R23
	OOM202	MSA 472062; Teledyne R24; Siemens 6419332 E380E, 9004979 E347E
	OOM102	Teledyne R22
	OOM106	Datascope 0600-00-0002
	OOM202-2	MSA 806572
	OOM110	Ohmeda 6050-0004-110

* Main compatible manufacturers for guidance purposes only. For additional details of other compatible manufacturers, please refer to EnviteC's cross reference table for medical oxygen sensors.

