



Handheld Mainstream and Sidestream
CO₂/SpO₂ Monitors VM-2500-M and VM-2500-S
User Manual

VM-2500 User Manual

Version: EN 2.0
Released: 12/2009

The information in this document is subject to change without notice.

Copyright © 2009 Viamed Ltd.. All Rights Reserved. This document contains proprietary information that is protected by copyright. All rights are reserved. No part of this document may be photocopied, reproduced or translated to another language without prior written consent of Viamed Ltd..

Contact information

Viamed Ltd.
15 Station Road
Cross Hills
Keighley
West Yorkshire BD20 7DT
United Kingdom

Tel: +44 (0)1535 634542
Fax: +44 (0)1535 635582
E-mail: info@viamed.co.uk

www.viamed-online.com

Liability

Viamed Ltd. shall in no event be liable for any direct, indirect, special or consequential damages including without limitation damages for loss of business profits, loss of income, business interruption, loss of business information, loss of use or other related exposures, however caused, arising from the faulty or incorrect use of the product.

Viamed Ltd. guarantees that the product delivered has been thoroughly tested to ensure that it meets its published specifications.

Warranty

Viamed Ltd. warrants the products manufactured or distributed by them to be free from faulty materials and workmanship for a period of 24 months from date of original shipment to first end user except for disposable products or products which have a stated guarantee longer or shorter than 24 months. Viamed Ltd. will perform warranty service at its factory.

The obligations for Viamed Ltd. under this guarantee shall be limited to repair, or at Viamed Ltd.'s discretion, replacement of necessary parts or assemblies and shall not include shipping costs or other incidental damages.

Claims for damages during shipment must be filed promptly with the transportation company. All correspondence concerning the products must specify both the name of the product and its serial number as written on the label on the product.

Use of the equipment for other than its intended use, or if it has been repaired by anyone except Viamed Ltd. or a Viamed Ltd. authorized service centre, or altered or modified or used without following the instructions in the user manual, will void this guarantee.

Patents

PHASEIN AB holds the following patents regarding products described in this manual:
SE519766; SE519779; SE523461; SE524086. Other patents pending

Trademarks

PHASEIN IRMA™, PHASEIN XTP™, ISA™ and LEGI™ are trademarks of PHASEIN AB.

Table of Contents

1	Introduction	6			
1.1	Intended Use	6	5.5.3	SpO ₂ Sensor	22
1.2	Warnings	6	5.6	Commencing Monitoring	23
1.3	Cautions	9	5.7	Switching off the Device	23
1.4	Symbol Description	10			
1.5	Terms and Definitions	10	6	Display Modes and Displayed Data	24
1.6	User Requirements	11	6.1	Toggling Between Display Modes	24
			6.2	Symbols and Indicators	25
			6.3	Pulse Tone	25
2	Theory of operation	12			
2.1	CO ₂ measurement	12	7	Alarms	26
2.2	SpO ₂ measurement	13	7.1	Alarm Priority and Appearance	26
			7.2	Audible Alarm Volume	26
3	Product overview	13	7.3	Default Alarm Limits	26
3.1	Available models	13	7.4	Limit Alarm	27
3.2	Mainstream Capnography Accessories	14	7.5	Alarm Messages	27
3.2.1	IRMA CO ₂ analyzer	14	7.6	Resetting of Alarm Signals	28
3.2.2	IRMA airway adapter	14			
3.3	Sidestream Capnography Accessories	15	8	Menu Structure	28
3.3.1	Nomo Adapter	15	8.1	Main Menu	28
3.3.2	Sidestream sampling line, airway adapter and other patient interfaces	16	8.1.1	Submenu: Alarm Settings	28
3.4	SpO ₂ Sensors	16	8.1.2	Submenu: Data Management	29
			8.1.2.1	Recording Data	29
			8.1.2.2	Data Management	29
4	Exterior View, Controls and Connectors	17	8.1.3	Submenu: Setup	30
4.1	Mainstream Capnograph/Pulse Oximeter - VM-2500-M	17	8.1.3.1	General Information	30
4.2	Sidestream Capnograph/Pulse Oximeter - VM-2500-S	18	8.1.3.2	Parameters	31
			8.1.4	Submenu: Patient ID	34
			8.1.5	Default Start Settings	34
5	Preparation for Use	19	8.2	Other	34
5.1	Selecting Power Supply	19	8.2.1	Volume Control Shortcut	34
5.1.1	Power Supply	19	8.2.2	Brightness Control Shortcut	34
5.1.2	Rechargeable Li-ion Battery or AA Alkaline Batteries	19	8.2.3	Power-Save Mode	34
5.2	Connecting Sensors/Sampling Line Configurations to the VM-2500	20	9	Adverse affects on performance	34
5.2.1	IRMA™ CO ₂ Analyzer (mainstream)	20	9.1	CO ₂ measurements	34
5.2.2	Nomo technology sampling line configuration (sidestream)	20	9.1.1	Humidity or Condensation	34
5.2.3	SpO ₂ Sensor	21	9.1.2	Interfering gases or vapours	35
5.3	Visual Check	21	9.2	SpO ₂ Measurement	35
5.4	Switching on the Device	21			
5.5	Connecting Sensors/Sampling Line Configurations to the Patient	21	10	Troubleshooting guide	36
5.5.1	IRMA™ CO ₂ Analyzer (mainstream)	21	10.1	Error Message – Cause – Corrective Action	36
5.5.2	Nomo Technology sampling line configuration (sidestream)	22	10.2	Failure – Cause – Corrective Action	37
			10.3	IRMA™ LED Status and LEGI™ Status	38
			10.4	Problems with EMI (Electromagnetic Interference)	38

11	Maintenance	38
11.1	Maintenance	38
11.2	Cleaning	38
11.3	Testing	39
11.4	Zeroing the Capnograph	39
11.4.1	Mainstream VM-2500-M	39
11.4.2	Sidestream VM-2500-S	39
12	VM-2500 PC-Software	40
13	Technical Specifications	41
14	Component Lists and Ordering	43
14.1	Packing List	43
14.2	Order Number	44
14.3	Accessories	44
14.4	Replacement Parts	45

1 Introduction

1.1 Intended Use

The VM-2500 CO₂/SpO₂ Monitor is intended to provide continuous monitoring of end-tidal CO₂ concentration (EtCO₂), inspired CO₂ concentration (FiCO₂), functional arterial oxygen saturation (SpO₂), respiration rate (RR) and pulse rate (PR) of adult, paediatric infant and neonatal patients. The CO₂ measurement can be performed either with the mainstream method using the VM-2500-M or with the sidestream method using the VM-2500-S.

It may be used in hospital, hospital type facilities, operating suites, intensive care units, emergency medicine/emergency transport as well as in the home care environment.

The VM-2500 CO₂/SpO₂ Monitor is not intended to be used as the only means of monitoring a patient. It shall always be used in combination with other vital sign monitoring devices and/or professional human judgments of a patient's condition. The VM-2500 is intended to be used by trained and authorized health care professionals only.

1.2 Warnings

Adhere to the following warnings for safe operation of the VM-2500-M mainstream CO₂/SpO₂ Monitor and the VM-2500-S sidestream CO₂/SpO₂ Monitor.

For VM-2500 in General:

 *Warning: The monitor is to be operated only by trained personnel and is for attended monitoring only.*

 *Warning: Do not make any clinical judgments based solely on the VM-2500. The monitor is intended only as an adjunct in patient assessment. It must be used in conjunction with clinical signs and symptoms. The interpretation of the measurement values should be done only by trained health care professionals.*

 *Warning: The monitor should only be used for the purpose and in the manner described in this manual.*

 *Warning: Explosion hazard. Do not use the monitor in the presence of flammable anaesthetic mixtures with air, oxygen, or nitrous oxide.*

 *Warning: Routinely monitor the patient to make sure that the VM-2500 is functioning correctly and*

the SpO₂ sensor, the mainstream IRMA™ CO₂ analyzer and the CO₂ sidestream sampling configuration are correctly placed.

 *Warning: Do not place the monitor in any position that might cause it to fall on to the patient.*

 *Warning: Certain environmental and physiological conditions, medical procedures, sensor application errors and external agents may interfere with the monitor's ability to detect and display accurate measurements. (Chapter 9 provides information on possible interferences)*

 *Warning: For the SpO₂ measurement, the monitor uses red and infrared light with specific fixed wavelengths. Consider that these wavelengths might influence diagnostic parameters of other optical applications. The specifications of the wavelengths used are listed in the 'Instructions for Use' of the specific sensor.*

 *Warning: The monitor detects respiratory effort via changes in CO₂ concentration of exhaled air; therefore, the CO₂ measurement can be used to detect apnoea. The device however is unable to discriminate between a patient not breathing and a sensor that is disconnected from the patient circuit. Always monitor and set alarms for SpO₂ when using the VM-2500 to monitor respiratory function.*

 *Warning: If you are uncertain about the accuracy of any measurement, check the patient's vital signs by alternate means; then make sure the monitor is functioning correctly.*

 *Warning: The use of accessories, sensors, and cables other than those specified may result in increased electromagnetic emission and/or create invalid readings of the monitor.*

 *Warning: In high ambient light conditions it is required to shield the SpO₂ sensor application site with opaque material. Too much ambient light may result in inaccurate measurements.*

 *Warning: Check all alarm settings and auditory alarm before use of the monitor.*

 *Warning: Do not silence the audible alarm function, or decrease the audible alarm volume if patient safety could be compromised.*

⚠ Warning: No modifications of the monitor are allowed without authorization of the manufacturer.

⚠ Warning: Measurements can be affected by mobile and RF communications equipment. Make sure that the monitor is used in the electromagnetic environment specified in this manual.

⚠ Warning: Disconnect the monitor and probes from the patient during computed tomography (CT).

⚠ Warning: Disconnect the monitor and probes from the patient during magnetic resonance imaging (MRI) scanning. An induced current could potentially cause burns.

Additional warnings regarding the power supply system: To prevent the possibility of the Li-ion battery CT-2500 from leaking, heating or explosion, observe the following precautions.

⚠ Warning: Do not immerse the battery in water or seawater. Store it in a cool and dry environment if not used.

⚠ Warning: Do not discard, store or use the battery near a heat source (e.g. a fire or heater).

⚠ Warning: Only charge the Li-ion battery CT-2500 while inserted in the VM-2500 monitor using the provided power supply FW 7660M/06.

⚠ Warning: Do not connect the positive and negative terminal with metal objects such as wire and do not transport or store the battery together with metal objects such as necklaces or hairpins as this may short-circuit the battery.

⚠ Warning: Do not strike, throw or trample the battery or pierce it with a nail or other sharp object.

⚠ Warning: Only use the Li-ion battery Model No. CT-2500 provided with the monitor. Contact the manufacturer for replacements.

⚠ Warning: Only use the power supply Model No. FW 7660M/06 provided with the monitor. The use of a power supply other than this may result in hazardous situation and affect patient's safety.

⚠ Warning: Do not use rechargeable AA sized batteries instead of alkaline AA sized batteries to operate the device, as this may affect the function of the device.


Additional warnings regarding the mainstream device - VM-2500-M:

⚠ Warning: Disposable IRMA™ airway adapters shall not be reused. Used airway adapters shall be disposed of in accordance with local regulations for medical waste.


⚠ Warning: Do not use the IRMA™ airway adapter (adult/paediatric) with infants as the adapter adds 6ml dead space to the patient circuit.

⚠ Warning: Do not use the IRMA™ airway adapter (infant) with adults or paediatric as this may cause excessive flow resistance.

⚠ Warning: Do not place the IRMA™ airway adapter between the endotracheal tube and an elbow as this may allow patient secretions to block the adapter windows and result in incorrect operation.



⚠ Warning: To keep secretions and moisture from pooling on the windows, always position the IRMA™ CO₂ analyzer in a vertical position.



⚠ Warning: Do not use the IRMA™ CO₂ analyzer with metered dose inhalers or nebulised medication as this may affect the light transmission of the airway adapter windows.

⚠ Warning: A successful zeroing requires the presence of ambient air (21% O₂ and 0% CO₂) in the IRMA™ airway adapter during zeroing. Incorrect zeroing of the IRMA™ CO₂ analyzer will result in false gas readings.

⚠ Warning: Replace the IRMA™ airway adapter if condensation occurs inside the adapter.

Additional warnings regarding the sidestream device – VM-2500-S:

 *Warning: Do not use the VM-2500-S to analyse gasses (e.g. anaesthetics like nitrous oxide etc.) that need to be returned to the patient circuit or a scavenging system. The gas outlet of the sidestream device is not designed to return the exhaust gases to the patient circuit or a scavenging system.*

 *Warning: Excessive positive or negative pressure in the patient circuit might cause incorrect readings and internal damage of the monitor.*

 *Warning: Carefully route the sampling line to reduce the risk of patient entanglement or strangulation.*

 *Warning: Do not lift the VM-2500-S by the sampling line as this could disconnect from the monitor, causing the monitor to fall on the patient.*

 *Warning: Do not use adult or paediatric type sampling line configurations with infants or neonates, as this may add dead space to the patient circuit.*

 *Warning: Do not use infant or neonatal type sampling line configurations with adults or children, as this may result in excessive flow resistance.*

 *Warning: Do not use the VM-2500-S with metered dose inhalers or nebulised medication as this may block the bacteria filter integrated in the Nomo Adapter.*

 *Warning: Since a successful zeroing requires the presence of ambient air (21% O₂ and 0% CO₂) in the device, ensure that the VM-2500-S is placed in a well ventilated environment. Avoid breathing near the monitor before or during the zeroing procedure. Incorrect zeroing of the integrated ISA™ CO₂ analyzer will result in false gas readings.*

 *Warning: Replace the blocked part of the sampling line configuration if a sampling system occlusion message is displayed by the monitor.*

 *Warning: Do not re-use the disposable sampling lines and airway adapters.*

 *Warning: Used disposable sampling lines shall be disposed of in accordance with local regulations for medical waste.*

 *Warning: The Nomo Adapter is reusable. It only needs to be replaced if occluded. Replace the Nomo Adapter if “Sampling system occlusion!” is displayed by the monitor whilst only the Nomo Adapter is connected.*

1.3 Cautions

Adhere to the following recommendations to avoid damage or malfunction of the VM-2500-M mainstream CO₂/SpO₂ Monitor and VM-2500-S sidestream CO₂/SpO₂ Monitor.

For VM-2500 in General:

-  *Caution: Do not spray, pour, or spill any liquid on the monitor, its accessories, connectors, switches, or openings in the enclosure as this may damage it.*
-  *Caution: Do not immerse the monitor or its accessories in liquid.*
-  *Caution: Do not autoclave or steam sterilize the monitor or its accessories.*
-  *Caution: Refer to the specific 'Instructions for Use' of the used SpO₂ sensor for correct cleaning and/or sterilization.*
-  *Caution: Do not apply excessive tension to any of the monitor cables.*
-  *Caution: Do not operate the monitor outside the specified operating temperature environment.*
-  *Caution: The monitor requires no routine calibration. A basic maintenance plan conducted by qualified service personnel is recommended. Please refer to the Service Manual for detailed information.*
-  *Caution: There are no user-serviceable parts inside the VM-2500. The cover should only be removed by qualified service personnel.*

Additional cautions regarding the power supply system:

-  *Caution: Do not use or store the Li-ion battery at very high temperature conditions e.g. strong direct sunlight or in a heated vehicle. Under these conditions the battery can overheat, causing it to burn or its performance will degenerate and its service life will be decreased.*
-  *Caution: Do not use the Li-ion battery in an electromagnetic environment other than specified in this manual, as this may damage the safety features of the battery and result in unforeseen danger.*

 *Caution: If the Li-ion battery leaks and the electrolyte gets into the eyes, immediately rinse the eyes with clean running water and seek medical assistance to prevent injury of the eyes.*

 *Caution: If the Li-ion battery gives off an odour, generates heat, becomes discoloured or deformed, or in any way appears abnormal during use, recharging or storage, immediately remove it from the device and stop using it.*

 *Caution: If the Li-ion battery terminals are dirty, clean the terminals with a dry cloth before use, otherwise a power failure or charging failure may occur due to poor connection with the device.*

 *Caution: Discharged batteries may cause fire; tape the terminals to isolate them.*

 *Caution: Remove the batteries if the device is to be stored or not used for a longer period of time.*

Additional cautions regarding the mainstream device - VM-2500-M:

-  *Caution: Do not immerse the IRMA™ CO₂ analyzer in liquid.*
-  *Caution: The IRMA™ CO₂ analyzer and the IRMA™ airway adapters are non-sterile devices. To avoid damage, do not autoclave these components.*
-  *Caution: Use only PHASEIN manufactured IRMA™ airway adapters.*

Additional cautions regarding the sidestream device - VM-2500-S:

-  *Caution: The Nomo Adapter, the sampling lines and their patient interfaces are non-sterile devices. To avoid damage, do not autoclave these components.*
-  *Caution: Use only Nomo Adapters manufactured by Viamed.*

1.4 Symbol Description

	Accompanied with “ <i>Warning: Supplementary text.</i> ” within this document. Warnings indicate potential harmful conditions that may lead to injury or death.
	Accompanied with “ <i>Caution: Supplementary text.</i> ” within this document. Cautions indicate conditions that may lead to damage to or malfunction of the device.
	Denoted as “ <i>Note: Supplementary text.</i> ” within this document. Notes inform the user to relevant facts and conditions in connection with the device.
	Consult User Manual for detailed operating information.
	Consult accompanying documents for important safety-related information
	Manufacturer
	Date of manufacture
	Type BF applied part
REF	Catalogue number
SN	Serial number
P/N	Part number
LOT	Batch code
	Use by [YYYY-MM-DD] (indicates that the device should not be put into service after the date accompanying the symbol)
	Temperature limitation
	Do not re-use
	Fragile, handle with care
	Do not dispose in the consumer waste. Electrical and electronic equipment shall be collected and recycled in accordance with (Directive 2002/96/EC)

	European Union approval (complies with 93/42/EEC Medical Device Directive)
IPXY	IP Code (International Protection Rating)

1.5 Terms and Definitions

BTPS	Body Temperature and Pressure Saturated
VM-2500	Handheld CO ₂ /SpO ₂ monitor developed by Viamed. Available as mainstream (VM-2500-M) or side-stream (VM-2500-S) version.
CO ₂	Carbon dioxide
EtCO ₂	End-tidal expired CO ₂ gas concentration
FiCO ₂	Inspired CO ₂ gas concentration
Hb	Deoxygenated haemoglobin
HbO ₂	Oxygenated haemoglobin
HME	Heat Moisture Exchanger
IR	Infrared
IRMA™	Infrared mainstream CO ₂ analyzer.
ISA™	Infrared sidestream CO ₂ analyzer
LEGI™	Light Emitting Gas Inlet: Status indicator integrated in the gas sample inlet port.
Li-ion battery	Lithium-ion rechargeable battery
MDD	Medical Device Directive
MRI	Magnetic Resonance Imaging
N/A	Not applicable. Data does not apply to the configuration.
Nomo Adapter	Adapter that is connected between the sidestream VM-2500-S and the selected sampling line. The Nomo Adapter removes water and water vapour from the sampling line and has a hydrophobic bacterial filter to reduce the potential for water intrusion and cross contamination.

Nomo technology	Unique technology to remove water vapour and aspirated or condensed water from the sidestream CO ₂ gas sample through a membrane-like surface into the surrounding air. (Nomo stands for “no-moisture”)
PR	Pulse rate
RF	Radio frequency
RH	Relative humidity
Rise time	Time required to achieve an increase from 10% to 90% of final value when step function change in concentration occurs at the sampling site.
RR	Respiration rate
Sampling line config.	A sampling line configuration consists of a Nomo Adapter connected to a sampling line with an appropriate patient interface for sampling gas, e.g. a nasal cannula, respiratory mask or the Y-piece of an intubated patient.
SpO ₂	Functional arterial oxygen saturation
TBD	To Be Determined. Value or property not yet decided; further investigations may be necessary.
Total system response time	Time from a step function change in gas level at the sampling site to the achievement of 90% of the final gas reading of the capnograph. Total system response time = Delay time + Rise time
USB	Universal Serial Bus
Zeroing	Ambient gas reference measurement used to establish zero concentration level for CO ₂ . Zeroing needs to be performed ONLY when an offset in gas measurement values is observed, or when an unspecified accuracy message is displayed.

1.6 User Requirements

The user(s) of the monitor shall have an in-depth knowledge of gas analyzing and non-invasive monitoring of functional arterial oxygen saturation.

2 Theory of operation

2.1 CO₂ measurement

Principle

The measurement of CO₂ in gas mixtures with the VM-2500 is based on the fact that different gases absorb infrared light at specific wavelengths. The absorption spectra for CO₂, N₂O and different anaesthetic agents are shown in Figure 1. The CO₂ analyzer of the VM-2500 uses the absorption peak at 4.2μm to determine the CO₂ concentration in the gas. Two additional wavelengths beside this absorption peak are used as reference.

To measure the absorption of light at these wavelengths, a broadband infrared radiation source is

Capnogram

A capnogram is a graph representing the CO₂ concentration in respiratory gasses plotted against time. The capnogram waveform is typically divided into 4 phases (Bhavani-Shankar & Philip, 2000). In the waveform below the inspiration (phase 0) is plotted in blue and the expiration (phase I - III) in red.




Figure 2: Normal capnogram waveform




Figure 1: Gas absorption spectra

used. The light transmitted from the infrared source passes through the gas sample and is then filtered using a set of narrow optical band pass filters. The individual filters are mounted in a rapidly rotating filter wheel that intersects the light path before the light reaches the infrared detector.

The infrared detector detects the portion of the light that is not absorbed by the gas. The amplitude of the detector output is an inverse function of the gas concentration. Thus, at a concentration of zero, the amplitude is at its maximum.

Phase I: Baseline (FiCO₂)

Phase II: Positive expiration slope (alveolar gas)

Phase III: Alveolar plateau (EtCO₂)

Phase 0: Inspiration

A capnogram provides valuable information regarding the health situation of patients in respiratory distress.

Measurement methods

Determining the CO₂ concentration of respiratory gases is done either using the mainstream method or

the sidestream method. In both cases the measurement principle is very similar (infrared spectroscopy) the position of the sensor in the measurement setup however differs.

In mainstream capnography an airway adapter is, for example, inserted between the endotracheal tube and the Y-piece of the breathing circuit. The airway adapter has an optical window over which the infrared mainstream CO₂ sensor is positioned. The respiratory CO₂ measurements are obtained by continuously measuring the infrared light absorption, in the gas flow, through the optical windows.

For sidestream capnography a sampling line is connected to an endotracheal tube with the help of a sidestream airway adapter or directly from the nose via a nasal prong. Gas samples are continuously taken from the patient at a constant flow rate (VM-2500-S has a suction rate of 50ml/min) and transferred to the infrared CO₂ sensor incorporated in the capnograph. The CO₂ measurement is taken within the device and after measurement the waste gas is exhausted through the gas outlet.

Typically water traps and bacteria filters are positioned between the sampling line and the device inlet to avoid damage to the device.

2.2 SpO₂ measurement

SpO₂ measurement is performed in transmission mode. The SpO₂ sensor consists of emitters which pass red and infrared light through peripheral sites such as a finger, toe or ear to a light-sensitive detector.

For both wavelengths the change in absorption is continuously measured. In this way the pulsatile signal due to the arterial blood alone is extracted, excluding the offset due to absorption by venous blood, skin, bone, muscle and fat.

This signal is used to determine the functional arterial oxygen saturation (SpO₂) based on the fact that the amount of absorbed infrared light and red light is different for oxygenated haemoglobin (HbO₂) and deoxygenated haemoglobin (Hb).

The amount of red and infrared light received is compared and the percent of haemoglobin molecules bound with oxygen molecules calculated:

$$SPO_2 = \frac{HbO_2}{HbO_2 + Hb}$$

These measurement values are continuously displayed as a waveform (plethysmogram) and also used to determine the pulse rate. Note that the pulse oximeter equipment is calibrated to display the functional oxygen saturation.

Certain physiological conditions, medical procedures, and external agents may interfere with the monitor's ability to detect and display accurate SpO₂ measurements (see chapter 9.2 for detailed information).

3 Product overview

3.1 Available models

Mainstream Capnograph/Pulse Oximeter – VM-2500-M

The VM-2500-M is used together with an IRMA™ CO₂ analyzer, an IRMA™ airway adapter and an application appropriate SpO₂ sensor.

Power is delivered via either the provided power supply, a rechargeable Li-ion battery or 4 x AA batteries.

Mainstream CO₂ monitoring is performed with the IRMA™ CO₂ analyzer. This key technology sets new standards in CO₂ mainstream measurement and provides reliable, safe and easy CO₂ monitoring.

Key features of VM-2500-M:

- Warm-up time < 10s
- Direct measurement without time delay
- Small, light-weight and shock-resistant:
IRMA™ CO₂ analyzer weighs less than 30 g
- Adult/paediatric and infant IRMA™ airway adapters available
- Easy plug and measure technology
- IRMA™ airway adapter with XTP™ non-condensing light transmission window
- Maintenance and calibration-free technology
- Wide range of high-quality SpO₂ sensors available
- 2 years warranty

Sidestream Capnograph/Pulse Oximeter – VM-2500-S

The VM-2500-S monitor is used together with a Nomo Adapter, a suitable sampling line configuration (e.g. sampling line, airway adapter, nasal/oral sampling) and an application appropriate SpO₂ sensor.

Power is delivered via either the provided power supply, a rechargeable Li-ion battery or 4 x AA batteries.

Sidestream CO₂ monitoring is performed with the ISA™ CO₂ analyzer incorporated in the VM-2500-S. Together with the Nomo technology that removes water from the sampled gas, the ISA™ CO₂ analyzer with integrated sampling pump delivers the highest performance in an ultra-compact format.

Key features of VM-2500-S:

- Warm-up time < 10s
- Specially designed for all applications using low flow sampling (50 ml/min); from adults to neonates
- For intubated and non-intubated patients
- Nomo technology incorporated in the Nomo Adapter to remove water and water vapour from the sampling line.
- Hydrophobic bacterial filter incorporated in the Nomo Adapter to reduce the potential for water intrusion and cross contamination.
- Maintenance and calibration-free technology
- Wide range of high-quality SpO₂ sensors available
- 2 years warranty

3.2 Mainstream Capnography Accessories

3.2.1 IRMA™ CO₂ analyzer

The IRMA™ CO₂ analyzer is an ultra compact measurement device for mainstream CO₂ measurement with the VM-2500-M. As all necessary calibration constants are stored within each IRMA™ CO₂ analyzer, the analyzer can be replaced without the need for recalibration.

The analyzer has a rugged mechanical design providing reliable shock resistance.



Figure 3: IRMA™ CO₂ analyzer with airway adapter

To perform gas measurements, the IRMA™ CO₂ analyzer requires an airway adapter. The IRMA™ CO₂ analyzer connects in place on top of the IRMA™ airway adapter. This airway adapter is, for example, inserted between the endotracheal tube and the Y-piece of the breathing circuit.

Respiratory CO₂ measurements are obtained by continuously measuring the infrared light absorption, through optical XTP™ windows, in the gas flow through the adapter.

3.2.2 IRMA™ airway adapter

The IRMA™ airway adapter is designed as a single-patient-use disposable product and is available as:

- IRMA™ airway adapter (adult/paediatric) for patients over 1 year of age or 10kg in weight.
- IRMA™ airway adapter (infant) for patients up to 1 year of age or 10kg in weight.

The IRMA™ airway adapter (infant) has specially designed connectors for minimizing the dead space and can be used on very small patients.



Figure 4: IRMA™ airway adapters: Infant (left) and Adult/Paediatric (right)

As the airway adapter is positioned directly in the airway, its performance can be affected by water vapour, patient secretions or nebulised medications that can accumulate on the adapter's windows.

The water vapour can condense on the surface of the adapter windows in the format of small discrete water droplets. This condensation can affect light absorption through the windows, thus affecting the precision of the measurement.

The XTP™ windows of the IRMA™ airway adapter therefore have special features that prevent a decrease in performance when vapour is present. Using the latest advances in material technology they are designed to provide a window minimizing the impact of water vapour on light transmission.

For optimal results, the airway adapter shall not be placed between an endotracheal tube and an elbow, as this may allow patients secretions to block the adapter window.

3.3 Sidestream Capnography Accessories

Sidestream CO₂ measurement is performed by continuously removing a gas sample at a flow of 50ml/min from the respiratory circuit and analyzing it within the sidestream ISA™ CO₂ analyzer incorporated in the VM-2500-S.

To transport the gas sample from the patient to the device, a sampling line configuration is required.

A sampling line configuration consists of a Nomo Adapter connected to a sampling line with an appropriate patient interface for sampling gas, e.g. a nasal cannula, respiratory mask or the Y-piece of an intubated patient.

3.3.1 Nomo Adapter

The gas sample coming from the sampling line is fed through the Nomo Adapter to the VM-2500-S.

The Nomo Adapter is specifically designed for 50ml/min low sampling flow applications. It has a very low dead space that results in an ultra-fast rise time, making CO₂ measurement possible, even at high respiratory rates. The VM-2500-S application therefore ranges from neonates to adults.

The Nomo Adapter contains an unique water separation section and a hydrophobic bacterial filter to protect the ISA™ CO₂ analyzer against water intrusion and cross contamination.



Figure 5: Nomo Adapter

No condensing water:

Because of condensing water within the connected sampling line, droplets can potentially result in occlusion of the sampling system and interfere with the CO₂ measurement. To prevent this situation the Nomo Adapter incorporates an unique water separation section, the NOMO section.

This section is made of a special polymer and a hydrophobic bacteria filter and removes water vapour and aspirated or condensed water. Water and water vapour passes through the membrane-like surface of the sampling line and evaporates into the surrounding air, while leaving CO₂ unaffected.

Bacterial filter:

To protect the sidestream ISA™ CO₂ analyzer against cross contamination, the Nomo Adapter includes a filter with bacterial filter efficiency of $\geq 99.9980\%$. It is important to be aware that secretion and nebulised medication may be absorbed by the surface of the bacteria filter, and may block the Nomo Adapter.

Replacement of Nomo Adapter:

The Nomo Adapter is reusable as the Nomo technology allows up to two weeks of continuous monitoring with high moisture containing breathing gases.

⚠ Warning: *The Nomo Adapter is reusable. It only needs to be replaced if occluded. Replace the Nomo Adapter if "Sampling system occlusion !" is displayed by the VM-2500-S whilst only the Nomo Adapter is connected.*

3.3.2 Sidestream sampling line, airway adapter and other patient interfaces

Sampling with intubated patient

Sampling with intubated patients is typically performed with an appropriate sidestream airway adapter connected to the Nomo Adapter with a disposable sampling line. The sidestream airway adapter is designed as a single-patient-use disposable product and is available in different sizes ranging from adults through to neonates.



Figure 6: Typical sampling line assembly for intubated patients:
1. Disposable sampling line,
2. Sidestream airway adapter (adult/paediatric),
3. Sidestream airway adapter (infant/neonatal)

Sampling with non-intubated patient

Sampling with non-intubated patients is typically performed with a nasal or oral sampling configuration. This sampling assembly is directly connected to the nose or mouth with a cannula. For further instructions, warnings and precautions refer to the 'Instructions for Use' provided with the selected sampling system assembly.

Note: Our sampling line configurations are continuously updated; please contact Viamed for detailed information. Specific patient sampling line configurations and interfaces are available upon request.

3.4 SpO₂ Sensors

Viamed supplies a large variety of disposable and reusable SpO₂ sensors for use with the VM-2500.

Depending on the sensor type and model, their application ranges from adults to neonates, providing application specific features and design.

The SC 6500 VM Sensor is typically used for adult applications and the W 6500 VM Sensor for infant and neonatal applications (see figure 7). Other sensors are available upon request.




Figure 7: SpO₂ sensor SC 6500 VM (left) and W 6500 VM (right)

All sensors operate in transmission mode. The light source emits red and infrared light with wavelengths of 660nm and 905nm respectively at a typical radiant power of 3.5mW.

The SpO₂ sensor is applied to peripheral areas of the body such as finger tips and toes for adults, and paediatrics, and the foot or palms for infants and neonates.

4 Exterior View, Controls and Connectors


4.1 Mainstream Capnograph/Pulse Oximeter – VM-2500-M




No.	Symbol	Feature/Button	Function
1		Arrow buttons (up/down)	Multifunction buttons used for: 1. Scrolling through menu items 2. Increasing/decreasing parameters 3. Shortcuts to volume/brightness control during monitoring
2		Display mode	Toggles between various display modes. Shortcut to return to display mode during menu mode.
3		Alarm silenced	The audible alarm can be silenced for a maximum period of two minutes. Optical alarm remains activated.
4		Menu	Menu selection. Shortcut to return to the previous menu level during menu mode.
5		ENTER button	Confirms selection
6		Pulse tone	Turns pulse tone on/off
7		USB	USB 2.0 interface
8		On/Off	To turn on the device: press and hold power button briefly. To turn off the device: press and hold power button for approx. 3 seconds.

9		Power input	Port to connect the external power supply (100-240V AC / 50-60Hz, Model No. FW 7660M/06)
10		SpO ₂ Sensor port	Port to connect the SpO ₂ sensor
11		CO ₂ Mainstream port	Port to connect the IRMA™ CO ₂ analyzer

4.2 Sidestream Capnograph/Pulse Oximeter – VM-2500-S



Rear view



Top View

Front and side views are the same as for VM-2500-M (see Chapter 4.1)

No.	Symbol	Feature/Button	Function
9		Power input	Port to connect the external power supply (100-240 V AC / 50-60Hz, Model No. FW 7660M/06)
12		Gas outlet port	Gas outlet of the sidestream module (not applicable for mainstream)
13		SpO ₂ sensor port	Port to connect the SpO ₂ sensor
14		CO ₂ Sidestream gas inlet port	Port with Light Emitting Gas Inlet (LEGI™) to connect the Nomo Adapter

5 Preparation for Use

5.1 Selecting Power Supply

Power is supplied to the monitor either via external power supply, rechargeable Li-Ion battery, or 4 x AA alkaline batteries.

5.1.1 Power Supply

The external power supply (100-240V AC/50-60Hz, Model No. FW 7660M/06) is used for continuous operation of the monitor and to charge the Li-ion battery.

Connecting the power supply (see figure 8):

1. To operate by mains, connect the power supply cable into the power input socket located at the back of the device.
2. Ensure that the correct power supply plug is connected to the power supply. It can be exchanged by pressing the release button (2.1) on the power supply. As standard the device is supplied with a European and United Kingdom plug. Additional plugs are available upon request.
3. Connect the power supply to an AC outlet.




Figure 8: Connect the power supply

⚠ Warning: Only use the power supply Model No. FW7660M/06 provided with the monitor. The use of a power supply other than this may result in hazardous situation and effect patient's safety.

5.1.2 Rechargeable Li-ion Battery or AA Alkaline Batteries

For convenient monitoring in emergency medicine or during patient transport the monitor is supplied by the rechargeable Li-ion battery (3.7 V/2500 mAh, Model No. CT-2500) or with 4 x AA alkaline batteries.

When the device is connected to an AC outlet the Li-ion battery will begin recharging. This is represented by the three segments of the battery level indicator illuminating in sequence. When the Li-ion battery is completely recharged the three segments of the battery level indicator will be displayed fully.

Note: The charging function is not available at the battery contacts of the 4 x AA alkaline batteries.

Battery Installation (see figure 9):

1. Slide down the cover of the battery compartment on the rear panel of the device.
2. Insert four alkaline batteries (1.5V, AA), ensuring the correct orientation in accordance with the polarity markings.
3. Alternatively, insert the rechargeable Li-ion battery (Model No. CT-2500), orientated according to the guiding grooves.
4. Slide the battery-compartment cover back into its initial position to close.




Figure 9: Insert a Li-ion battery or 4 x AA alkali batteries

⚠ Warning: Only charge the Li-ion battery CT-2500 while inserted in the VM-2500 monitor using the provided power supply FW 7660M/06.

⚠ Caution: Do not dispose the Li-ion battery or the alkaline batteries in the consumer waste, if they are empty or can no longer be recharged. Batteries may contain substances which are harmful to the environment and to the health. Please dispose the batteries at the available battery collecting sites or the recycling yards of the municipalities. Please only dispose discharged batteries into the designated containers. Tape the terminals of the Li-ion batteries at the poles before disposing.


⚠ Caution: Remove the batteries if the device is to be stored or not used for a longer period of time.

5.2 Connecting Sensors / Sampling Line Configurations to the VM-2500

5.2.1 IRMA™ CO₂ Analyzer (mainstream)

For CO₂ mainstream capnography with VM-2500-M inspect the IRMA™ CO₂ analyzer and connector cables for any external damage.

Insert the connector of the IRMA™ CO₂ analyzer into the CO₂ Mainstream port located on the top edge of VM-2500-M. Secure the IRMA™ CO₂ analyzer on top of the IRMA™ airway adapter. It will click into place when correctly seated.



⚠ Warning: Do not use the IRMA™ airway adapter (adult/paediatric) with infants as the adapter adds 6ml dead space to the patient circuit.

⚠ Warning: Do not use the IRMA™ airway adapter (infant) with adults or paediatric as this may cause excessive flow resistance.

⚠ Caution: Use only PHASEIN manufactured IRMA™ airway adapters.

5.2.2 Nomo technology sampling line configuration (sidestream)

For CO₂ sidestream capnography with the VM-2500-S inspect the Nomo Adapter and the selected sampling line assembly for any external damage.

- Insert the Nomo Adapter head into the CO₂ sidestream inlet port (LEGI™) located on the top edge of the VM-2500-S.
- Now connect the selected sampling line assembly to the female Luer Lock connection at the Nomo Adapter. A typical sampling line assembly for intubated patients consists of a sampling line and a sidestream airway adapter. For non-intubated patients, a sampling line configuration with nasal and/or oral sampling is typical.




Figure 10: Nomo Adapter with sampling line configuration for intubated patients

⚠ Warning: The Nomo Adapter is reusable. It only needs to be replaced if occluded. Replace the Nomo Adapter if “Sampling system occlusion !” is displayed by the VM-2500-S whilst only the Nomo Adapter is connected.

⚠ Warning: Do not re-use disposable sampling lines and airway adapters.

⚠ Warning: Used disposable sampling lines and airway adapters shall be disposed of in accordance with local regulations for medical waste.

⚠ Caution: Use only Nomo Adapters manufactured by Viamed.

5.2.3 SpO₂ Sensor

Inspect the SpO₂ sensor and connector cables for any external damage.

Insert the SpO₂ sensor cable into the SpO₂ sensor port located on the top edge of the device, ensuring correct orientation of the sensor connector and the port.

5.3 Visual Check

Before commencing operation, ensure that the device, power supply, sensors or sampling line configurations are not damaged.

During the visual check of the VM-2500-S side-stream device, ensure that the gas outlet located at the rear part of the device is clear of any obstruction.

⚠ Warning: Do not use sensors, cables or lines that appear to be damaged. Do not use sensors when optical components are exposed. Do not use a device that appears damaged. Replace the monitor immediately in cases of visible damage.

⚠ Warning: Ensure that the speaker is clear of any obstruction and that the speaker holes are not covered. Failure to do so could result in an inaudible alarm tone.

⚠ Warning: Ensure that the gas outlet of the side-stream device is clear of any obstruction. Failure to do so could result in CO₂ sidestream measurement errors.

⚠ Warning: Check the compatibility of the monitor, probe and cable before use. Incompatible components can result in degraded performance.

5.4 Switching on the Device

Press and hold the ON/OFF button  briefly until an opening “welcome screen” appears. The power-on self-test is successfully completed after a single loud tone sounds.

5.5 Connecting Sensors / Sampling Line Configurations to the Patient


5.5.1 IRMA™ CO₂ Analyzer (mainstream)

A green LED indicates that the IRMA™ CO₂ Analyzer is powered and ready for use.


Perform the following tests prior to connecting the IRMA™ CO₂ Analyzer to the patient circuit:

1. Breathe into the airway adapter and check that valid CO₂ waveforms and values are displayed by the monitor.
2. Remove the airway adapter and wait for 5 seconds.
3. Check that the airway adapter alarm is displayed and that the LED at the IRMA™ CO₂ Analyzer shows a flashing red light.


Now connect the IRMA™ airway adapter to the patient circuit:



1. Connect the 15 mm male connector of the IRMA™ airway adapter to the breathing circuit Y-piece.



2. Connect the 15 mm female connector of the IRMA™ airway adapter to the endotracheal tube with or without an angled connector.




Alternatively, connect a HME (Heat Moisture Exchanger) between the patient's endo-tracheal tube and the IRMA™ CO₂ analyzer. Placing a HME in front of the IRMA™ CO₂


analyzer protects the airway adapter from secretions and effects of water vapour and eliminates the need of changing the adapter. It allows free positioning of the IRMA™ CO₂ analyzer as well.

3. Perform the tightness check of the patient circuit with the IRMA™ CO₂ analyzer connected on the airway adapter.

⚠ Warning: Do not place the IRMA™ airway adapter between the endotracheal tube and an elbow as this may allow patient secretions to block the adapter windows and result in incorrect operation.



⚠ Warning: To keep secretions and moisture from pooling on the windows, always position the IRMA™ CO₂ analyzer in a vertical position.



⚠ Warning: Replace the adapter if condensation occurs inside the airway adapter.

5.5.2 Nomo Technology sampling line configuration (sidestream)

A constant green light at the Light Emitting Gas Inlet (LEGI™) at the VM-2500-S indicates that the system is working correctly.

Perform the following tests prior to connecting the assembled sampling line configuration to the patient circuit:

1. Breathe into the sampling line configuration and check that valid CO₂ waveforms and values are displayed on the VM-2500-S.
2. Occlude the sampling line with a fingertip and wait for 10 seconds.
3. Check that an occlusion alarm is displayed and that the LED shows a flashing red light.

Now connect the selected sampling line configuration to the patient circuit:

A. In the case of intubated patients

The sidestream airway adapter is connected to the tubing of the patient circuit.

- First connect the male connector of the sidestream airway adapter to the breathing circuit Y-piece.
- Now connect the female connector of the sidestream airway adapter to the endotracheal tube with or without an angled connector.

B. In the case of non-intubated patients

The nasal or oral sampling configuration is directly connected to the nose or mouth respectively. For further instructions, warnings and precautions refer to the 'Instructions for Use' provided with the sampling system configuration.

Note: Our sampling line configurations are continuously updated; please contact Viamed for detailed information. Specific patient sampling line configurations and interfaces are available upon request.


⚠ Warning: Do not use adult or paediatric type sampling line configurations with infants or neonates, as this may add dead space to the patient circuit.

⚠ Warning: Do not use infant or neonatal type sampling line configurations with adults or paediatric, as this may result in excessive flow resistance.


⚠ Warning: Do not use the VM-2500-S with metered dose inhalers or nebulised medication as this may block the bacteria filter.

5.5.3 SpO₂ Sensor


1. Refer to the sensor 'Instructions for Use' to determine if an appropriate sensor is being used, and if it is applied correctly.



Adult



Paediatric



Neonatal

2. Confirm that all connections have been made correctly by verifying an actual SpO₂ waveform on the monitor display.

Warning: Avoid application of the SpO₂ sensors to oedematous or fragile tissue.

Warning: Do not use the SpO₂ sensor if it is damaged. Use of a damaged sensor could cause patient injury or equipment failure.

Warning: Excessive patient motion, excessive ambient light, electromagnetic interference, dysfunctional haemoglobin, low perfusion, intravascular dyes, finger nail polish and long or artificial finger nails may affect the sensor performance and the accuracy of the measurement.

Warning: Do not autoclave the SpO₂ sensor.

Warning: For the SpO₂ measurement, the monitor uses red and infrared light with specific fixed wavelengths. Consider that these wavelengths might influence diagnostic parameters of other optical applications. The specifications of the wavelengths used are listed in the 'Instructions for Use' of the specific sensor.

Caution: Depending on the sensor model, specifications can differ. For sensor specific information on wavelengths and radiant power, please refer to the 'Instructions for Use' provided with your SpO₂ Sensor.

Caution: For further instructions, warnings and precautions refer to the 'Instructions for Use' provided with the SpO₂ Sensor.

5.6 Commencing Monitoring

Once the sensors/sidestream sampling line configuration is connected and correctly positioned on the patient, monitoring begins automatically.

An audiovisual alarm appears, if any of the sensors or the Nomo Adapter is disconnected from the device.

The VM-2500 can be reset to the start-up configuration by resetting the alarms (refer to Chapter 7.6).


5.7 Switching off the Device

Press and hold the on/off button ① for approx. 3 seconds to switch off the device. The VM-2500 will also power off automatically after 5 minutes when not in use.


6 Display Modes and Displayed Data

6.1 Toggling Between Display Modes


The operator can toggle between various display modes by pressing the  button.



Display 1: Standard



Display 2: Numerical




Display 3 to 5: Trend
(15 min, 1 h and 6 h Trend, parallel to ongoing measurement)

The small numbers on the right hand side of the measurement value of a parameter indicate the upper and lower alarm limits. In all displays the measurement value, alarm limits, curve, name and unit of each parameter are displayed in the same colour.

- 1 End-tidal expired CO₂ gas concentration in vol %, kPa or mmHg
- 2 Inspired CO₂ gas concentration in vol %, kPa or mmHg
- 3 Respiration rate in breaths per minute
- 4 CO₂ waveform (Capnogram)
The default setting of the amplitude scale is Auto scaling. Here the reading is automatically adjusted to the signal strength; therefore, a waveform with strong amplitude should be visible at all times. The scale however can also be defined by the user (refer to chapter 8.1.3).
- 5 Functional blood oxygen saturation in units of percent SpO₂
- 6 Bar graph for pulse amplitude
Indicates the dynamic pulse amplitude and rate. As the detected pulse becomes stronger, more bars light with each pulse. The reverse is true for weak pulses.
- 7 Pulse rate in beats per minute
- 8 Pulse waveform (plethysmogram)
The reading is automatically adjusted to the pulse strength; therefore, a waveform with strong amplitude should be visible at all times.
- 9 Combined EtCO₂ and FiCO₂ trend waveform
- 10 SpO₂ trend waveform
- 11 Start and end times

6.2 Symbols and Indicators



No.	Symbol/Indicator	Function
1		Battery level indicator The three segments represent the battery charge level. The symbol flashes yellow when the battery capacity is low. During charging of the Li-Ion battery, the three segments of the indicator are illuminated in sequence.
2		Alarm silence indicator The audible alarm can be silenced for a maximum period of two minutes. Optical alarm remains activated.
3		Pulse tone off
4		Memory full indicator The device's memory for measurement data is full. No new data can be stored. Old data can be erased or are overwritten.
5	DEMO	Demo mode indicator The device is operated in demo mode. All measurement data are simulated. This mode is accessed via the service menu.

6		Real-Time Mode indicator The device is operated in the Real-Time mode. Data is available at the USB port
7		Neonatal mode The device is operated in the neonatal mode. On activation the alarm limits are changed to neonatal defaults.
8	17:35	Current time , displayed in 12h or 24h format.
9		Respiration rate , displayed in breaths per minute.
10		Pulse rate , displayed in beats per minute.
11		The colour of the bar graph is an indicator of the signal quality Green : good signal quality, very accurate measurement Yellow : average signal quality, measurement may be inaccurate Red : poor signal quality, unreliable measurement

6.3 Pulse Tone

During monitoring a pulse beep is sounded for every detected pulse. The pitch of the pulse tone is dependent on the measured SpO₂ value. A higher pitch is indicative of higher oxygen saturation.

The pulse tone volume can be adjusted under the menu item 'Volume'. The pulse tone can be also silenced using the button. Pressing this button for a second time will reactivate the pulse tone.

7 Alarms

7.1 Alarm Priority and Appearance

The VM-2500 differentiates between alarms of high, medium and low priority.

An alarm of higher priority will always overlay alarms of respective lower priority. Vice versa, an alarm of high priority cannot be silenced by an alarm of lower priority. In this way the VM-2500 will always display the alarm with the highest priority, if more than one alarm condition exists at the same time.

7.2 Audible Alarm Volume

The alarm volume is not adjustable; however, it is possible to silence the alarm for a period of two minutes using the  button. It has an audibility of at least 55 dB (A) in 1 meter distance of the device.

7.3 Default Alarm Limits

Default settings are active according to the selected device mode “Adult” or “Neonatal”.

Alarm Priority and Appearance

Priority Level	Audio Signal Sequence (repeatedly)	Visual Alarm	Condition
High (Warning)	5 tone beep + 2 seconds pause + 5 tone beep and 3 seconds pause	 Red and !!!	For potentially life-threatening situations
Medium (Caution)	3 tone beep and 5 seconds pause	 Yellow and !!	For potentially serious problems that are presumed to be not life threatening
Low (Advisory)	2 tone beep and 16 seconds pause	 Yellow and !	Advisory alarms

Default Alarm Limits

Limit	Unit	Range	“Adult”	“Neonatal”
<i>EtCO₂ High</i>	%	0.1...9.9 / off	7.3	7.3
<i>EtCO₂ Low</i>	%	off / 0.1...9.9	3.2	3.2
<i>FiCO₂ High</i>	%	0.1...9.9 / off	1.2	1.2
<i>RR High</i>	/min	4...150 / off	off	off
<i>Apnoea</i>	sec	20, 40, 60	20	20
<i>SpO₂ High</i>	%	1...99 / off	off	95
<i>SpO₂ Low</i>	%	off / 1...99	85	85
<i>PR High</i>	/min	1...250 / off	140	150
<i>PR Low</i>	/min	off / 1...250	45	30

7.4 Limit Alarm

The table below shows the limit alarms, alarm conditions and priority of the VM-2500. Depending on the priority of the alarm the respective measurement value will change colour and an audible alarm sounds.

Limit Alarm	Condition	Priority
<i>EtCO₂ High</i>	EtCO ₂ value above set alarm limit	Medium
<i>EtCO₂ Low</i>	EtCO ₂ value below set alarm limit	Medium
<i>FiCO₂ High</i>	FiCO ₂ value above set alarm limit	Medium
<i>RR High</i>	Respiratory rate above set alarm limit	Medium
<i>Apnoea !!</i>	No breath detected within set terms (if $\leq$ 1min)	Medium
<i>Apnoea !!!</i>	No breath detected within set terms (if $>$ 1min)	High
<i>SpO₂ High</i>	SpO ₂ value above set alarm limit in the standard mode	Low
<i>SpO₂ High (neonatal mode)</i>	SpO ₂ value above set alarm limit in the neonatal mode	Medium
<i>SpO₂ Low</i>	SpO ₂ value below set alarm limit	Medium
<i>PR High</i>	Pulse rate above set alarm limit	Medium
<i>PR Low</i>	Pulse rate below set alarm limit	Medium

7.5 Alarm Messages

The table below presents the alarm messages, alarm condition and priority of the VM-2500. The colour of the alarm message and its audible alarm depend on the priority of the alarm message.


Alarm Messages	Condition	Priority
<i>Bad signal quality !!!</i>	Poor-quality pulse signal, for example as a result of low perfusion	High
<i>No CO₂ Sensor !</i>	No connection to the mainstream IRMA™ CO ₂ analyzer	Low
<i>Check CO₂ Adapter !</i>	The mainstream IRMA™ adaptor is dirty, not positioned correctly etc.	Low
<i>No Nomo Adapter !</i>	No connection to the sidestream Nomo Adapter.	Low
<i>Sampling system occlusion !</i>	A component of the sidestream sampling system (sampling line, Nomo Adapter or gas outlet) is blocked (occluded)	Low
<i>CO₂ Over range !</i>	Measured CO ₂ concentration is outside of the specified accuracy range	Low
<i>No SpO₂ Sensor !</i>	No connection to the SpO ₂ sensor	Low
<i>SpO₂ Probe off !</i>	SpO ₂ sensor is connected to the device, a signal was detected and then the finger has been removed/slipped off	Low
<i>Excess light !</i>	High ambient light sources near the SpO ₂ sensor, e.g. surgical lights	Low
<i>Battery low !</i>	Low battery level at start up results in error message and dysfunction. Critical level during monitoring results in audible alarm and battery indicator to flash yellow. After 3 minutes device switches off.	Low

⚠ Warning: The monitor detects respiratory effort via changes in CO_2 concentration of exhaled air; therefore, the CO_2 measurement can be used to detect apnoea. The device however is unable to discriminate between a patient not breathing and a sensor that is disconnected from the patient circuit. Always monitor and set alarms for SpO_2 when using the VM-2500 to monitor respiratory function.

7.6 Resetting of Alarm Signals

Once triggered, an alarm will only be reset if the cause of the alarm has been resolved. Individual alarm limits can also be completely deactivated if required.

Alarm signals can be reset by pressing and holding the  button for 3 seconds. If the initial condition for the warning is still present after resetting the warning signals, the warning will return immediately. In case of the alarm signals “ SpO_2 Probe off!”, “No Nomo Adapter” and “No CO_2 Sensor!” the device reverts to the on-position. Parameters which have been set by the user will remain once an alarm is reset.



Press and hold button 


Display after reset of alarms

8 Menu Structure

8.1 Main Menu

Navigating the Menu

Use the  buttons to scroll through menu items. The currently selected menu item is highlighted by a coloured frame. Press the  button to confirm your selection. Select the menu item ‘Back’ to return to the previous menu level. Alternatively use the menu button  as shortcut for “Back”.



All important and frequently used settings are accessible through the main menu, which can be accessed by pressing the  button.

Entering Data

In some submenus it is possible to adjust a certain parameter. In this case the parameter can be increased or decreased using the  buttons. The value will increase or decrease more quickly when the respective button is held down. Press the  button to confirm the new value.

Exiting Menu and Returning to Display

Press the display button  at any time, in any menu to return immediately to the monitoring display. If no button has been pressed for more than 30 seconds, the device will automatically return to the monitoring screen.

8.1.1 Submenu: Alarm Settings


General Information

The VM-2500 alarm limits for EtCO_2 , FiCO_2 , SpO_2 , respiration rate and pulse rate can be set individually. The current alarm limits are displayed as small numbers on the right hand side of the measurement value. If a measured value either exceeds the upper limit or falls below the lower limit, visual and audible alarms will be triggered immediately.

Visual alarm

When an alarm has been triggered the critical value will flash and turn yellow together with the violated alarm limit.

An alarm will also be triggered if the SpO₂ sensor is removed from the patient, the SpO₂ signal quality remains poor over a longer period of time, the mainstream IRMA™ airway adapter is dirty or the side-stream Nomo Adapter is occluded. This is also the case if the SpO₂ sensor, the IRMA™ CO₂ analyzer or the Nomo Adapter is disconnected from the device, provided that valid measurement data has been recorded beforehand.



Visual alarm which was triggered by a violation of the upper respiration rate alarm limit

Selection with ▲▼ buttons / confirmation with ← button

Adjusting Settings

Use the Alarm Settings menu to set the upper and lower alarm limit of EtCO₂, SpO₂, respiration rate, pulse rate and FiCO₂. Select “-” to deactivate the respective alarm limit. “Apnoea” represents the lower alarm limit of the respiration rate. If no breath is detected within the set time, the apnoea alarm is activated. After restarting the device, the default alarm limits will be reset.

Use the Alarm settings menu to set the upper and lower alarm limit of EtCO₂, SpO₂, respiration rate, pulse rate and FiCO₂. Select “off” to deactivate the respective alarm limit.

“Apnoea” represents the lower alarm limit of the respiration rate. If no breath is detected within the set time, the apnoea alarm is activated. After restarting the device, the default alarm limits will be reset. This is also the case, if the power supply is interrupted.

8.1.2 Submenu: Data Management

8.1.2.1 Recording Data


The VM-2500 device can store more than 400 hours of monitoring data. Each individual data set, regardless of its actual length, uses at least 80 minutes of

memory space. A new data set is generated automatically each time the device is turned on or the current Patient ID is changed.

When the device is turned off or the current Patient ID is changed, all of the measurements that were taken are automatically stored in the devices memory, together with the respective alarm limits, patient ID, date and time.

The device warns the user when the memory is almost full by displaying the ⇄ symbol. If the memory is full the oldest data set is overwritten upon confirmation by the user. Stored data sets can be retrieved and erased under the menu item “Data management”. The data sets can also be downloaded on PC with the user-friendly VM-2500 PC-Software.

8.1.2.2 Data Management



Use the Data Management menu to access the list of stored data sets, to delete all data in the memory or to view the remaining recording time.

Retrieve the selected data set by pressing the ← button.




The stored measurements are displayed in graphic form together with the date, start time, duration of the recording and Patient ID. The colour code of the displayed data is as follows: EtCO₂ – white, SpO₂ - green, respiration rate – blue, pulse rate - orange.

Select “Back” to return to the list of stored data or “Delete” to erase the data set.


8.1.3 Submenu: Setup

8.1.3.1 General Information




This submenu offers access to various device settings; confirm selection by pressing the  button.

8.1.3.2 Parameters



Volume


Adjust the pulse tone volume using the ▲▼ buttons. Confirm the new setting by pressing the ← button.



Brightness

Adjust the display brightness using the ▲▼ buttons. Confirm the new setting by pressing the ← button.

Note: Very high brightness settings will shorten the battery life considerably!




SpO₂ Averaging

Stable: When this setting is selected any strong and sudden variations in data will not immediately affect the reading (data incorporated over time); minor irregularities have little or no effect on the displayed reading.

Standard: Averaging parameters used for this setting are between those of the stable and sensitive settings.


Sensitive: The reading is more sensitive to irregularities but reacts very quickly to any changes in measured parameters.

Refer to Chapter 13 “Technical Specifications” for further details on the influence of the SpO₂ averaging settings on the SpO₂ reaction time.




Neonatal Mode

The neonatal symbol  indicates that the neonatal mode is activated. In this mode the default alarm limits are adjusted to neonatal default settings (refer to Chapter 7.3). Please ensure that accessories appropriate for neonates are connected.



Date and Time

First, select between Y/M/D and D/M/Y mode (respectively 12h mode and 24h mode), then set the date and time. Settings for the date and time are not erased when the batteries are temporarily removed.




CO₂ Unit

By changing the CO₂ unit, the measurement values and default limits of EtCO₂ and FiCO₂ are converted accordingly. For conversion to the units kPa and mmHg an automatic barometric pressure compensation is performed.




Language

All messages and menus will be displayed in the selected language. The standard language package comprises up to 16 languages. Please refer to the manufacturer for detailed information on the current languages available.




CO₂ Scale

The scale maximum of the capnogram and CO₂ trend can be fixed to 6 %, 10 % or 15 %. Select the option “Auto Scaling” for optimal amplitude scaling of the data. Depending on the selected CO₂ scale, the values will be adjusted accordingly.



CO₂ Time Base


For an optimal time resolution the time base of the capnogram can be adjusted to 15 seconds or 30 seconds.



Real-Time Mode

The Real-Time mode symbol  indicates that the Real-Time mode is activated. Activation of the Real-Time mode enables visualization and storage of measurement data on a PC.

In this mode the ongoing measurement values of EtCO₂, FiCO₂, SpO₂, respiration rate and pulse rate are available every 4 seconds at the USB port for download to the PC. For more information please refer to the enclosed Software Manual.




Gas Compensation

The presence of oxygen and nitrous oxide can cause some interference in the CO₂ measurement. These interferences are compensated by setting the range of O₂ and N₂O concentration under the menu point “Gas compensation” accordingly.

When the O₂/N₂O concentrations are set correctly, the measurement is within the accuracy range stated in the Technical Specifications (refer to Chapter 13).

Note: Only the mainstream VM-2500-M may be used to analyse gasses which contain nitrous oxide. Be sure that the IRMA™ airway adapter is correctly connected to the patient circuit.


⚠ Warning: Do not use VM-2500-S to analyse gasses (e.g. anaesthetics like nitrous oxide etc.) that need to be returned to the patient circuit or a scavenging system. The gas outlet of the sidestream device is not designed to return the exhaust gases to the patient circuit or a scavenging system.



Zeroing

Zeroing needs to be performed only when an offset in EtCO₂ and FiCO₂ values is observed, or when an unspecified accuracy message is displayed.


⚠ Warning: Incorrect zeroing will result in false CO₂ gas readings. Refer to chapter 11.4 for detailed information.



Service

The Service submenu is protected by a PIN code; only authorized service personnel can access this menu.

8.1.4 Submenu: Patient ID



The VM-2500 features the possibility of saving a 12 digit patient ID together with every created data file. If the Patient ID is changed via the menu during measurement, the current data file is closed and a new data file with the new Patient ID is opened. A message is displayed to inform the user.

8.1.5 Default Start Settings

Changed settings are in effect only as long as the VM-2500 remains switched on. Once the VM-2500 has been switched off, at the next start up, the default settings will be in effect. The start up defaults can be changed in the PIN protected Service Menu. Only authorized service personnel can access this menu.

8.2 Other

8.2.1 Volume Control Shortcut

If the ▲ button is pressed during any monitoring display mode, the volume control screen will open. Adjust the volume using the ▲▼ buttons. Confirm the new setting by pressing the ◀ button.

8.2.2 Brightness Control Shortcut

If the ▼ button is pressed during any monitoring display mode, the brightness control screen will open. Adjust the brightness using the ▲▼ buttons. Confirm the new setting by pressing the ◀ button.

8.2.3 Power-Save Mode



The device's display can be turned off to save power and extend battery life. This can be accomplished by pressing and holding the ▼ button. A countdown will start, after which the display will be switched off. The device is now in economy power mode. The pressing of any button will reactivate the display.

9 Adverse affects on performance

9.1 CO₂ measurements

9.1.1 Humidity or Condensation

The partial pressure and the volume percentage of CO₂ will depend on the amount of water vapour in the measured gas.

Both the mainstream and sidestream devices (VM-2500-M and VM-2500-S) will always show the actual CO₂ partial pressure at the current humidity level.

For sidestream sampling (VM-2500-S) the following has to be considered: In the alveoli of the patient, the breathing gas is saturated with water vapour at body temperature (BTPS). When the breathing gas is sampled, and passing the sampling line, the gas temperature will almost reach the ambient temperature before reaching the VM-2500-S monitor. As the Nomo Adapter removes all condensed water, no water will reach the measurement unit. The relative humidity of the sampled gas will be approximately 95%.

If CO₂ values at BTPS are required, the following equation can be used:

$$EtCO_2 (BTPS) = EtCO_2 \times (1 - (3.8/P_{amb}))$$

With:

- $EtCO_2$ = $EtCO_2$ value displayed by VM-2500-S [vol%]
- P_{amb} = Ambient pressure
- 3.8 = Typical partial pressure of water vapour condensed between the patient circuit and the VM-2500-S [kPa]
- $EtCO_2 (BTPS)$ = $EtCO_2$ gas concentration at BTPS [vol%]

9.1.2 Interfering gases or vapours

Gas or Vapour	Gas Level	CO ₂ Measurement Interference
N ₂ O ⁴⁾	60 vol%	- ²⁾
HAL ⁴⁾	4 vol%	- ¹⁾
ENF, ISO, SEV ⁴⁾	5 vol%	+8% of reading ³⁾
DES ⁴⁾	15 vol%	+12% of reading ³⁾
Xe (Xenon)	80 vol%	-10% of reading ³⁾
He (Helium)	50 vol%	-6% of reading ³⁾
Metered dose inhaler propellants ⁴⁾	Not for use with metered dose inhalers propellants	
C ₂ H ₅ OH (Ethanol) ⁴⁾	0.3 vol%	- ¹⁾
C ₃ H ₇ OH (Iso-propanol) ⁴⁾	0.5 vol%	- ¹⁾
CH ₃ COCH ₃ (Acetone) ⁴⁾	1 vol%	- ¹⁾
CH ₄ (Methane) ⁴⁾	3 vol%	- ¹⁾
CO (Carbon monoxide) ⁵⁾	1 vol%	- ¹⁾
NO (Nitrogen monoxide) ⁵⁾	0.02 vol%	- ¹⁾
O ₂ ⁵⁾	100 vol%	- ²⁾

Note 1: Negligible interference, effect included in the specifications “accuracy, incl. interfering gases” (refer to Chapter 13)

Note 2: Negligible interference with N₂O/O₂ concentrations correctly set, effect included in the specifications “accuracy, incl. interfering gases” (refer to Chapter 13)

Note 3: Interference at indicated gas level

Note 4: According to the EN ISO 21647:2004 standard

Note 5: In addition to the EN ISO 21647:2004 standard

9.2 SpO₂ Measurement

Physiological conditions, medical procedures, or external agents that may interfere with the monitor's ability to detect and display accurate SpO₂ measurements include:

- Incorrect application of the SpO₂ sensor
- Placement of the SpO₂ sensor on an extremity with a blood pressure cuff, arterial catheter, or intravascular line
- Excessive patient activity
- Intravascular dyes
- Externally applied colouring agents, such as nail polish
- Failure to cover the sensor site with opaque material in high ambient light conditions
- Venous pulsation
- Dysfunctional haemoglobin, e.g. caused by carbon monoxide intoxication
- Low perfusion

10 Troubleshooting guide

10.1 Error Message – Cause – Corrective Action

Error Message	Cause	Corrective Action
<i>Battery low !</i>	Low battery level at start up results in error message and dysfunction. Critical level during monitoring results in audible alarm and battery indicator to flash yellow. After 3 minutes device switches off.	Replace batteries immediately or connect the external power supply to the AC line.
<i>Device defective !</i>	Defective hardware, e.g. PCB.	Send the device to qualified service personnel.
<i>No SpO₂ Sensor !</i>	No connection to the SpO ₂ sensor.	Check sensor connection.
<i>SpO₂ Probe off !</i>	The SpO ₂ sensor is connected to the device, a signal was detected and then the finger has been removed/slipped of.	Check that the sensor is correctly attached to the patient.
<i>SpO₂ Sensor fault !</i>	The connected SpO ₂ sensor is either defective or not compatible with the device.	Replace sensor.
<i>Excess light !</i>	High ambient light sources near the SpO ₂ sensor, e.g. surgical lights.	Shield sensor more effectively from external light.
<i>Bad signal quality !!!</i>	Poor-quality pulse signal, for example as a result of low perfusion.	Check other vital signs to ensure sufficient perfusion. Move the sensor to a different site on the patient or provide more effective monitoring conditions.
<i>CO₂ Over range !</i>	The measured CO ₂ concentration is outside of the specified accuracy range.	Ensure that device is used in an environment with a CO ₂ concentration within the specified accuracy range.
<i>No Nomo Adapter !</i>	No connection to the sidestream Nomo Adapter.	Check Nomo Adapter connection.
<i>Sampling system occlusion !</i>	A component of the sidestream sampling system (sampling line, Nomo Adapter or gas outlet) is blocked (occluded).	Remove the occlusion by e.g. replacing the sampling line or Nomo Adapter.
<i>No CO₂ Sensor !</i>	No connection to the mainstream IRMA™ CO ₂ analyzer.	Check connection.
<i>Check CO₂ Adapter !</i>	The mainstream IRMA™ adaptor is dirty, not positioned correctly etc.	Check airway adapter and replace if required
<i>CO₂ Sensor fault !</i>	The mainstream IRMA™ CO ₂ analyzer is defective.	Replace IRMA™ CO ₂ analyzer.
<i>Data memory full, overwrite? Yes/No</i>	Message is displayed at every startup of device if the memory is full.	Delete files or overwrite.
<i>Zeroing disabled</i>	The IRMA™ CO ₂ analyzer / Nomo Adapter is not connected.	Connect the IRMA™ CO ₂ analyzer / Nomo Adapter and try again.
	The CO ₂ measurement module has not reached its working temperature after power on or after changing the IRMA™ airway adapter / Nomo Adapter	Wait 10 seconds for warm up of the IRMA™ CO ₂ analyzer / ISA module before proceeding with the zeroing procedure.
	Other causes	Send the device to qualified service personnel.

10.2 Failure – Cause – Corrective Action

Failure	Cause	Corrective Action
<i>No response to the power button</i>	Power button is not fully depressed.	Ensure that the power button is fully depressed.
	The batteries may be missing, discharged, or oriented incorrectly or no power supply is connected.	Install new batteries or connect the power supply to the AC line.
<i>Audible medium priority alarm sounds while the device is off and can not be restarted again</i>	Power supply is disconnected during operation while no batteries are installed. Medium priority audible alarm occurs for 2 minutes whereupon device switches off.	Reconnect power supply immediately or insert battery.
<i>No pulse signal found or the pulse signal cannot be found anymore</i>	Patient has no pulse signal.	Check the patient.
	The incorrect SpO ₂ sensor is used.	Check the sensor 'Instructions for Use' to determine if an appropriate sensor is being used and if it is applied correctly.
	SpO ₂ sensor or extension cable is defective.	Check the sensor and extension cable connections. Test the sensor on another subject. Try another sensor or extension cable.
	Perfusion may be too low for the monitor to track the pulse.	Check the patient. Test the monitor on yourself. Change the sensor site. Try another sensor.
	Interference due to patient activity may be preventing the monitor from tracking the pulse.	Keep the patient still, if possible. Verify that the sensor is securely applied and replace it if necessary. Change the sensor site.
	The sensor may be too tight, there may be interference due to ambient light, or the sensor may be on an extremity with a blood pressure cuff, arterial catheter, or intravascular line.	Reposition sensor, as necessary.
	Electromagnetic interference may be preventing the monitor from tracking the pulse.	Remove the source of interference.
<i>No pulse tone</i>	Pulse beep volume is off.	Switch volume on.
	Speaker/audio has malfunctioned. Signal is corrupted. The VM-2500 has stopped functioning.	Contact qualified service personnel.
<i>EtCO₂ values are inconsistent</i>	Physiological reasons.	Check the patient.
	Leakage in the system.	Check the tubes and connections to the patient
<i>EtCO₂ values are continuously higher or lower than expected.</i>	Physiological reasons.	Check the patient.
	Zeroing or calibration required.	Contact qualified service personnel.

10.3 IRMA™ LED Status and LEGI™ Status

Indication	Status	Action
● <i>Steady green light</i>	System OK	-
⦿ <i>Blinking green light</i>	Zeroing in progress	-
● <i>Steady red light</i>	System Error	check device error message
⦿ <i>Blinking red light at IRMA™ LED</i>	IRMA™ adapter defective	Check IRMA™ adapter
⦿ <i>Blinking red light at LEGI™</i>	Sampling system defective	Check sampling system

10.4 Problems with EMI (Electromagnetic Interference)

The VM-2500 has been tested and found to comply with the limits for medical devices according to: EN 60601-1-2, (second edition) and the Medical Device Directive 93/42/EEC.

These limits are designed to provide reasonable protection against harmful interference in a typical medical installation.

Due to the proliferation of radio-frequency transmitting equipment and other sources of electrical noise in healthcare environments it is possible that high levels of such interference due to close proximity or strength of a source may result in disruption of performance of this device. Examples of noise sources in healthcare environments that could cause electromagnetic interference include:

- Electrosurgical units
- Cellular phones
- Mobile two-way radios
- Electrical appliances
- High-definition televisions (HDTVs)

The measurement values of VM-2500 can be obscured by electromagnetic interference. During such interference measurements may seem inappropriate or the monitor may not seem to operate correctly.

Disruption may be evidenced by erratic readings, cessation of operation or other incorrect functioning.

If this occurs, the operating environment should be surveyed to determine the source of disruption and the following actions taken to eliminate the source:

- Turn equipment in the vicinity off and on to isolate the offending equipment.
- Reposition or relocate the interfering equipment.
- Increase the distance between the interfering equipment and this equipment.

The VM-2500 will generate, use and radiate radio frequency energy. Failure to follow these instructions may cause harmful interference with other devices in the vicinity.

11 Maintenance

11.1 Maintenance

The monitor requires no routine calibration. If service is necessary, contact qualified service personnel or your local sales representative.

 *Caution: There are no user-serviceable parts within the VM-2500. The cover should only be removed by qualified service personnel.*

 *Caution: The VM-2500 requires no routine calibration. A basic maintenance plan conducted by qualified service personnel is recommended. Please refer to the Service Manual for detailed information.*

11.2 Cleaning

Surface-clean

The VM-2500 and its accessories should be cleaned on a regular basis. Use a soft cloth dampened with either a commercial, nonabrasive cleaner, or a solution of 70% alcohol in water to clean the device. Lightly wipe the surface of the monitor.

To prevent cleaning liquids and dust from entering the VM-2500-S sidestream device through its LEGI™ connector, keep the Nomo Adapter connected while cleaning the monitor.

 *Caution: Do not immerse the VM-2500 or its accessories in liquid.*

 *Caution: Do not spray, pour, or spill any liquid on the VM-2500, its accessories, connectors, switches, or openings in the enclosure as this may result in damage to the unit.*

Disinfection

Use a soft cloth saturated with a solution of 10% chlorine bleach in tap water to disinfect the device housing. Use 70% alcohol to disinfect the IRMA CO₂ Analyzer and the SpO₂ sensor.

 **Caution:** Do not autoclave or steam sterilize the VM-2500 device, or its disposable accessories.

 **Caution:** Do not autoclave or steam sterilize the IRMA CO₂ Analyzer or the Nomo Adapter and sampling configuration.

11.3 Testing

Test of the alarm system

In order to trigger an alarm for test purposes during monitoring set the upper alarm limit of SpO₂ or pulse rate below the currently indicated measurement value. The device will react with a visual and audible alarm.

Test of the measurement accuracy SpO₂

The only reliable method of testing the SpO₂ measurement accuracy of the monitor is the clinical validation of the measurement data, indicated by the system monitor with SpO₂ sensor on the basis of a blood gas analysis. During extensive clinical studies, the monitor combined with the approved sensors evidenced the accuracy required. Verify SpO₂ readings at regular intervals with a reference instrument.

Note that a functional tester cannot be used to assess the accuracy of a pulse oximeter probe or a pulse oximeter monitor.

Test of the measurement accuracy CO₂

The CO₂ units of the VM-2500 (sidestream and mainstream) are permanently factory calibrated. Reliable testing of the measurement accuracy of the CO₂ measurement requires the use of a suitable calibration gas mixture. During extensive measurements with calibration gas, the CO₂ measurement units (sidestream and mainstream) evidenced the accuracy required. Verify CO₂ gas readings at regular intervals with a reference instrument.

11.4 Zeroing the Capnograph

11.4.1 Mainstream VM-2500-M

Zeroing needs to be performed only when an offset in gas values is observed, or when an unspecified accuracy message is displayed:

- First connect a new IRMA™ airway adapter onto the IRMA™ CO₂ analyzer, without connecting the airway adapter to the patient circuit.
- Allow 10 seconds for warm up of the IRMA™ CO₂ analyzer after power on or after changing the IRMA™ airway adapter before proceeding with the zeroing procedure.
- Select MAIN MENU > SETUP > ZEROING at the monitor.
- Ensure that ambient air (21% O₂ and 0% CO₂) is present in the IRMA™ airway adapter.
- Start the zeroing by selecting “Yes” when the message “Start Zeroing?” is displayed.
- The message “Zeroing completed!” indicates that the zeroing was successful.

Special care should be taken to avoid breathing near the airway adapter before or during the zeroing procedure. Always perform a pre-use check after zeroing the IRMA™ CO₂ analyzer.

 **Warning:** A successful zeroing requires the presence of ambient air (21% O₂ and 0% CO₂) in the IRMA™ airway adapter during zeroing. Incorrect zeroing of the IRMA™ CO₂ analyzer will result in false gas readings.

11.4.2 Sidestream VM-2500-S

The VM-2500-S performs zeroing automatically by switching the gas sampling from the respiratory circuit to ambient air. The automatic zeroing is performed every 24 hours, and takes less than 3 seconds.

Next to the automatic zeroing a manual induced zeroing is also possible, however needs to be performed only when an offset in gas values is observed, or when an unspecified accuracy message is displayed:

- First connect a Nomo Adapter to the VM-2500-S.
- Allow 10 seconds for warm up of the ISA™ module after power on or after changing the Nomo Adapter before proceeding with the zeroing procedure.
- Select MAIN MENU > SETUP > ZEROING at the monitor.
- Ensure that the monitor is placed in a well ventilated environment. Avoid breathing near the monitor housing.
- Start the zeroing by selecting “Yes” when the message “Start Zeroing?” is displayed.
- The message “Zeroing completed!” indicates that the zeroing was successful.

 **Warning:** Since a successful zeroing requires the presence of ambient air (21% O₂ and 0% CO₂) in the device, ensure that the VM-2500-S is placed in a well ventilated environment. Avoid breathing near the VM-2500-S before or during the zeroing procedure. Incorrect zeroing of the integrated ISA™ CO₂ analyzer will result in false gas readings.

12 VM-2500 PC-Software

With the user-friendly VM-2500 PC-Software all measurement data, selected alarm limits and occurred alarm can be stored on a PC via the USB interface. Here the data can be viewed and patient data added. The file can be printed or exported as CSV file for processing with additional software.

Furthermore the software can be used to display and save measurement values and alarm messages on a PC, parallel to ongoing measurements.

To enable this function the Real-Time mode has to be selected on the device. During this mode the device sends the current EtCO₂, FiCO₂, SpO₂, respiration rate and pulse rate measurement values every 4 seconds via USB to the PC.

For more information please refer to the enclosed Software Manual.

13 Technical Specifications

GENERAL

Parameters displayed

- Numerical: End-tidal CO₂ concentration (EtCO₂), Inspired CO₂ concentration (FiCO₂), Oxygen saturation (SpO₂), respiration rate (RR), pulse rate (PR)
- Graphical: Capnogram, Plethysmogram, trends of numerical data (15min/1h/6h)

Indicators

Signal strength and signal quality, pulse amplitude, battery status, alarm mute, pulse tone mute, storage status, Real-Time mode, neonatal mode, time

Alarms

- Limits: Adjustable limits for all numerical parameters

- Alerts: Audible and visual alarms (complies with EN 60601-1-8)

Storing Data

- Communication interface: USB 2.0
- Data memory on device: up to 400 hours in total
- Real-Time mode: Visualise and save numerical parameters every 4 seconds on PC
- PC-Software: VM-2500 PC-Software (for data download and Real-Time mode)

CAPNOGRAPH

Warm up time

< 10 seconds (concentrations reported and full accuracy)

CAPNOGRAPH	Mainstream	Sidestream
Operation principle	State-of-the-art, single path, non-dispersive infrared (NDIR), mainstream sensor	Ultra-compact infrared spectrometer with integrated low-flow pump
Barometric pressure compensation	The total pressure of the gas mixture is estimated by measuring the actual atmospheric pressure in the IRMA™ analyzer. For conversion to the units kPa and mmHg an automatic barometric pressure compensation is performed.	The total pressure of the gas mixture is measured by a pressure sensor in the sidestream CO ₂ analyzer. For conversion to the units kPa and mmHg an automatic barometric pressure compensation is performed.
Flow rate	–	50 ± 10 ml/min
Calibration	No span calibration required. Manual zeroing required when offset in gas values is observed.	No calibration required Automatic zeroing is performed at start-up and then every 24 hours.
Rise time	≤ 90 ms (at 10 l/min)	≤ 200ms (at 50 ml/min sample flow)
Total system response time	< 1 second	< 3 second (with 2 m sampling line)
Airway adapter	- Disposable Adult/Paediatric: < 6 ml dead space, < 0.3 cm H₂O pressure drop at 30 l/min - Disposable Infant: < 1 ml dead space, < 1.3 cm H₂O pressure drop at 10 l/min	Disposable Adult/Paediatric: < 6 ml dead space, < 0.3 cm H ₂ O pressure drop at 30 l/min
Sampling interface	–	Nomo Adapter with female Luer Lock connection
Length of interface	Cable length mainstream sensor: 2.55 m	Sampling line length: 2 m
Water handling	XTP™ windows of the IRMA™ airway adapter with special features that prevent a decrease in performance when vapour is present.	Nomo Adapter with proprietary water removal tubing.

Measurement Range

- EtCO₂ and FiCO₂: 0 - 15%
- Respiration rate: 0 - 150 breaths/min

Accuracy

- EtCO₂ and FiCO₂: +/- (0.2 vol%, + 2% of reading), +/- (0.3 vol% + 4% of reading) incl. interfering gases
- Respiration rate: +/-1 digit

Parameter renews

- EtCO₂ and FiCO₂: Displayed after one breath and then a continually updated breath average.
- Respiration rate: Displayed after three breaths and then average value updated every breath.

PULSE OXIMETER

Measurement Range

- SpO₂: 0 - 100%
- Pulse Rate: 20 - 300 beats/min

Accuracy

- SpO₂¹⁾: +/- 2% (70 to 100%)
- Pulse Rate: +/- 1 digit (up to 100 beats/min) or +/- 1% (> 100 beats/min)

Parameter renews (see Table)

First displayed value after application:

- SpO₂: Between 3 seconds and 7 seconds, depending on measurement conditions.
- Pulse rate: Between 5 seconds and 8 seconds, depending on measurement conditions.

PHYSICAL CHARACTERISTICS

Display

Active OLED colour graphic display,
262 000 colours, 240 x 320 pixels

Dimensions

(L x W x H): 15 cm x 7.5 cm x 3.5 cm

Weight

< 400 g (complete device with batteries)

POWER SUPPLY

Power may be supplied by battery, rechargeable battery or by AC power supply

4 x Batteries

Alkaline battery (AA / LR6 / AM3 / MN1500 / Mignon), 1.5V

Li-ion battery, Model No. CT-2500

Li-ion battery, 3.7V, 2500mAh, charging time and working time with full function: 5-6 hours

AC power supply, Model No. FW 7660M/06

Medical power supply with option for country-specific input plug

- Input: 100-240V AC / 50-60Hz / 250mA
- Output: 6V DC / 1.4 A

PULSE OXIMETER Parameter renews

Measurement dynamics		Beat to beat min/max	Sensitive min/max	Standard min/max	Stable min/max
SpO ₂ ²⁾	First reaction after	N/A	1 sec	2 sec	4 sec
	Determined value reached after another	N/A	4 sec	8 sec	12 sec
Pulse-rate ³⁾	First reaction after	1 sec / 7 sec	1 sec / 7 sec	1 sec / 7 sec	1 sec / 7 sec
	Determined value reached after another	N/A	1 sec / 4 sec	1 sec / 6 sec	1 sec / 8 sec

1) As inherent to their functional principle, pulse oximetry measurements underlie statistical spread, therefore only two thirds of the measurement data are within the specific range of +/- ARMS

2) Measured at de-saturation / re-saturation between 96 % and 84 % SpO₂ under favourable measurement conditions. The values can be extended by a bad pulsation strength or motion artefacts.

3) Maximum values are measured with sudden change from 40 to 200 beats/minute and vice versa. The reaction depends on the difference (variance) of the beats among themselves.

ENVIRONMENTAL CONDITIONS

Operating conditions

0 - 50°C, 15 - 95% R.H. (non-condensing),
60 - 120 kPa (excl. Li-ion battery ¹⁾)

Storage conditions

-30 - 70°C, 10 - 95% R.H. (non-condensing),
60 - 120 kPa (excl. Li-ion battery ²⁾)

CLASSIFICATION

General

- The device is not suitable for use in the presence of a flammable anaesthetic mixture with air or with oxygen or nitrous oxide.
- No sterile parts are included.
- Mode of operation: Continuous operation

Construction

Water-resistant construction of class IPX1 (with silicone cover)

Classification (according to MDD 93/42/EEC)

Class IIb

Electrical safety

Class of protection II / Type BF – Type and degree of protection against electrical shock

APPLIED STANDARDS

The applied standards are listed in the directory COMPLIANCE on the CD-ROM provided with the device.

14 Component Lists and Ordering

14.1 Packing List

VM-2500-M:

- Mainstream device
- IRMA™ CO₂ analyzer
- IRMA™ airway adapter (adult/paediatric)
- Reusable SpO₂ sensor (selectable, see order numbers)
- USB data cable
- Power supply (FW7660M/06)
- Power supply adapter (Europe plug)
- Power supply adapter (UK plug)
- Li-ion battery (CT-2500)
- 4 x Batteries (AA Mignon)
- Silicone protective cover
- User manual + PC software (CD-ROM)

VM-2500-S:

- Sidestream device
- Nomo Adapter
- Disposable sampling line
- Airway adapter (adult/paediatric)
- Reusable SpO₂ sensor (selectable, see order numbers)
- USB data cable
- Power supply (FW7660M/06)
- Power supply adapter (Europe plug)
- Power supply adapter (UK plug)
- Li-ion battery (CT-2500)
- 4 x Batteries (AA Mignon)
- Silicone protective cover
- User manual + PC software (CD-ROM)

1) Should the device experience condensation it should be stored for more than 24 hours in an environment with relative moisture content below 95%RH (non-condensing).

2) With Li-ion battery the conditions are reduced to: operating while charging at 0 – 45°C and 86 – 106 kPa storage (1 month) at -20 – 60°C and 86 – 106 kPa

14.2 Order Number

Please indicate language version when ordering.

VM-2500 Starter Kit Versions

Device	Language Version	Order number
Mainstream VM-2500-M	Europe	4410500
	Asian	4410501
Sidestream VM-2500-S	Europe	4410520
	Asian	4410521

Language Version ¹⁾

- Europe: EN, DE, ES, FR, IT, NL, SE, RU (additional languages available upon request)
- Asian: EN, CN, JP (additional languages available on request)

SpO₂ Sensor Style

- SC 6500 VM
- SCP 6500 VM
- W 6500 VM
- EP 6500 VM
- SF 6500 VM

14.3 Accessories

CO₂ Mainstream Accessories

- IRMA™ airway adapter (adult/paediatric), P/N 4420511, Mainstream airway adapter for adult and paediatric use, box of 25
- IRMA™ airway adapter (infant), P/N 4420521, Mainstream airway adapter for infant use, box of 10

CO₂ Sidestream Accessories ²⁾

- Nomo Adapter, P/N 4420525, long term use water separator and bacterial filter for sidestream sampling, female Luer Lock connector
- Disposable sampling lines, P/N 4420550, disposable hydrophobic sampling line made of PE, 2m length, male/male Luer Lock connectors, box of 25

- Sidestream airway adapter (adult/paediatric), P/N 4420531, Sidestream airway adapter for adult and paediatric use, female Luer Lock connector, box of 25
- Airway adapter ISO 15mm / I.D. 2.0mm, P/N 4420570, airway adapter for infant and neonatal use, disposable, Luer Lock sideport – FLL, box of 10
- Airway adapter ISO 15mm / I.D. 2.5mm, P/N 4420571, airway adapter for infant and neonatal use, disposable, Luer Lock sideport – FLL, box of 10
- Airway adapter ISO 15mm / I.D. 3.0mm, P/N 4420572, airway adapter for infant and neonatal use, disposable, Luer Lock sideport – FLL, box of 10
- Airway adapter ISO 15mm / I.D. 3.5mm, P/N 4420573, airway adapter for infant and neonatal use, disposable, Luer Lock sideport – FLL, box of 10
- Airway adapter ISO 15mm / I.D. 4.0mm, P/N 4420574, airway adapter for infant and neonatal use, disposable, Luer Lock sideport – FLL, box of 10
- Airway adapter ISO 15mm / I.D. 5.0mm, P/N 4420575, airway adapter for infant and neonatal use, disposable, Luer Lock sideport – FLL, box of 10
- Airway adapter ISO 15mm / I.D. 6.0mm, P/N 4420576, airway adapter for infant and neonatal use, disposable, Luer Lock sideport – FLL, box of 10
- Airway adapter ISO 15mm / I.D. 7.0mm, P/N 4420577, airway adapter for infant and neonatal use, disposable, Luer Lock sideport – FLL, box of 10
- Airway adapter ISO 15mm / I.D. 8.0mm, P/N 4420578, airway adapter for infant and neonatal use, disposable, Luer Lock sideport – FLL, box of 10

1) Refer to the manufacturer for detailed information on the current languages available.

2) Note: Our sampling line configurations are continuously updated; please visit Viamed for detailed information. Specific patient sampling line configurations and interfaces are available upon request.

- Airway adapter ISO 15mm / I.D. 9.0mm, P/N 4420579, airway adapter for infant and neonatal use, disposable, Luer Lock sideport – FLL, box of 10
- Airway adapter ISO 15mm / I.D. 10.0mm, P/N 4420580, airway adapter for infant and neonatal use, disposable, Luer Lock sideport – FLL, box of 20

SpO₂ Accessories

- SC 6500 VM, P/N 0014750, 3rd generation adult soft sensor, 1.2m silicone cable
- SF 6500 VM, P/N 0014650, Fingerclip sensor, 1.2 m long cable, PVC cable
- SCP 6500 VM, P/N 0014751, 3rd generation paediatric soft sensor, 1.2m silicone cable
- SCA 6500 VM, P/N 0014590, 3rd generation adult soft sensor, 1.2m silicone cable, ≥ 200 autoclaving cycles
- SCPA 6500 VM, P/N 0014591, 3rd generation paediatric soft sensor, 1.2m silicone cable, ≥ 200 autoclaving cycles
- W 6500 VM, P/N 0014835, silicone wrap sensor, 1.2m silicone cable
- EP 6500 VM, P/N 0014825, Ear sensor, 1.2m long cable, PUR cable, PUR ear hanger
- XT 6500 VM Extension cable, P/N 0014895, 1.2m cable length, PVC cable
- XT 6501 VM Extension cable, P/N 0014896, 2.4m cable length, PVC cable

Additional SpO₂ sensors are available upon request

Other Accessories

- Universal Mounting Kit, P/N 4420600, V-adapter with female pole-mount thread
- Universal Mounting Clamp, P/N 0121200, Adapter with vertical and horizontal adjustment

- Carry case, P/N 0022170, Carrying bag for main unit and sensor, with shoulder strap
- Europe power supply plug, P/N 4430604, Europe adapter for the power supply FW 7660M/06
- UK power supply plug, P/N 4430603, UK adapter for the power supply FW 7660M/0

14.4 Replacement Parts

- IRMA™ CO₂ analyzer, P/N 4420505, mainstream CO₂ analyzer
- Power supply, FW7660M/06, P/N 4420595, Power supply for continuous operation of the VM-2500 and charging of Li-ion battery, Input: 100-240V AC/50-60Hz, Output: 6V DC/1.4A
- Silicone Protective Cover, P/N 4420610, Protective cover for the VM-2500 device
- Li-Ion battery, CT-2500, P/N 4420590, special rechargeable battery for use with the VM-2500, 3.7 V/2500 mAh
- USB Data Cable, P/N 0022174, Data cable for data transfer between the VM-2500 device and PC
- 4 x Batteries (AA Mignon), P/N 9950054
- VM-2500 CD-ROM, CD with the VM-2500 User Manual + PC software