

# Viamed Returns Shipping Invoice

Delivery Address:	Date: 09 Sep 2021	Shipper Address:
Teledyne Analytical Instruments	RMA:	Viamed
Vasu Narasimhan	Viamed RN: RTS1100764	15 Station Road
16830 Chestnut Street		Cross Hills
		Keighley
91748-1020		West Yorkshire
0016269341589		BD20 7DT
		+44 (0) 1535 634542
		+44 (0) 1535 635582
		info@viamed.co.uk

---

Returns Under Warranty Value for Customs : \$100.00  
See Attached Listing for Serial numbers and Faults

---

Part Number	P/N / Value	Description	Quantity
R17A	0110117 (\$16.00)	Oxygen Sensor - Automotive, R17A	4
	Tariff Code: HTS Code 9027.10.2000		
	Country of Origin: USA		
R22A	0110122 (\$16.00)	Oxygen Sensor - Automotive, R22A	22
	Tariff Code: HTS Code 9027.10.2000		
	Country of Origin: USA		
R17AVAN	0110127 (\$16.00)	Oxygen Sensor - Automotive, R17AVAN	1
	Tariff Code: HTS Code 9027.10.2000		
	Country of Origin: USA		
C44611-R22AV	0110132 (\$16.00)	Oxygen Sensor - Automotive, R22AV	3
	Tariff Code: HTS Code 9027.10.2000		
	Country of Origin: USA		
R17AH	0110137 (\$16.00)	Oxygen sensor - Automotive, R17AH	12
	Tariff Code: HTS Code 9027.10.2000		
	Country of Origin: USA		

Part Number	Serial Number	RMA	Fault Description
R17A	557894		Output ( 7mV - 13mV ) 14.6 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
R17A	557895		Output ( 7mV - 13mV ) 14.5 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
R17A	557901		Output ( 7mV - 13mV ) 13.8 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
R17A	557903		Output ( 7mV - 13mV ) 14.3 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
R22A	536303		Non-linear
R22A	573039		Output ( 7mV - 13mV ) 13.7 n Move Item to location 126552 n Failed in Q.A. n Did not reach end user
R22A	573057		Output ( 7mV - 13mV ) 14.3 n Move Item to location 126552 n Failed in Q.A. n Did not reach end user
R22A	573106		Output ( 7mV - 13mV ) 14.1 n Move Item to location 126552 n Failed in Q.A. n Did not reach end user
R22A	573148		Output ( 7mV - 13mV ) 13.4 n Move Item to location 126552 n Failed in Q.A. n Did not reach end user
R22A	573160		Output ( 7mV - 13mV ) 13.6 n Move Item to location 126552 n Failed in Q.A. n Did not reach end user
R22A	573182		Output ( 7mV - 13mV ) 14.3 n Move Item to location 126552 n Failed in Q.A. n Did not reach end user
R22A	573228		Output ( 7mV - 13mV ) 14.6 n Move Item to location 126552 n Failed in Q.A. n Did not reach end user
R22A	573233		Output ( 7mV - 13mV ) 13.6 n Move Item to location 126552 n Failed in Q.A. n Did not reach end user
R22A	573238		Output ( 7mV - 13mV ) 13.8 n Move Item to location 126552 n Failed in Q.A. n Did not reach end user
R22A	573253		Output ( 7mV - 13mV ) 15.3 n Move Item to location 126552 n Failed in Q.A. n Did not reach end user
R22A	573262		Output ( 7mV - 13mV ) 13.3 n Move Item to location 126552 n Failed in Q.A. n Did not reach end user
R22A	573288		Output ( 7mV - 13mV ) 13.5 n Move Item to location 126552 n Failed in Q.A. n Did not reach end user
R22A	573306		Output ( 7mV - 13mV ) 13.5 n Move Item to location 126552 n Failed in Q.A. n Did not reach end user
R22A	573331		Output ( 7mV - 13mV ) 13.6 n Move Item to location 126552 n Failed in Q.A. n Did not reach end user
R22A	573338		Output ( 7mV - 13mV ) 13.9. n Move Item to location 126552 n Failed in Q.A. n Did not reach end user
R22A	573393		Output ( 7mV - 13mV ) 14.3 n Move Item to location 126552 n Failed in Q.A. n Did not reach end user
R22A	573422		Output ( 7mV - 13mV ) 13.6 n Move Item to location 126552 n Failed in Q.A. n Did not reach end user
R22A	573492		Output ( 7mV - 13mV ) 13.6 n Move Item to location 126552 n Failed in Q.A. n Did not reach end user
R22A	575658		Output ( 7mV - 13mV ) 14.1 n Move Item to location 126552 n Failed in Q.A. n Did not reach end user
R22A	575671		Output ( 7mV - 13mV ) 13.3 n Move Item to location 126552 n Failed in Q.A. n Did not reach end user
R22A	575707		Output ( 7mV - 13mV ) 14.0 n Move Item to location 126552 n Failed in Q.A. n Did not reach end user
R17AVAN	534108		Output ( 9mV - 13mV ) 00.1 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C44611-R22AV	577380		Output ( 9mV - 13mV ) 8.8 n Move Item to location 126553 n Failed in Q.A. n Did not reach end user

Part Number	Serial Number	RMA	Fault Description
C44611-R22AV	577518		Output ( 9mV - 13mV ) 14.5 n Move Item to location 126553 n Failed in Q.A. n Did not reach end user
C44611-R22AV	577567		Output ( 9mV - 13mV ) 7.5 n Move Item to location 126553 n Failed in Q.A. n Did not reach end user
R17AH	568827		Output ( 10.5mV - 13.5mV ) 00.1 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
R17AH	568831		Output ( 10.5mV - 13.5mV ) 15.2 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
R17AH	568838		Output ( 10.5mV - 13.5mV ) 13.8 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
R17AH	568850		Output ( 10.5mV - 13.5mV ) 13.6 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
R17AH	568855		Output ( 10.5mV - 13.5mV ) 14.1 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
R17AH	568857		Output ( 10.5mV - 13.5mV ) 13.7 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
R17AH	568867		Output ( 10.5mV - 13.5mV ) 13.6 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
R17AH	568872		Output ( 10.5mV - 13.5mV ) 14.0 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
R17AH	568878		Output ( 10.5mV - 13.5mV ) 13.9 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
R17AH	568882		Output ( 10.5mV - 13.5mV ) 13.7 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
R17AH	568897		Output ( 10.5mV - 13.5mV ) 14.1 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
R17AH	568899		Output ( 10.5mV - 13.5mV ) 13.7 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user