

Viamed Returns Shipping Invoice

Delivery Address:
Teledyne Analytical Instruments
Vasu Narasimhan
16830 Chestnut Street

91748-1020

0016269341589

Date: 09 Sep 2021
RMA:
Viamed RN: RTS1100778

Shipper Address:
Viamed
15 Station Road
Cross Hills
Keighley
West Yorkshire
BD20 7DT
+44 (0) 1535 634542
+44 (0) 1535 635582
info@viamed.co.uk

Returns Under Warranty Value for Customs : \$100.00
See Attached Listing for Serial numbers and Faults

Part Number	P/N / Value	Description	Quantity
C43690-R17MED	0110017 (\$23.00)	Oxygen sensor, R17MED	9
	Tariff Code: 9019200000		
	Country of Origin: USA		
C44611-R22MED	0110022 (\$23.00)	Oxygen sensor, R22MED	18
	Tariff Code: 9019200000		
	Country of Origin: USA		
C-73894	0110081	Oxygen sensor, UFO 130-2	2
	Tariff Code: 9019200000		

Part Number	Serial Number	RMA	Fault Description
C43690-R17MED	411596		Non Linear
C43690-R17MED	411597		Non Linear
C43690-R17MED	593136		Output (7mV - 13mV) 00.1 n Move Item to location 126559 n Failed in Q.A. n Did not reach end user
C43690-R17MED	593184		Output (7mV - 13mV) 00.1 n Move Item to location 126559 n Failed in Q.A. n Did not reach end user
C43690-R17MED	593279		Output (7mV - 13mV) 14.8 n Move Item to location 126559 n Failed in Q.A. n Did not reach end user
C43690-R17MED	593281		Output (7mV - 13mV) 14.5 n Move Item to location 126559 n Failed in Q.A. n Did not reach end user
C43690-R17MED	593292		Output (7mV - 13mV) 13.7 n Move Item to location 126559 n Failed in Q.A. n Did not reach end user
C43690-R17MED	605632		Output (7mV - 13mV) 13.3 n Move Item to location 126559 n Failed in Q.A. n Did not reach end user
C43690-R17MED	605788		Output (7mV - 13mV) 14.3 n Move Item to location 126559 n Failed in Q.A. n Did not reach end user
C44611-R22MED	506866		
C44611-R22MED	506873		
C44611-R22MED	506875		
C44611-R22MED	601807		Output (7mV - 13mV) 13.2 n Move Item to location 126560 n Failed in Q.A. n Did not reach end user
C44611-R22MED	601810		Output (7mV - 13mV) 13.5 n Move Item to location 126560 n Failed in Q.A. n Did not reach end user
C44611-R22MED	601829		Output (7mV - 13mV) 13.8 n Move Item to location 126560 n Failed in Q.A. n Did not reach end user
C44611-R22MED	601871		Output (7mV - 13mV) 13.9 n Move Item to location 126560 n Failed in Q.A. n Did not reach end user
C44611-R22MED	601882		Output (7mV - 13mV) 14.7 n Move Item to location 126560 n Failed in Q.A. n Did not reach end user
C44611-R22MED	601883		Output (7mV - 13mV) 13.5 n Move Item to location 126560 n Failed in Q.A. n Did not reach end user
C44611-R22MED	601888		Output (7mV - 13mV) 13.4 n Move Item to location 126560 n Failed in Q.A. n Did not reach end user
C44611-R22MED	601900		Output (7mV - 13mV) 13.9 n Move Item to location 126560 n Failed in Q.A. n Did not reach end user
C44611-R22MED	601944		Output (7mV - 13mV) 13.5 n Move Item to location 126560 n Failed in Q.A. n Did not reach end user
C44611-R22MED	601949		Output (7mV - 13mV) 13.8 n Move Item to location 126560 n Failed in Q.A. n Did not reach end user
C44611-R22MED	601953		Output (7mV - 13mV) 14.3 n Move Item to location 126560 n Failed in Q.A. n Did not reach end user
C44611-R22MED	601956		Output (7mV - 13mV) 13.7 n Move Item to location 126560 n Failed in Q.A. n Did not reach end user
C44611-R22MED	601958		Output (7mV - 13mV) 14.0 n Move Item to location 126560 n Failed in Q.A. n Did not reach end user
C44611-R22MED	601977		Output (7mV - 13mV) 14.2 n Move Item to location 126560 n Failed in Q.A. n Did not reach end user
C44611-R22MED	601981		Output (7mV - 13mV) 14.3 n Move Item to location 126560 n Failed in Q.A. n Did not reach end user
C-73894	479035		Untested
C-73894	500812		Untested