

Viamed Returns Shipping Invoice

Delivery Address:
Teledyne Analytical Instruments
Vasu Narasimhan
16830 Chestnut Street

Date: 09 Sep 2021
RMA:
Viamed RN: RTS1100779

Shipper Address:
Viamed
15 Station Road
Cross Hills
Keighley
West Yorkshire
BD20 7DT
+44 (0) 1535 634542
+44 (0) 1535 635582
info@viamed.co.uk

91748-1020

0016269341589

Returns Under Warranty Value for Customs : \$100.00

See Attached Listing for Serial numbers and Faults

Part Number	P/N / Value	Description	Quantity
R17A	0110117 (\$16.00) Tariff Code: HTS Code 9027.10.2000 Country of Origin: USA	Oxygen Sensor - Automotive, R17A	9
R22A	0110122 (\$16.00) Tariff Code: HTS Code 9027.10.2000 Country of Origin: USA	Oxygen Sensor - Automotive, R22A	5
R17AVAN	0110127 (\$16.00) Tariff Code: HTS Code 9027.10.2000 Country of Origin: USA	Oxygen Sensor - Automotive, R17AVAN	1
R17AH	0110137 (\$16.00) Tariff Code: HTS Code 9027.10.2000 Country of Origin: USA	Oxygen sensor - Automotive, R17AH	3
R22AH	0110142 Tariff Code: HTS Code 9027.10.2000 Country of Origin: Mexico	Oxygen sensor - Automotive, R22AH	4

Part Number	Serial Number	RMA	Fault Description
R17A	589841		Output (7mV - 13mV) 13.5 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
R17A	601038		Output (7mV - 13mV) 13.4 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
R17A	601041		Output (7mV - 13mV) 14.7 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
R17A	601047		Output (7mV - 13mV) 00.1 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
R17A	601063		Output (7mV - 13mV) ERRATIC n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
R17A	601065		Output (7mV - 13mV) 13.8 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
R17A	601072		Output (7mV - 13mV) 13.4 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
R17A	601074		Output (7mV - 13mV) 13.8 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
R17A	601077		Output (7mV - 13mV) 14.5 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
R22A	595305		Output (7mV - 13mV) 1.4 n Move Item to location 126552 n Failed in Q.A. n Did not reach end user
R22A	595595		Output (7mV - 13mV) 13.5 n Move Item to location 126552 n Failed in Q.A. n Did not reach end user
R22A	595599		Output (7mV - 13mV) 13.7 n Move Item to location 126552 n Failed in Q.A. n Did not reach end user
R22A	595612		Output (7mV - 13mV) 13.8 n Move Item to location 126552 n Failed in Q.A. n Did not reach end user
R22A	595651		Output (7mV - 13mV) 4.2 n Move Item to location 126552 n Failed in Q.A. n Did not reach end user
R17AVAN	608175		Output (9mV - 13mV) 13.7 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
R17AH	608027		Output (10.5mV - 13.5mV) 14.3 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
R17AH	608052		Output (10.5mV - 13.5mV) 14.4 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
R17AH	608082		Output (10.5mV - 13.5mV) 13.9 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
R22AH	396432		Output (10.5mV - 13.5mV) 13.8 n Move Item to location 126550 n Failed in Q.A. n Did not reach end user
R22AH	608379		Output (10.5mV - 13.5mV) 14.3 n Move Item to location 126550 n Failed in Q.A. n Did not reach end user
R22AH	608396		Output (10.5mV - 13.5mV) 13.8 n Move Item to location 126550 n Failed in Q.A. n Did not reach end user
R22AH	608417		Output (10.5mV - 13.5mV) 15.1 n Move Item to location 126550 n Failed in Q.A. n Did not reach end user