

Viamed Returns Shipping Invoice

Delivery Address:
Teledyne Analytical Instruments
Vasu Narasimhan
16830 Chestnut Street

91748-1020

0016269341589

Date: 09 Sep 2021
RMA:
Viamed RN: RTS1100781

Shipper Address:
Viamed
15 Station Road
Cross Hills
Keighley
West Yorkshire
BD20 7DT
+44 (0) 1535 634542
+44 (0) 1535 635582
info@viamed.co.uk

Returns Under Warranty Value for Customs : \$100.00
See Attached Listing for Serial numbers and Faults

Part Number	P/N / Value	Description	Quantity
C61340-R24MED	0110024 (\$25.00)	Oxygen sensor, R24MED	56
	Tariff Code: 9019200000		
	Country of Origin: USA		

Part Number	Serial Number	RMA	Fault Description
C61340-R24MED	609274		Output (10mV - 17mV) ERRATIC n Move Item to location 126561 n Output (10mV - 17mV) ERRATIC n Move Item to location 126561 n Failed in Q.A. n Did not reach end user
C61340-R24MED	609282		Output (10mV - 17mV) ERRATIC n Move Item to location 126561 n Failed in Q.A. n Did not reach end user
C61340-R24MED	609287		Output (10mV - 17mV) ERRATIC n Move Item to location 126561 n Failed in Q.A. n Did not reach end user