

# Viamed Returns Shipping Invoice

Delivery Address:	Date: 09 Sep 2021	Shipper Address:
Teledyne Analytical Instruments	RMA:	Viamed
Luz Acosta	Viamed RN: RTS1100801	15 Station Road
16830 Chestnut Street		Cross Hills
		Keighley
91748-1020		West Yorkshire
0016269341589		BD20 7DT
		+44 (0) 1535 634542
		+44 (0) 1535 635582
		info@viamed.co.uk

Returns Under Warranty Value for Customs : \$100.00  
See Attached Listing for Serial numbers and Faults

Part Number	P/N / Value	Description	Quantity
C43690-R17MED	0110017 (\$23.00)	Oxygen sensor, R17MED	3
	Tariff Code: 9019200000		
	Country of Origin: USA		
C44611-R22MED	0110022 (\$23.00)	Oxygen sensor, R22MED	17
	Tariff Code: 9019200000		
	Country of Origin: USA		
C61340-R24MED	0110024 (\$25.00)	Oxygen sensor, R24MED	10
	Tariff Code: 9019200000		
	Country of Origin: USA		
A37016-T1	0110051 (\$27.00)	Oxygen sensor, T1	1
	Tariff Code: 9019200000		
	Country of Origin: USA		
A51327-T7	0110057 (\$22.00)	Oxygen sensor, T7	1
	Tariff Code: 9019200000		
	Country of Origin: USA		

Part Number	Serial Number	RMA	Fault Description
C43690-R17MED	593247		Non-linear
C43690-R17MED	617363		Output ( 7mV - 13mV ) 13.6 Failed in Q.A. Did not reach end user
C43690-R17MED	625571		Output ( 7mV - 13mV ) erratic Failed in Q.A. Did not reach end user
C44611-R22MED	531997		Non-linear
C44611-R22MED	609160		Output ( 7mV - 13mV ) 13.8 Failed in Q.A. Did not reach end user
C44611-R22MED	609162		Output ( 7mV - 13mV ) 13.8 Failed in Q.A. Did not reach end user
C44611-R22MED	618742		Output ( 7mV - 13mV ) 13.7 Failed in Q.A. Did not reach end user
C44611-R22MED	618744		Output ( 7mV - 13mV ) 14.2 Failed in Q.A. Did not reach end user
C44611-R22MED	618754		Output ( 7mV - 13mV ) 13.5 Failed in Q.A. Did not reach end user
C44611-R22MED	618756		Output ( 7mV - 13mV ) 13.6 Failed in Q.A. Did not reach end user
C44611-R22MED	618760		Output ( 7mV - 13mV ) 14.0 Failed in Q.A. Did not reach end user
C44611-R22MED	618787		Output ( 7mV - 13mV ) 13.7 Failed in Q.A. Did not reach end user
C44611-R22MED	618796		Output ( 7mV - 13mV ) 13.5 Failed in Q.A. Did not reach end user
C44611-R22MED	618807		Output ( 7mV - 13mV ) 13.4 Failed in Q.A. Did not reach end user
C44611-R22MED	618819		Output ( 7mV - 13mV ) 13.5 Failed in Q.A. Did not reach end user
C44611-R22MED	618851		Output ( 7mV - 13mV ) 13.5 Failed in Q.A. Did not reach end user
C44611-R22MED	618875		Output ( 7mV - 13mV ) 13.7 n Move Item to location 126560 n Failed in Q.A. n Did not reach end user
C44611-R22MED	622705		Output ( 7mV - 13mV ) 13.4 Failed in Q.A. Did not reach end user
C44611-R22MED	622729		Output ( 7mV - 13mV ) 13.4 Failed in Q.A. Did not reach end user
C44611-R22MED	625113		Output ( 7mV - 13mV ) erratic Failed in Q.A. Did not reach end user
C61340-R24MED	608494		Output ( 10mV - 17mV ) erratic Failed in Q.A. Did not reach end user
C61340-R24MED	608497		Output ( 10mV - 17mV ) erratic Failed in Q.A. Did not reach end user
C61340-R24MED	608498		Output ( 10mV - 17mV ) erratic Failed in Q.A. Did not reach end user
C61340-R24MED	608501		Output ( 10mV - 17mV ) erratic Failed in Q.A. Did not reach end user
C61340-R24MED	608502		Output ( 10mV - 17mV ) erratic Failed in Q.A. Did not reach end user
C61340-R24MED	608505		Output ( 10mV - 17mV ) erratic Failed in Q.A. Did not reach end user
C61340-R24MED	609292		Output ( 10mV - 17mV ) erratic Failed in Q.A. Did not reach end user
C61340-R24MED	609293		Output ( 10mV - 17mV ) erratic Failed in Q.A. Did not reach end user

Part Number	Serial Number	RMA	Fault Description
C61340-R24MED	628294		Output ( 10mV - 17mV ) 2.5 Erratic Failed in Q.A. Did not reach end user
C61340-R24MED	628298		Output ( 10mV - 17mV ) 2.6 Erratic Failed in Q.A. Did not reach end user
A37016-T1	553531		Leaking Failed in Q.A. Did not reach end user
A51327-T7	590756		Calibration Failed