



STILL VALIDATING THIS DATA

Multiple Suppliers

z₂

Use Honeywell Healthcare Solutions GmbH References

Use VMS / Nammu-Tech References

Unique Faults QA , Demo & Customer 26 Oct 2019 - 27 Oct 2021

	Cleaned Fault	8010004	8010006	8010007	8010008	8010012	8010014	8010018	8010020	8010022	8010023	8010026	8010050
Super Clean	Casing Out Of Specification - Too Large												173
Super Clean	Connector - Unstable Connection			1									
Super Clean	Electrolyte Leakage						1						
Super Clean	High Magnetic Signature												9
Super Clean	High Output						1		1				2
Super Clean	Linearity Error		1	1									
Super Clean	Low Output		4	1									1
Super Clean	No Output	13	6	2	2	2			1	3			3
Super Clean	Response Time Out Of Specification												1
Super Clean	Unknown Fault					8				1			
Super Clean	Unstable Output Signal		1	1	3				3	2	1		
Super Clean	Zero Offset Signal Out Of Tolerance									2			

Months to Failure From Invoice Customers

26 Oct 2019 - 27 Oct 2021

	From Invoice to Customer to Booked into Returns system (Customers)															
Clean Code	0	1	2	3	4	5	6	7	8	9	10	11	12	13	15	
	Months	Months	Months	Months	Months	Months	Months	Months	Months	Months	Months	Months	Months	Months	Months	
Casing Out Of Specification - Too Large					130			2			24			12		

Unstable Output Signal	1	2	1	2				2	1	2	2		
High Output Zero Offset Signal Out Of Tolerance	1			1				1	1				
No Output	3	2	1	4	1	5	2	4	4	1			1
Unknown Fault							6	1			2		
Response Time Out Of Specification													1
Linearity Error	1							1					
Low Output	1		2		1	1	2	3					
Electrolyte Leakage					1								1
Connector - Unstable Connection		1											
High Magnetic Signature	5			3									

Stock References Months to Failure From Invoice Customers 26 Oct 2019 - 27 Oct 2021

From Invoice to Customer to Booked into Returns system (Customers)															
Reference	0 Months	1 Months	2 Months	3 Months	4 Months	5 Months	6 Months	7 Months	8 Months	9 Months	10 Months	11 Months	12 Months	13 Months	15 Months
8010007				1					1	1	2			1	
8010022				1					1	1	3	2			1
8010004						3	2	3		2	1		2		
8010012			1	1	1				5	3			2		
8010050	5		1		130	4		5		1	24			13	
8010006		2	1	1	2	1			3		1			1	
8010014							1								
8010023				1											
8010020					1	2						1			
8010008			1						1	3					
8010018			1												

Break Down Parts To Fault Code

Stock Ref	Description	Fault Code	Cleaned Fault Code	True Fault?	Qty
8010050	Oxygen Sensor - JFD (non-mag).	Casing out of specification - too large	Casing Out Of Specification - Too Large	Fault	173
8010007	Oxygen Sensor - NaNS01	Connector - unstable connection	Connector - Unstable Connection	Fault	1
8010014	Oxygen Sensor - K-22D	Electrolyte Leakage	Electrolyte Leakage	Fault	1
8010050	Oxygen Sensor - JFD (non-mag).	High magnetic signature	High Magnetic Signature	Fault	9
8010018	Oxygen Sensor - Triton T22D	High Output	High Output	Fault	1
8010022	Oxygen Sensor - APD16	High Output	High Output	Fault	1
8010050	Oxygen Sensor - JFD (non-mag).	High Output	High Output	Fault	2
8010006	Oxygen Sensor - rEvo CR22D	Linearity Error	Linearity Error	Fault	1
8010007	Oxygen Sensor - NaNS01	Linearity Error	Linearity Error	Fault	1
8010006	Oxygen Sensor - rEvo CR22D	Low Output	Low Output	Fault	4
8010007	Oxygen Sensor - NaNS01	Low Output	Low Output	Fault	1
8010050	Oxygen Sensor - JFD (non-mag).	Low Output	Low Output	Fault	1
8010004	Oxygen Sensor - R17JJ-CCR	No Fault Found	No Fault Found	No Fault	11

8010006	Oxygen Sensor - rEvo CR22D	No Fault Found	No Fault Found	No Fault	4
8010007	Oxygen Sensor - NaNS01	No Fault Found	No Fault Found	No Fault	1
8010026	Oxygen Sensor - MDS100	No Fault Found	No Fault Found	No Fault	1
8010050	Oxygen Sensor - JFD (non-mag).	No Fault Found	No Fault Found	No Fault	1
8010004	Oxygen Sensor - R17JJ-CCR	No Output	No Output	Fault	13
8010006	Oxygen Sensor - rEvo CR22D	No Output	No Output	Fault	6
8010007	Oxygen Sensor - NaNS01	No Output	No Output	Fault	2
8010008	Oxygen Sensor - NaNM01	No Output	No Output	Fault	2
8010012	Oxygen Sensor - SC-I22D	No Output	No Output	Fault	2
8010020	Oxygen Sensor - type 22D	No Output	No Output	Fault	1
8010022	Oxygen Sensor - APD16	No Output	No Output	Fault	3
8010050	Oxygen Sensor - JFD (non-mag).	No Output	No Output	Fault	3
8010004	Oxygen Sensor - R17JJ-CCR	Out Of Warranty	Out Of Warranty	No Fault	20
8010006	Oxygen Sensor - rEvo CR22D	Out Of Warranty	Out Of Warranty	No Fault	3
8010007	Oxygen Sensor - NaNS01	Out Of Warranty	Out Of Warranty	No Fault	3
8010050	Oxygen Sensor - JFD (non-mag).	Out Of Warranty	Out Of Warranty	No Fault	10
8010050	Oxygen Sensor - JFD (non-mag).	Response Time Out Of Specification	Response Time Out Of Specification	Fault	1
8010012	Oxygen Sensor - SC-I22D	Unknown Fault	Unknown Fault	Fault	8
8010022	Oxygen Sensor - APD16	Unknown Fault	Unknown Fault	Fault	1
8010006	Oxygen Sensor - rEvo CR22D	Unstable Output Signal	Unstable Output Signal	Fault	1
8010007	Oxygen Sensor - NaNS01	Unstable Output Signal	Unstable Output Signal	Fault	1
8010008	Oxygen Sensor - NaNM01	Unstable Output Signal	Unstable Output Signal	Fault	3
8010020	Oxygen Sensor - type 22D	Unstable Output Signal	Unstable Output Signal	Fault	3
8010022	Oxygen Sensor - APD16	Unstable Output Signal	Unstable Output Signal	Fault	2
8010023	Oxygen Sensor - APD8	Unstable Output Signal	Unstable Output Signal	Fault	1
8010022	Oxygen Sensor - APD16	Zero Offset Signal Out Of Tolerance	Zero Offset Signal Out Of Tolerance	Fault	2

Z_2 - Returns Statistics 2021-10-26

Start From 26/10/2019

End at 26/10/2021