


[Serial number lists for opera](#)

Company / Hospital	Ambient Pressure Diving Ltd
Contact Name	M Wroe
Department	
Position	
Company Phone	01326 563834
Direct Phone	01326 561040
Opera Account	VST0002
Viamed ORD Number	
Email	
Order Number	
Type	Warranty

SRS67423 Warranty Repair

Ambient Pressure Diving Ltd,M Wroe 01326 561040

SRS67423 Notes:

12/Feb/2020 Robert Connor

12/Feb/2020 Robert Connor

12/Feb/2020 Robert Connor

12/Feb/2020 Robert Connor

Received 1 x APD16 s/n APD117680

17/May/2021 Catherine Spence

replacing under warranty as per steve nixon

All items for repair

[Add Blank Line For new item linked to repair](#)

SRN	Equipment	Reference	Serialnumber Barcode	Serialnumber	Description	Customer Warranty New	Customer Warranty Repair	Our Supplier Warranty	Status	History
SRN32356	Oxygen Sensor VST	8010022	1371891	APD117680	Oxygen Sensor APD16	Yes Change	No Change	Check	Replaced	History

SRN32356 Oxygen Sensor VST 8010022 APD117680 Oxygen Sensor APD16

[Global fill fields across multiple 8010022's records on SRS](#)
[Create Issue - Use to link files and or photos to the SRS](#)

Internal Engineer Notes against SRN32356

12/Feb/2020 Robert Connor

Sensor tested in air at 998mbar, 10.7 Celsius. Output 10.5 mV.

17/May/2021 Catherine Spence

replaced under warranty as per steve nixon

(add separately to completing a repair)

[Quote Item Against SRN32356](#)
[Request Item Against SRN32356](#)
[Add Labour Item SRN32356](#)
[Refresh List](#)

Notes Customer will read

Complete Repair

Chargable Time

Zero Seconds

Actual Repair Time

Zero Seconds

Fault Code

Unknown Fault

Status

Replaced

Submit

Quote Preview

Repair SRS67423 Values in Pounds @ £1: £1.00

Parts Cost

£0.00