

Service Repair Sheet SRS67969

Date Promised to Return: 01 Jan 1970 -3600

Contact Name

Dennis Vessey

Company/ Hospital Name

Megalodon UK

Department

Accounts

Position

Direct Phone

01652 688 904

General Phone

01652 688 904

Opera Account

VST010

Email

dennis.vessey@btconnect.com

Order Number

Date Received

06/Oct/2021

Booked in By

Robert Connor

Main Company

VandagraphST

Type Return

Warranty

Notes 30/Sep/2021 Ryan Swaine

30/Sep/2021 Ryan Swaine

Sensor SN:116198. intermittent reading, if you wiggle the connector it sometimes works.

06/Oct/2021 Robert Connor

Received 1 x SC-I22D s/n ISC116198

06/Oct/2021 Robert Connor

Cell added to RMA box 821.

18/Oct/2021 Catherine Spence

Sent back to supplier on tracking number 1ZA8Y7290454093180 today

06/Nov/2021 Catherine Spence

with supplier

08/Nov/2021 Catherine Spence

Received credit for item

08/Nov/2021 Catherine Spence

Replacing as a good will gesture, to replace on next order

Ready For quote

Repair Complete Signed

CGreen 8-11-2021

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN34226	Oxygen Sensor VST	8010012	ISC116198	Yes

Please replace 1x 8010012 on next order
No fault found
Replacing as good will gesture. cg 8-11-2021