

*Viamed*  
**Internal Audit Check list**

<b>Design</b>			
Created:	17/May 1995	<b>Audit No 03</b>	VOP 17
Revised:	26 July 2022		Page 1 of 12
Audit Date	26-7-22	Auditor <i>Heather Lombs</i>	

Sub Processes Linked to Audit 03

Review the below processes tasks and audits and ensure they are completed in a timely manner.

VST Ltd                   **Operational planning and control**  
 ISO9001:2015  
 8.1                        The organization shall plan, implement and control the processes (see 4.4) needed to meet the requirements for the provision of products and services, and to implement the actions determined in Clause 6, by:  
                           a) determining the requirements for the products and services;  
                           b) establishing criteria for:  
                             1) the processes;  
                             2) the acceptance of products and services;  
                           c) determining the resources needed to achieve conformity to the product and service requirements;  
                           d) implementing control of the processes in accordance with the criteria;  
                           e) determining, maintaining and retaining documented information to the extent necessary:  
                             1) to have confidence that the processes have been carried out as planned;  
                             2) to demonstrate the conformity of products and services to their requirements.  
                           The output of this planning shall be suitable for the organizations operations.  
                           The organization shall control planned changes and review the consequences of unintended changes, taking action to mitigate any adverse effects, as necessary.  
                           The organization shall ensure that outsourced processes are controlled (see 8.4). 

VST Ltd                   The organization shall establish, implement and maintain a design and development process that is appropriate to ensure the subsequent provision of products and services. 

VST Ltd                   **Design and development planning**  
 ISO9001:2015  
 8.3.1 General           In determining the stages and controls for design and development, the organization shall consider:  
                           a) the nature, duration and complexity of the design and development activities;  
                           b) the required process stages, including applicable design and development reviews;  
                           c) the required design and development verification and validation activities;

Internal Audit Check list			
Design			
Created:	17/May 1995	<b>Audit No 03</b>	VOP 17
Revised:	26 July 2022		Page 2 of 12
Audit Date		Auditor	

- d) the responsibilities and authorities involved in the design and development process;
- e) the internal and external resource needs for the design and development of products and services;
- f) the need to control interfaces between persons involved in the design and development process;
- g) the need for involvement of customers and users in the design and development process;
- h) the requirements for subsequent provision of products and services;
- i) the level of control expected for the design and development process by customers and other relevant interested parties;
- j) the documented information needed to demonstrate that design and development requirements have been met.



VST Ltd  
ISO9001:2015  
8.3.3

**Design and development inputs**

The organization shall determine the requirements essential for the specific types of products and services to be designed and developed. The organization shall consider:

- a) functional and performance requirements;
- b) information derived from previous similar design and development activities;
- c) statutory and regulatory requirements;
- d) standards or codes of practice that the organization has committed to implement;
- e) potential consequences of failure due to the nature of the products and services.



Inputs shall be adequate for design and development purposes, complete and unambiguous.

Conflicting design and development inputs shall be resolved. The organization shall retain documented information on design and development inputs.

VST Ltd  
ISO9001:2015  
8.3.4

**Design and development controls**

The organization shall apply controls to the design and development process to ensure that:

- a) the results to be achieved are defined;
- b) reviews are conducted to evaluate the ability of the results of design and development to meet requirements;
- c) verification activities are conducted to ensure that the design and development outputs meet the input requirements;
- d) validation activities are conducted to ensure that the resulting products and services meet the requirements for the specified application or intended use;



Internal Audit Check list			
Design			
Created:	17/May 1995	<b>Audit No 03</b>	VOP 17
Revised:	26 July 2022		Page 3 of 12
Audit Date		Auditor	

e) any necessary actions are taken on problems determined during the reviews, or verification and validation activities;  
 f) documented information of these activities is retained.  
 NOTE Design and development reviews, verification and validation have distinct purposes. They can be conducted separately or in any combination, as is suitable for the products and services of the organization.

VST Ltd  
 ISO9001:2015  
 8.3.5

**Design and development outputs**

The organization shall ensure that design and development outputs:

- a) meet the input requirements;
- b) are adequate for the subsequent processes for the provision of products and services;
- c) include or reference monitoring and measuring requirements, as appropriate, and acceptance criteria;
- d) specify the characteristics of the products and services that are essential for their intended purpose and their safe and proper provision.

The organization shall retain documented information on design and development outputs.

✓

VST Ltd  
 ISO9001:2015  
 8.3.6

**Design and development changes**

The organization shall identify, review and control changes made during, or subsequent to, the design and development of products and services, to the extent necessary to ensure that there is no adverse impact on conformity to requirements.

The organization shall retain documented information on:

- a) design and development changes;
- b) the results of reviews;
- c) the authorization of the changes;
- d) the actions taken to prevent adverse impacts.

✓

VST Ltd  
 ISO9001:2015  
 8.5.1

**Control of production and service provision**

The organization shall implement production and service provision under controlled conditions.

Controlled conditions shall include, as applicable:

- a) the availability of documented information that defines:
  - 1) the characteristics of the products to be produced, the services to be provided, or the activities to be performed;
  - 2) the results to be achieved;
- b) the availability and use of suitable monitoring and measuring resources;
- c) the implementation of monitoring and measurement activities at appropriate stages to verify that criteria for control of processes or outputs, and acceptance criteria for products and services, have been met;

✓

Internal Audit Check list			
Design			
Created:	17/May 1995	<b>Audit No 03</b>	VOP 17
Revised:	26 July 2022		Page 4 of 12
Audit Date		Auditor	

- d) the use of suitable infrastructure and environment for the operation of processes;
- e) the appointment of competent persons, including any required qualification;
- f) the validation, and periodic revalidation, of the ability to achieve planned results of the processes for production and service provision, where the resulting output cannot be verified by subsequent monitoring or measurement;
- g) the implementation of actions to prevent human error;
- h) the implementation of release, delivery and post-delivery activities

Viamed Ltd  
ISO13485:2016  
4.2.3 Medical device file

**Documentation requirements**

For each medical device type or medical device family, the organization shall establish and maintain one or more files either containing or referencing documents generated to demonstrate conformity with the requirement of this International Standard and compliance with applicable regulatory requirements.

The content of the file(s) shall include, but is not limited to:

- a) general description of the medical device, intended use/purpose, and labelling, including any instructions for use;
- b) specifications for product;
- c) specifications or procedures for manufacturing, packaging, storage, handling and distribution;
- d) procedures for measuring and monitoring;
- e) as appropriate, requirements for installation;
- f) as appropriate, procedures for servicing.

Viamed Ltd  
ISO13485:2016  
7.1

**Planning of product realization**

The organization shall plan and develop the processes needed for product realization. Planning of product realization shall be consistent with the requirements of the other processes of the quality management system. The organization shall document one or more processes for risk management in product realization. Records of risk management activities shall be maintained (see 4.2.5).

In planning product realization, the organization shall determine the following, as appropriate:

- a) quality objectives and requirements for the product;
- b) the need to establish processes and documents (see 4.2.4) and to provide resources specific to the product, including infrastructure and work environment;
- c) required verification, validation, monitoring, measurement, inspection and test, handling, storage, distribution and traceability activities specific to the

✓  
Doc index  
Tech files  
Rules +  
tasks  
Procedures

Doc index  
Tech files  
Management  
Review  
Route map  
Bar codes  
QA system

Internal Audit Check list			
Design			
Created:	17/May 1995	<b>Audit No 03</b>	VOP 17
Revised:	26 July 2022		Page 5 of 12
Audit Date		Auditor	

product together with the criteria for product acceptance;  
 d) records needed to provide evidence that the realization processes and resulting product meet requirements (see 4.2.5).

The output of this planning shall be documented in a form suitable for the organization's method of operations.

NOTE Further information can be found in ISO 14971.

Viamed Ltd

ISO13485:2016

7.3.1

#### General

The organization shall document procedures for design and development

Viamed Ltd

ISO13485:2016

7.3.10

#### Design and development files

The organization shall maintain a design and development file for each medical device type or medical device family. This file shall include or reference records generated to demonstrate conformity to the requirements for design and development and records for design and development changes.

Doc index  
marketing index

tech files

Procedures

Viamed Ltd

ISO13485:2016

7.3.2

#### Design and development planning

The organization shall plan and control the design and development of product. As appropriate, design and development planning documents shall be maintained and updated as the design and development progresses.

During design and development planning, the organization shall document:

- a) the design and development stages;
- b) the review(s) needed at each design and development stage;
- c) the verification, validation, and design transfer activities that are appropriate at each design and development stage;
- d) the responsibilities and authorities for design and development;
- e) the methods to ensure traceability of design and development outputs to design and development inputs;
- f) the resources needed including necessary competence of personnel

tech files  
marketing index  
issues  
Doc index

Review meetings

Viamed Ltd

ISO13485:2016

7.3.3

#### Design and development inputs

Inputs relating to product requirements shall be determined and records maintained (see 4.2.5). These inputs shall include:

- a) functional, performance, usability and safety requirements, according to the intended use;
- b) applicable regulatory requirements and standards;

Doc index  
tech files  
QA system

Role maps  
Roles + tasks  
marketing system

Internal Audit Check list			
Design			
Created:	17/May 1995	<b>Audit No 03</b>	VOP 17
Revised:	26 July 2022		Page 6 of 12
Audit Date		Auditor	

- c) applicable output(s) of risk management;
- d) as appropriate, information derived from previous similar designs;
- e) other requirements essential for design and development of the product and processes.

These inputs shall be reviewed for adequacy and approved. Requirements shall be complete, unambiguous, able to be verified or validated, and not in conflict with each other. NOTE Further information can be found in IEC 62366-1.

Viamed Ltd

ISO13485:2016

7.3.4

#### **Design and development outputs**

Design and development outputs shall:

- a) meet the input requirements for design and development;
- b) provide appropriate information for purchasing, production and service provision;
- c) contain or reference product acceptance criteria;
- d) specify the characteristics of the product that are essential for its safe and proper use.

The outputs of design and development shall be in a form suitable for verification against the design and development inputs and shall be approved prior to release.

Records of the design and development outputs shall be maintained (see 4.2.5).

*Tech files*

*Doc index*

*Review meetings*

Viamed Ltd

ISO13485:2016

7.3.5

#### **Design and development review**

At suitable stages, systematic reviews of design and development shall be performed in accordance with planned and documented arrangements to:

- a) evaluate the ability of the results of design and development to meet requirements;
- b) identify and propose necessary actions.

Participants in such reviews shall include representatives of functions concerned with the design and development stage being reviewed, as well as other specialist personnel.

Records of the results of the reviews and any necessary actions shall be maintained and include the identification of the design under review, the participants involved and the date of the review (see 4.2.5).

*tech files*

*Doc index*

*Review meetings*

*QA system*

*management*

*Review*

Viamed Ltd

ISO13485:2016

7.3.6

#### **Design and development verification**

Design and development verification shall be performed in accordance with planned and documented arrangements to ensure that the design and development outputs have met the design and development input requirements.

The organization shall document verification plans that include methods, acceptance criteria and, as appropriate,

*tech files*

*management*

*Review*

*Doc index*

*QA system*

Internal Audit Check list			
Design			
Created:	17/May 1995	<b>Audit No</b> 03	VOP 17
Revised:	26 July 2022		Page 7 of 12
Audit Date		Auditor	

statistical techniques with rationale for sample size. If the intended use requires that the medical device be connected to, or have an interface with, other medical device(s), verification shall include confirmation that the design outputs meet design inputs when so connected or interfaced. Records of the results and conclusions of the verification and necessary actions shall be maintained (see 4.2.4 and 4.2.5).

Viamed Ltd

ISO13485:2016

7.3.7

**Design and development validation**

Design and development validation shall be performed in accordance with planned and documented arrangements to ensure that the resulting product is capable of meeting the requirements for the specified application or intended use. The organization shall document validation plans that include methods, acceptance criteria, and, as appropriate, statistical techniques with rationale for sample size.

Design validation shall be conducted on representative product. Representative product includes initial production units, batches or their equivalents. The rationale for the choice of product used for validation shall be recorded (see 4.2.5).

As part of design and development validation, the organization shall perform clinical evaluations or performance evaluations of the medical device in accordance with applicable regulatory requirements.

A medical device used for clinical evaluation or performance evaluation is not considered to be released for use to the customer.

If the intended use requires that the medical device be connected to, or have an interface with, other medical device(s), validation shall include confirmation that the requirements for the specified application or intended use have been met when so connected or interfaced.

Validation shall be completed prior to release for use of the product to the customer.

Records of the results and conclusion of validation and necessary actions shall be maintained (see 4.2.4 and 4.2.5).

Doc index  
tech files  
procedures

Roles +  
tasks.

Viamed Ltd

ISO13485:2016

7.3.8

**Design and development transfer**

The organization shall document procedures for transfer of design and development outputs to manufacturing. These procedures shall ensure that design and development outputs are verified as suitable for manufacturing before becoming final production specifications and that production capability can meet product requirements.

Results and conclusions of the transfer shall be recorded (see

procedures  
Doc index  
tech files

Internal Audit Check list			
Design			
Created:	17/May 1995	<b>Audit No 03</b>	VOP 17
Revised:	26 July 2022		Page 8 of 12
Audit Date		Auditor	

4.2.5).

Viamed Ltd      **Control of design and development changes**  
 ISO13485:2016      The organization shall document procedures to control design and development changes. The organization shall determine the significance of the change to function, performance, usability, safety and applicable regulatory requirements for the medical device and its intended use.

7.3.9      Design and development changes shall be identified. Before implementation, the changes shall be:

- reviewed;
- verified;
- validated, as appropriate;
- approved.

management review

tech files  
doc index

The review of design and development changes shall include evaluation of the effect of the changes on constituent parts and product in process or already delivered, inputs or outputs of risk management and product realization processes. Records of changes, their review and any necessary actions shall be maintained (see 4.2.5).

Viamed Ltd      **Validation of processes for production and service provision**  
 ISO13485:2016

7.5.6      The organization shall validate any processes for production and service provision where the resulting output cannot be or is not verified by subsequent monitoring or measurement and, as a consequence, deficiencies become apparent only after the product is in use or the service has been delivered.

Doc index  
Barcode system

procedures  
management review

Training Record  
calibration index

Validation shall demonstrate the ability of these processes to achieve planned results consistently.

The organization shall document procedures for validation of processes including:

- defined criteria for review and approval of the processes;
- equipment qualification and qualification of personnel;
- use of specific methods, procedures and acceptance criteria;
- as appropriate, statistical techniques with rationale for sample sizes
- requirements for records (see 4.2.5);
- revalidation, including criteria for revalidation;
- approval of changes to the processes.

The organization shall document procedures for the validation of the application of computer software used in production and service provision. Such software applications shall be validated prior to initial use and, as appropriate, after changes to such software or its application. The specific approach and

## Internal Audit Check list

### Design

Created:	17/May 1995	<b>Audit No</b> 03	VOP 17
Revised:	26 July 2022		Page 9 of 12
Audit Date		Auditor	

activities associated with software validation and revalidation shall be proportionate to the risk associated with the use of the software including the effect on the ability of the product to conform to specifications.

Records of the results and conclusion of validation and necessary actions from the validation shall be maintained (see 4.2.4 and 4.2.5).

	<b>QUESTION:</b>	<b>RESPONSE:</b>	Y/ N
1	Review Last years Audit. Update processes if required. Are all follow on Issue resolved satisfactory.		
2	Technical File Reviewed ID: <i>Task SO #261987 ✓</i>		Y
3	Check that the final design responsibility is a Sole Authority.	Top management VOP 17	Y
4	Check that all products are C.E. marked and Viamed products have a C.E. file.	Intrastats <i>No products</i>	N/A
5	Pick 5 Products from the Files product list	<i>No products</i>	N/A
6	Declaration on Conformance Certificates present	<i>No products</i>	N/A
7	Verify that EMC testing has been identified where required.	<i>No products</i>	N/A
8	Are the latest BS ISO MDD requirements are available List DOC IDs	<i>No products</i>	N/A
9	Check that product classification is done to MDD principles.	<i>No products</i>	N/A
10	Verify that each design was initiated from a job description & specification	<i>No products</i>	N/A
11	Has each design has received a job number and a job progress form	<i>No products</i>	N/A
12	Verify the existence of a design documentation checklist.	<i>No products</i>	N/A
13	Check that estimated times have been noted. Electronic timing being introduced	<i>No products</i>	N/A
14	Have final testing requirements, and test criteria, been identified	<i>No products</i>	N/A
15	Have concession notes have been raised on non-approved suppliers	<i>No products</i>	N/A
16	Check that current status is identified on a regular basis.	<i>No products</i>	N/A
17	Verify that design reviews are undertaken and that records are retained	<i>No products</i>	N/A
18	Check that any amendments to design are logged	<i>No products</i>	N/A

## Internal Audit Check list

### Design

Created:	17/May 1995	<b>Audit No 03</b>	VOP 17
Revised:	26 July 2022		Page 10 of 12
Audit Date		Auditor	

19	Check that design output records are verified against design input	No products	N/A
20	Does design verification comply with COP 16 - 7.7.1 - .4	No products	N/A
21	Check that clinical trials have been carried out and relevant records retained	No products	N/A
22	Verify that design validation has been carried out as required by form QC30	No products	N/A
23	Check that any design changes have been identified, recorded and approved	No products	N/A
24	Check that CE files are complete, correct and maintained <i>via N/A</i> — files	PMS	Reviewed Y
25	Are design components kept separate from stock and adequately stored	Barcode tracking	Y
26	Are design component stocks labelled	Barcode tracking	Y
27	Check the existence of design compliance forms	No products	N/A
28	Check that these files are maintained	only PMS	Y

### List Processes Per Title

<u>Product Controller</u>						
Process Scope	Roll Task	Roll Audit	Risk	Action	Notes / Issues	
PROCESSID 7043			Freq Risk Overall			
PROCESSID 7045 Check that no Products / Designs have changed significantly to warrant informing MDD / Bsi / CMDCAS or any other related Body. First Check the Stockbox -> QA Fails Review (Slow Page) in case there are any products that have a New repair code that requires a risk assesment Carry out the Technical File PMS Form for Each OBL and Viamed Product Using procedure VM3COP27.11. See VM3COP18	50 Managing Director <i>26/7/27</i>		Freq 1 Risk 1 Overall 1	Task 2M		
PROCESSID 7047			Freq Risk Overall			

# Internal Audit Check list

## Design

Created:	17/May 1995	<b>Audit No 03</b>	VOP 17
Revised:	26 July 2022		Page 11 of 12
Audit Date		Auditor	

<u>ISO Controller</u>		<u>Roll Task</u>	<u>Roll Audit</u>	<u>Risk</u>	<u>Action</u>	<u>Notes / Issues</u>
<b>Process Scope</b>						
PROCESSID <b>5887</b> To Keep Products and Services up-to date with current regulations and standards		235 Managing Director	266759 ✓	Freq 2 Risk 2 Overall 4	Task 3M	
<u>Marketing Controller</u>						
<b>Process Scope</b>		<u>Roll Task</u>	<u>Roll Audit</u>	<u>Risk</u>	<u>Action</u>	<u>Notes / Issues</u>
PROCESSID <b>6942</b> Organisation of the process of actual marketing campaings etc..				Freq Risk Overall		
PROCESSID <b>7173</b> Making the Promotional Material leaflets etc..				Freq Risk Overall		
<u>Accounts Processes</u>						
<b>Process Scope</b>		<u>Roll Task</u>	<u>Roll Audit</u>	<u>Risk</u>	<u>Action</u>	<u>Notes / Issues</u>
PROCESSID <b>7919</b> send a report to John of what is happening with the debtors from the last month, include problems and payments due.		928 Company Secretary	929 266792 ✓	Freq 1 Risk 1 Overall 1	Task 1M Audit 6M 261872 ✓	
Can add to issue and redirect						
<u>Audits</u>						
<b>Process Scope</b>		<u>Roll Task</u>	<u>Roll Audit</u>	<u>Risk</u>	<u>Action</u>	<u>Notes / Issues</u>
PROCESSID <b>7716</b> To carry out Audit 03 Design Control Viamed		266765 this Audit Compan y Secretar y	266765 ✓	Freq 1 Risk 2 Overall 2	Audit 12M	
PROCESSID <b>7764</b> To carry out Audit 03 Design Control VST	VST Audit	193 Compan y Secretar y	193 ✓	Freq 1 Risk 2 Overall 2	Audit 12M	

Internal Audit Check list			
Design			
Created:	17/May 1995	<b>Audit No 03</b>	VOP 17
Revised:	26 July 2022		Page 12 of 12
Audit Date		Auditor	

<u>Design Products</u>							
Process Scope		Roll Task	Roll Audit	Risk	Action	Notes / Issues	
PROCESSID 42				Freq Risk Overall			