

# Viamed Returns Shipping Invoice

Delivery Address:

Teledyne Analytical Instruments  
Margie Read  
16830 Chestnut Street  
City Of Industry  
  
91748-1020

Date: 24 Jan 2022

RMA:  
Viamed RN: RTS1100840

Shipper Address:

Viamed  
15 Station Road  
Cross Hills  
Keighley  
West Yorkshire  
BD20 7DT  
+44 (0) 1535 634542  
+44 (0) 1535 635582  
info@viamed.co.uk

---

Returns Under Warranty Value for Customs : \$100.00

See Attached Listing for Serial numbers and Faults

---

Part Number	P/N / Value	Description	Quantity
C43690-R17MED	0110017 (\$22.00 Ea.) Tariff Code: 9019209000 Country of Origin: USA	Teledyne Oxygen sensor R-17MED	25
A37016-T1	0110051 (\$27.00) Tariff Code: 9019200000 Country of Origin: USA	Teledyne Oxygen sensor T-1	3

Part Number	Serial Number	RMA	Fault Description
C43690-R17MED	673681		No Output 00.1mV Failed in Q.A.Did not reach end user
C43690-R17MED	673765		No Output 00.1mV Failed in Q.A. Did not reach end user
C43690-R17MED	673767		No Output 00.1mVFailed in Q.A. Did not reach end user
C43690-R17MED	673776		No Output 00.1mV Failed in Q.A. Did not reach end user
C43690-R17MED	673777		No Output 00.1mVFailed in Q.A. Did not reach end user
C43690-R17MED	673781		No Output 00.1mV erratic Failed in Q.A. Did not reach end user
C43690-R17MED	673806		No Output 00.1mV Failed in Q.A. Did not reach end user
C43690-R17MED	673885		No Output 00.1mV Failed in Q.A Did not reach end user
C43690-R17MED	673888		No Output 00.1mVFailed in Q.A. Did not reach end user
C43690-R17MED	674208		No Output 00.1mV erratic Failed in Q.A. Did not reach end user
C43690-R17MED	687973		No Output 00.1mV erratic Failed in Q.A. Did not reach end user
C43690-R17MED	688000		No Output 00.1mv Failed in Q.A. Did not reach end user
C43690-R17MED	688018		No Output 00.1mV Failed in Q.A. Did not reach end user
C43690-R17MED	688127		No Output 00.1mV Failed in Q.A. n Did not reach end user
C43690-R17MED	688203		No Output 00.1mV Failed in Q.A. Did not reach end user
C43690-R17MED	688214		High Output13.3mV Failed in Q.A. n Did not reach end user
C43690-R17MED	688284		No Output 00.1mV Failed in Q.A. Did not reach end user
C43690-R17MED	688393		No Output 00.1mV erratic Failed in Q.A. Did not reach end user
C43690-R17MED	688427		No Output 00.1mV Failed in Q.A. Did not reach end user
C43690-R17MED	688483		High Output 13.3mV Failed in Q.A. Did not reach end user
C43690-R17MED	688782		No Output 00.1mV Failed in Q.A. Did not reach end user
C43690-R17MED	688797		No Output 00.1 mV Failed in Q.A. Did not reach end user
C43690-R17MED	688822		High Output 13.6 Failed in Q.A. Did not reach end user
C43690-R17MED	696958		Output erratic 0.1 -9.3 Failed in Q.A. Did not reach end user
C43690-R17MED	696991		Low Output 3.0mV Failed in Q.A. Did not reach end user
A37016-T1	664640		High Output Failed in QA Did not reach end user
A37016-T1	664719		No Output Failed in Q.A. Did not reach end user
A37016-T1	664734		High Output Did not reach end user Failed in QA