


oxygen monitoring

Flowsensor A

for flow measurement of gas delivered by anaesthesia systems and ventilators

- Hot-wire flow sensor with use with Dräger ventilators.
- Replacement for the Dräger Spirolog, Anemone, PM8030, PM8060, Eva, Cierco, Cato, Julian, Evita 4, Evita XL, Primus...
- Supplied in a pack of five, packed individually in sterilised packaging, and are for single patient use.



Instructions for use

- Alignment, operation and sensor replacement must be carried out in accordance with the relevant device-specific instruction manual.
- The sensor cannot be autoclaved.
- The sensor can be sterilised with ETO before usage.
- Do NOT clean inside of the sensor in a mechanical way, with compressed air or jet of water because this will lead to damage of the thin measuring wires.
- Several disinfection cycles with Ethanol solution 70% are possible, afterwards leave to dry in air.
- If there is a possibility that the sensor is still contaminated, a new one must be fitted.
- **Life cycle of the sensor is limited. It can be used as long as calibration in the ventilator is possible.**



VIAMED

Viamed Limited · 15 Station Road · Cross Hills
Keighley · West Yorkshire · BD20 7DT · United Kingdom
Tel: +44 (0)1535 634 542 / 636 757
Fax: +44 (0)1535 635 582
Email: info@viamed.co.uk Website: www.viamed.co.uk