



Teledyne Analytical Instruments

CID3988 Opera **00009147** Type:**Company Supplier**
Customer
Teledyne Analytical Instruments 16830 Chestnut Street, ,
 City Of Industry, California, 91748-1020, USA
 Tel **0016269341589**  Log Call



Teledyne Analytical Instruments - Returns Status Page

Summary Goods Waiting to Return to Teledyne Analytical Instruments

Stock Ref	Description	Total To Return Supplier
`0110017	Oxygen sensor R-17MED	41
`0110051	Oxygen sensor T-1	3
`0110022	Oxygen sensor R-22MED	4
`0110024	Oxygen sensor R-24MED	2
`0110122	Oxygen sensor - Automotive R-22A	2
`0110117	Oxygen sensor - Automotive R-17A	2
`0110127	Oxygen sensor - Automotive R-17AV	1
`0110137	Oxygen sensor - Automotive R-17AH	1
`0110361	Oxygen sensor R-22AHJR Jikco OEM sensor 2	

Summary Goods Back at Teledyne Analytical Instruments

Stock Ref	Description	Total With Supplier
0110081	Oxygen sensor UFO 130-2	88
0110024	Oxygen sensor R-24MED	66
0110017	Oxygen sensor R-17MED	19
0110022	Oxygen sensor R-22MED	26
0110051	Oxygen sensor T-1	1
0110057	Oxygen sensor T-7	1

0110023	Oxygen sensor R-23V	1
0110122	Oxygen sensor - Automotive R-22A	5
0110137	Oxygen sensor - Automotive R-17AH	3
0110361	Oxygen sensor R-22AHJR Jikco OEM sensor 2	

All Outstanding Items for Teledyne Analytical Instruments Supplier return File

Count	Return Box	RMA	Date Shipped to Supplier	Stock Reference	Description	Serialnumber	Fault
1	BOX778		29 Dec 2021	0110081	Oxygen sensor UFO 130-2	479035	Untested SRS67687
2	BOX778		29 Dec 2021	0110081	Oxygen sensor UFO 130-2	500812	Untested SRS67687
3	BOX781		29 Dec 2021	0110024	Oxygen sensor R-24MED	609181	Output (10mV - 17mV) ERRATIC n Move Item to location 126561 n Failed in Q.A. n Did not reach end user
4	BOX781		29 Dec 2021	0110024	Oxygen sensor R-24MED	609191	Output (10mV - 17mV) ERRATIC n Move Item to location 126561 n Failed in Q.A. n Did not reach end user
5	BOX781		29 Dec 2021	0110024	Oxygen sensor R-24MED	609187	Output (10mV - 17mV) ERRATIC n Move Item to location 126561 n Failed in Q.A. n Did not reach end user
6	BOX781		29 Dec 2021	0110024	Oxygen sensor R-24MED	608460	Output (10mV - 17mV) ERRATIC n Move Item to location 126561 n Failed in Q.A. n Did not reach end user
7	BOX781		29 Dec 2021	0110024	Oxygen sensor R-24MED	608476	Output (10mV - 17mV) ERRATIC n Move Item to location 126561 n Output (10mV - 17mV) ERRATIC n Move Item to location 126561 n Failed in Q.A. n Did not reach end user
8	BOX781		29 Dec 2021	0110024	Oxygen sensor R-24MED	608486	Output (10mV - 17mV) erratic n Move Item to location 126561 n Failed in Q.A. n Did not reach end user
9	BOX781		29 Dec 2021	0110024	Oxygen sensor R-24MED	609184	Output (10mV - 17mV) ERRATIC n Move Item to location 126561 n Failed in Q.A. n Did not reach end user
10	BOX781		29 Dec 2021	0110024	Oxygen sensor R-24MED	608458	Output (10mV - 17mV) ERRATIC n Move Item to location 126561 n Failed in Q.A. n Did not reach end user
11	BOX781		29 Dec 2021	0110024	Oxygen sensor R-24MED	608464	Output (10mV - 17mV) ERRATIC n Move Item to location 126561 n Failed in Q.A. n Did not reach end user
12	BOX781		29 Dec 2021	0110024	Oxygen sensor R-24MED	608474	Output (10mV - 17mV) ERRATIC n Move Item to location 126561 n Failed in Q.A. n Did not reach end user