

Invoice Correction 5259448773

ORIGINAL

Invoice Correction Date: 17.03.2022

Honeywell Healthcare Solutions GmbH

Alter Holzhafen 18

Purchase Order No: RTS3100852

23966 Wismar

Germany

Bill to: 1797448
 Vandagraph Sensor Technologies Ltd.
 15 Station Road
 Cross Hills, West Yorkshire
 KEIGHLEY
 BD20 7DT
 United Kingdom

Sold to: 1797448
 Vandagraph Sensor Technologies Ltd.
 15 Station Road
 Cross Hills, West Yorkshire
 KEIGHLEY
 BD20 7DT
 United Kingdom

Order Information:

Credit Memo Request: 63837659
 Sales Office: 2170 DE SPS Wismar
 Incoterms: FCA Origin
 Payment Terms: Net 30 Days After Invoice Date

Ship to: 1797448

Vandagraph Sensor Technologies Ltd.
 15 Station Road
 Cross Hills, West Yorkshire
 KEIGHLEY
 BD20 7DT
 United Kingdom

ITEM NO.	Material No.	QTY	UOM	UNIT PRICE	EXT. PRICE	VAT %
000010	E1002226 Oxygen Sensor OOD103-APD16 Customer Part Nbr: 8010022 credit note for SN: 120075 121132	2	EA	20,00	-40,00	0%

Totals

Total Before VAT -40,00 EUR
 VAT-Total 0,00 EUR
 Total Amount -40,00 EUR

Tax Information

Tax Exempt Reason: Tax free export - Goods

Integrity and Compliance at Honeywell

Honeywell places the highest value on the integrity of its business operations. We expect our employees and business partners to comply with our Code of Business Conduct and all applicable laws and regulations. If you become aware of illegal or improper conduct by a Honeywell employee please contact your Honeywell relationship manager. You can also report misconduct directly to us either via e-mail to our Integrity and Compliance Office at access.integrity.helpline@honeywell.com or via our Access helpline accessible worldwide at 800-237-5982.

Duns #:

Bank Details: No Payment Due

Account Name:

Account No:

Bank Key:

Swift Code:

IBAN:

Payment method:

Registered No: Schwerin HRB 1838

VAT Reg No: DE137437213

Federal ID #:

Registered Address: Honeywell Healthcare Solutions GmbH, Alter Holzhafen 18, Wismar, 23966, Germany

Credit Analyst: Xenia Musat

Sales Rep: Stefan Dreyer

Customer Care Rep: Jessica Wagner +49 3841 360 206

Honeywell Healthcare Solutions GmbH

WEEE Reg.No.: 57798680

Managing Director: Marcus Ostländer

Internet: www.envitec.comemail: info@envitec.com

Delivery exclusive to our General Terms (February 2016). We retain title to our goods and deliveries until all the payment from the sales contract have been received. Your order will be executed exclusively to our terms and conditions. These can be sent by separate request or viewed at www.envitec.com.

PRÜFBERICHT – TEST REPORT

Sauerstoffsensoren/Oxygen Sensors OOA, OOI, OOD

Kunde/ Customer	Vadagraph	Retour/ RMA- No.	Kunde/Customer	309250 133	EnvieC
Artikel/ Article	093 103 -APD 16	Anzahl/ Quantity	2	Eing. Datum/ Inc. Date	01.03.22
Bearbeitet/ processed	U. Bue	Prüfdatum/ Insp. Date	03.03.22	Geprüft/ Inspected	Bue

Produkt/ Identification		Messwert/ Test Result				Bewertung/ Evaluation				
Nr.	Seriennummer/ Serial no.	Lieferlos/ Manufact. code	Luftwert/ Air signal @ 20,9% O ₂ [mV]	Nullsignal/ Offset @ 0% O ₂ [µV]	Linearität/ Linearity [%]	Bemerkung/ Remark	außerhalb Garantie/ out of warranty	kein Fehler festgestellt/ no failure found	Fehlercode/ Failure- code	Austausch / Replacement
1	120075	03121	0						04	
2	121132	12121	12.4	38	fehlerhaft				02	
3										
4										
5										
6										
7										
8										
9										
10										
11										
12										
13										
14										
15										
16										
17										
18										
19										
20										

Fehler O2-S001: Luftwert außerhalb Toleranz/ Air signal out of tolerance
 Fehler O2-S003: Offset außerhalb Toleranz/ Offset out of tolerance
 Fehler O2-S005: Gehäuse defekt/ Housing defective
 Fehler O2-S007: Sonstiges (unter Angabe in Bemerkung) / Others (see remark)

Fehler O2-S002: Signal instabil / Signal not stable
 Fehler O2-S004: Signal 0mV / Signal 0mV
 Fehler O2-S006: Elektrolyt Austritt / Electrolyte leakage