
Tom Thumb Infant Resuscitator.

Part Number: 0310030

Performance Evaluation

Test Date	15/Dec/2010 9:35 am		
Serial number	VMA0109		
		Pass	Fail
Check Ports / Fittings for Leaks	Yes		
Adj. Valve @ 10Lpm Min reading <8 cmH2O	Yes		
Adj. Valve @ 10Lpm Min reading 42<->47cmH2O	42.4		
Prec. Valve 2+ cmH2O Over Adj. Value & <51 cmH2O	45.0		
Gauge Test @ 30 cmH2O	29.2		
Gauge cannot be Remove with out tools	Yes		
Gauge appears straight	Yes		
All Settings set to minimum	Yes		
Labels are attached	Yes		
Calibrated Manometer Barcode ID	102075		

Performance evaluation: PASS

Signed: _____

Name: Philip Crossley