



VIAMED Ltd.

15 Station Road, Cross Hills, Keighley, West Yorkshire, BD20 7DT, UK.

Website: www.viamed.co.uk. Email: info@viamed.co.uk.

Tel: +44 (0)1535 634542. Fax: +44 (0)1535 635582.

EC Declaration of Conformity

We hereby declare under sole responsibility that the product

SpO2 Finger Sensor - Adult. SF4013 GE DATEX Compatible

For use with pulse oximeters in order to determine saturation of haemoglobin non-invasively from light signals of two wavelengths transmitted through tissue or reflected from tissues

Product No(s).

0014625

conforms with the essential requirements of Annex I of the Council Directive 93/42/EEC of 14 June 1993 concerning medical devices considering the amendments by the directive 2007/47/EC.

In accordance with Annex IX of the Directive 93/42 EEC considering the amendments by the directive 2007/47/EC the product has been classified as

Class IIb

Application of the CE-Marking:

CE0086

Issuer:

Viamed Ltd.

15 Station Road

Cross Hills

Keighley

West Yorkshire, BD20 7DT

United Kingdom

Place, Date:

Keighley , 5 Nov 2010.

Legally binding signature:

.....
Derek Lamb (Managing Director)