

Viamed Returns Shipping Invoice

Delivery Address:
Teledyne Analytical Instruments
Margie Read
16830 Chestnut Street
City Of Industry
California
91748-1020

Date: 30 Sep 2022
RMA:
Viamed RN: RTS1100874

Shipper Address:
Viamed
15 Station Road
Cross Hills
Keighley
West Yorkshire
BD20 7DT
+44 (0) 1535 634542
+44 (0) 1535 635582
info@viamed.co.uk

Returns Under Warranty Value for Customs : \$100.00

See Attached Listing for Serial numbers and Faults

Part Number	P/N / Value	Description	Quantity
C43690-R17MED	0110017 (\$22.00 Ea.) Tariff Code: 9019209000 Country of Origin: USA	Teledyne Oxygen sensor R-17MED	1
C44611-R22MED	0110022 (\$25.00 Ea.) Tariff Code: 9019200000 Country of Origin: USA	Teledyne Oxygen sensor R-22MED	3
C43981-R21A	0110121 (\$16.00) Tariff Code: HTS Code 9027.10.2000 Country of Origin: USA	Automotive Oxygen Sensor R-21A	9

Part Number	Serial Number	RMA	Fault Description
C43690-R17MED	617320		Unstable
C44611-R22MED	622758		Non Linear
C44611-R22MED	622759		Unstable
C44611-R22MED	624826		Unstable
C43981-R21A	608262		High Output
C43981-R21A	608267		High Output
C43981-R21A	608269		No Output
C43981-R21A	608272		High Output
C43981-R21A	608281		Unstable
C43981-R21A	608361		Non Linear
C43981-R21A	608371		Unstable
C43981-R21A	608373		Non Linear
C43981-R21A	669463		Non Linear