

# Viamed Returns Shipping Invoice

Delivery Address:	Date: 30 Sep 2022	Shipper Address:
Teledyne Analytical Instruments	RMA:	Viamed
Margie Read	Viamed RN: RTS1100884	15 Station Road
16830 Chestnut Street		Cross Hills
City Of Industry		Keighley
California		West Yorkshire
91748-1020		BD20 7DT
0016269341589		+44 (0) 1535 634542
		+44 (0) 1535 635582
		info@viamed.co.uk

Returns Under Warranty Value for Customs : \$100.00

See Attached Listing for Serial numbers and Faults

Part Number	P/N / Value	Description	Quantity
C43690-R17A-LV	0110107 (\$16.00)	Automotive Oxygen Sensor R-17A	29
	Tariff Code: HTS Code 9027.10.2000		
	Country of Origin: USA		
C44611-R22A	0110122 (\$16.00)	Automotive Oxygen Sensor R-22A	2
	Tariff Code: HTS Code 9027.10.2000		
	Country of Origin: USA		
C43690-R17AV	0110127 (\$16.00)	Automotive Oxygen Sensor R-17A	1
	Tariff Code: HTS Code 9027.10.2000		
	Country of Origin: USA		
C43690-R17AH	0110137 (\$16.00)	Automotive Oxygen Sensor R-17A	3
	Tariff Code: HTS Code 9027.10.2000		
	Country of Origin: USA		
C44611-R22AHJR	0110361 (\$16.50)	Oxygen sensor R-22AHJR Jikco	22
	Tariff Code: HTS Code 9027.10.2000		
	Country of Origin: USA		

Part Number	Serial Number	RMA	Fault Description
C43690-R17A-LV	843983		Output ( 7mv - 11.5mv) 11.8 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	843984		Output ( 7mv - 11.5mv) 12.5 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	843988		Output ( 7mv - 11.5mv) 12.0 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	843989		Output ( 7mv - 11.5mv) 2.2 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	843994		Output ( 7mv - 11.5mv) 12.2 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	843996		Output ( 7mv - 11.5mv) 11.7 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	843997		Output ( 7mv - 11.5mv) 12.0 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	843999		Output ( 7mv - 11.5mv) 12.2 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	844000		Output ( 7mv - 11.5mv) 11.9 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	844002		Output ( 7mv - 11.5mv) 12.0 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	844005		Output ( 7mv - 11.5mv) 11.9 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	844008		Output ( 7mv - 11.5mv) 11.9 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	844011		Output ( 7mv - 11.5mv) 00.7 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	844012		Pass Yes n Move Item to location 126541 n Output ( 7mv - 11.5mv) 12.0 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	844017		Output ( 7mv - 11.5mv) 12.1 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	844019		Output ( 7mv - 11.5mv) 11.7 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	844022		Output ( 7mv - 11.5mv) 2.8 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	844029		output 12.4mV failed qa did not reach end user
C43690-R17A-LV	844037		Output ( 7mv - 11.5mv) 11.7 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	844038		Output ( 7mv - 11.5mv) 11.7 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	844039		Output ( 7mv - 11.5mv) 11.7 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	844045		Output ( 7mv - 11.5mv) 12.2 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	844047		Output ( 7mv - 11.5mv) 12.0 n Move Item to location 126541 n Output ( 7mv - 11.5mv) 11.9 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	844049		Output ( 7mv - 11.5mv) 11.9 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	844055		Output ( 7mv - 11.5mv) 11.6 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	844060		Output ( 7mv - 11.5mv) 12.3 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	844065		Output ( 7mv - 11.5mv) 12.3 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17A-LV	844072		Output ( 7mv - 11.5mv) 11.8 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user

Part Number	Serial Number	RMA	Fault Description
C43690-R17A-LV	844073		Output ( 7mv - 11.5mv) 11.7 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C44611-R22A	824470		Output ( 7mV - 13mV ) 4.5 n Move Item to location 126552 n Failed in Q.A. n Did not reach end user
C44611-R22A	832094		Output ( 7mV - 13mV ) 00.1 n Move Item to location 126552 n Failed in Q.A. n Did not reach end user
C43690-R17AV	844120		Output ( 9mV - 13mV ) 00.1 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17AH	843753		Output ( 10.5mV - 13.5mV ) 00.6 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17AH	843781		Output ( 10.5mV - 13.5mV ) 1.6 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C43690-R17AH	843818		Output ( 10.5mV - 13.5mV ) 1.1 n Move Item to location 126541 n Failed in Q.A. n Did not reach end user
C44611-R22AHJR	A838455		Output ( 10.5 - 13.5 mV ) 14.3 n 126550 n Failed in Q.A. n Did not reach end user
C44611-R22AHJR	A838459		Output ( 10.5 - 13.5 mV ) 14.0 n 126550 n Failed in Q.A. n Did not reach end user
C44611-R22AHJR	A838464		Output ( 10.5 - 13.5 mV ) 14.5 n 126550 n Failed in Q.A. n Did not reach end user
C44611-R22AHJR	A838465		Output ( 10.5 - 13.5 mV ) 15.4 n 126550 n Failed in Q.A. n Did not reach end user
C44611-R22AHJR	A838466		Output ( 10.5 - 13.5 mV ) 13.6 pressure 997 MB n 126550 n Failed in Q.A. n Did not reach end user
C44611-R22AHJR	A838491		Output ( 10.5 - 13.5 mV ) 13.5 n Pass Yes n 126550 n Output ( 10.5 - 13.5 mV ) 13.7 Pressure 997 Mb n Pass Yes n 126550 n Output ( 10.5 - 13.5 mV ) 13.7 Pressure 997 Mb n 126550 n Output ( 10.5 - 13.5 mV ) 13.7 Pressure 997 Mb n 126550 n Failed in Q.A. n Did not reach end user
C44611-R22AHJR	A838500		Output ( 10.5 - 13.5 mV ) 15.3 n 126550 n Failed in Q.A. n Did not reach end user
C44611-R22AHJR	A838505		Output ( 10.5 - 13.5 mV ) 14.0 n 126550 n Failed in Q.A. n Did not reach end user
C44611-R22AHJR	A838506		Output ( 10.5 - 13.5 mV ) 13.8 n 126550 n Failed in Q.A. n Did not reach end user
C44611-R22AHJR	A838517		Output ( 10.5 - 13.5 mV ) 14.0 n 126550 n Failed in Q.A. n Did not reach end user
C44611-R22AHJR	A838521		Output ( 10.5 - 13.5 mV ) 13.8 n Pass Yes n 126550 n Output ( 10.5 - 13.5 mV ) 13.8 n 126550 n Failed in Q.A. n Did not reach end user
C44611-R22AHJR	A838535		Output ( 10.5 - 13.5 mV ) 13.9 n 126550 n Failed in Q.A. n Did not reach end user
C44611-R22AHJR	A838537		Output ( 10.5 - 13.5 mV ) 13.7 Pressure 997 MB n 126550 n Failed in Q.A. n Did not reach end user
C44611-R22AHJR	A838540		Output ( 10.5 - 13.5 mV ) 15.0 n 126550 n Failed in Q.A. n Did not reach end user
C44611-R22AHJR	A838547		Output ( 10.5 - 13.5 mV ) 14.4 n 126550 n Failed in Q.A. n Did not reach end user
C44611-R22AHJR	A838564		Output ( 10.5 - 13.5 mV ) 14.5 n 126550 n Failed in Q.A. n Did not reach end user
C44611-R22AHJR	A838575		Output ( 10.5 - 13.5 mV ) 15.5 n 126550 n Failed in Q.A. n Did not reach end user
C44611-R22AHJR	A838595		Output ( 10.5 - 13.5 mV ) 15.0 n 126550 n Failed in Q.A. n Did not reach end user
C44611-R22AHJR	A838620		Output ( 10.5 - 13.5 mV ) 13.7 n 126550 n Failed in Q.A. n Did not reach end user
C44611-R22AHJR	A838632		Output ( 10.5 - 13.5 mV ) 13.5 Pressure 997 Mb n 126550 n Failed in Q.A. n Did not reach end user
C44611-R22AHJR	A838636		Output ( 10.5 - 13.5 mV ) 13.4 Pressure 997 Mb n 126550 n Failed in Q.A. n Did not reach end user

Part Number	Serial Number	RMA	Fault Description
C44611-R22AHJR	A838638		Output ( 10.5 - 13.5 mV ) 13.8 n 126550 n Failed in Q.A. n Did not reach end user