

Viamed Returns Shipping Invoice

Delivery Address:
Teledyne Analytical Instruments
Margie Read
16830 Chestnut Street
City Of Industry
California
91748-1020
0016269341589

Date: 30 Sep 2022
RMA:
Viamed RN: RTS1100892

Shipper Address:
Viamed
15 Station Road
Cross Hills
Keighley
West Yorkshire
BD20 7DT
+44 (0) 1535 634542
+44 (0) 1535 635582
info@viamed.co.uk

Returns Under Warranty Value for Customs : \$100.00
See Attached Listing for Serial numbers and Faults

Part Number	P/N / Value	Description	Quantity
C44611-R22A	0110122 (\$16.00)	Automotive Oxygen Sensor R-22A	6
	Tariff Code: HTS Code 9027.10.2000		
	Country of Origin: USA		
C43690-R17S1	0110237	R-17S1 Oxygen Sensor	1
	Tariff Code: HTS Code 9027.90.58.10		

Part Number	Serial Number	RMA	Fault Description
C44611-R22A	606401		High output and non linear
C44611-R22A	663097		Non Linear
C44611-R22A	774334		Non Linear
C44611-R22A	774516		Non Linear
C44611-R22A	793095		Low Output
C44611-R22A	822424		Non Linear
C43690-R17S1	685451		High Output