

Viamed Returns Shipping Invoice

Delivery Address:

Teledyne Analytical Instruments
Vasu Narasimhan
16830 Chestnut Street

91748-1020
0016269341589

Date: 05 Oct 2022

RMA:
Viamed RN: RTS1100812

Shipper Address:

Viamed
15 Station Road
Cross Hills
Keighley
West Yorkshire
BD20 7DT
+44 (0) 1535 634542
+44 (0) 1535 635582
info@viamed.co.uk

Returns Under Warranty Value for Customs : \$100.00
See Attached Listing for Serial numbers and Faults

Part Number	P/N / Value	Description	Quantity
C43690-R17MED	0110017 (\$22.00 Ea.)	Teledyne Oxygen sensor R-17MED	7
	Tariff Code: 9019209000		
	Country of Origin: USA		

Part Number	Serial Number	RMA	Fault Description
C43690-R17MED	646428		Output (7mV - 13mV) 12.4 n Pass Yes n Move Item to location 126559 n Output (7mV - 13mV) 13.3 n Move Item to location 126559 n Failed in Q.A. n Did not reach end user
C43690-R17MED	646438		Output (7mV - 13mV) 13.4 n Move Item to location 126559 n Output (7mV - 13mV) 13.4 n Move Item to location 126559 n Failed in Q.A. n Did not reach end user
C43690-R17MED	646440		Output (7mV - 13mV) 13.4 n Move Item to location 126559 n Failed in Q.A. n Did not reach end user
C43690-R17MED	646457		Output (7mV - 13mV) 13.5 n Move Item to location 126559 n Failed in Q.A. n Did not reach end user
C43690-R17MED	646461		Output (7mV - 13mV) 13.2 sensor doesnt fit on tester very loose. n Move Item to location 126559 n Failed in Q.A. n Did not reach end user
C43690-R17MED	646467		Output (7mV - 13mV) 1823656 n Pass Yes n Move Item to location 126559 n Output (7mV - 13mV) 13.6 n Move Item to location 126559 n Failed in Q.A. n Did not reach end user
C43690-R17MED	646471		Output (7mV - 13mV) 00.1 n Move Item to location 126559 n Failed in Q.A. n Did not reach end user