

Viamed Returns Shipping Invoice

Delivery Address:
Teledyne Analytical Instruments
Vasu Narasimhan
16830 Chestnut Street

91748-1020
0016269341589

Date: 30 Sep 2022
RMA:
Viamed RN: RTS1100812

Shipper Address:
Viamed
15 Station Road
Cross Hills
Keighley
West Yorkshire
BD20 7DT
+44 (0) 1535 634542
+44 (0) 1535 635582
info@viamed.co.uk

Returns Under Warranty Value for Customs : \$100.00
See Attached Listing for Serial numbers and Faults

Part Number	P/N / Value	Description	Quantity
C43690-R17MED	0110017 (\$22.00 Ea.) Tariff Code: 9019209000 Country of Origin: USA	Teledyne Oxygen sensor R-17MED	13
C44611-R22MED	0110022 (\$25.00 Ea.) Tariff Code: 9019200000 Country of Origin: USA	Teledyne Oxygen sensor R-22MED	9
0110023	0110023 Tariff Code: 9019200000	Oxygen sensor R-23V	1

Part Number	Serial Number	RMA	Fault Description
C43690-R17MED	465227		Unstable
C43690-R17MED	524654		Fluctuating O/p - During Use
C43690-R17MED	524655		Fluctuating O/p - During Use
C43690-R17MED	554239		Non Linear
C43690-R17MED	571933		Connector Fault
C43690-R17MED	601211		Connector Fault
C43690-R17MED	646428		Output (7mV - 13mV) 12.4 n Pass Yes n Move Item to location 126559 n Output (7mV - 13mV) 13.3 n Move Item to location 126559 n Failed in Q.A. n Did not reach end user
C43690-R17MED	646438		Output (7mV - 13mV) 13.4 n Move Item to location 126559 n Output (7mV - 13mV) 13.4 n Move Item to location 126559 n Failed in Q.A. n Did not reach end user
C43690-R17MED	646440		Output (7mV - 13mV) 13.4 n Move Item to location 126559 n Failed in Q.A. n Did not reach end user
C43690-R17MED	646457		Output (7mV - 13mV) 13.5 n Move Item to location 126559 n Failed in Q.A. n Did not reach end user
C43690-R17MED	646461		Output (7mV - 13mV) 13.2 sensor doesnt fit on tester very loose. n Move Item to location 126559 n Failed in Q.A. n Did not reach end user
C43690-R17MED	646467		Output (7mV - 13mV) 1823656 n Pass Yes n Move Item to location 126559 n Output (7mV - 13mV) 13.6 n Move Item to location 126559 n Failed in Q.A. n Did not reach end user
C43690-R17MED	646471		Output (7mV - 13mV) 00.1 n Move Item to location 126559 n Failed in Q.A. n Did not reach end user
C44611-R22MED	473358		Unstable
C44611-R22MED	582573		Output (7mV - 13mV) 12.9 n Pass Yes n Move Item to location 126560 n Output (7mV - 13mV) 14.0 n Move Item to location 126560 n Failed in Q.A. n Did not reach end user
C44611-R22MED	589356		Output (7mV - 13mV) 11.0 n Pass Yes n Move Item to location 126560 n Output (7mV - 13mV) 00.3 n Move Item to location 126560 n Failed in Q.A. n Did not reach end user
C44611-R22MED	589364		Output (7mV - 13mV) 12.7 n Pass Yes n Move Item to location 126560 n Output (7mV - 13mV) 13.6 n Move Item to location 126560 n Failed in Q.A. n Did not reach end user
C44611-R22MED	589378		Output (7mV - 13mV) 9.6 n Pass Yes n Move Item to location 126560 n Output (7mV - 13mV) 00.3 n Move Item to location 126560 n Failed in Q.A. n Did not reach end user
C44611-R22MED	589382		Output (7mV - 13mV) 10.5 n Pass Yes n Move Item to location 126560 n Output (7mV - 13mV) 00.7 n Move Item to location 126560 n Failed in Q.A. n Did not reach end user
C44611-R22MED	589385		Output (7mV - 13mV) 9.9 n Pass Yes n Move Item to location 126560 n Output (7mV - 13mV) 5.6 n Move Item to location 126560 n Output (7mV - 13mV) 5.5 n Move Item to location 126560 n Failed in Q.A. n Did not reach end user
C44611-R22MED	589419		Output (7mV - 13mV) 10.0 n Pass Yes n Move Item to location 126560 n Output (7mV - 13mV) 01.4 n Move Item to location 126560 n Failed in Q.A. n Did not reach end user
C44611-R22MED	597205		Unstable
0110023	V121059		Unstable