

Viamed Returns Shipping Invoice

Delivery Address:
Teledyne Analytical Instruments
Margie Read
16830 Chestnut Street
City of Industry
California
91748-1020
0016269341589

Date: 30 Sep 2022
RMA:
Viamed RN: RTS1100861

Shipper Address:
Viamed
15 Station Road
Cross Hills
Keighley
West Yorkshire
BD20 7DT
+44 (0) 1535 634542
+44 (0) 1535 635582
info@viamed.co.uk

Returns Under Warranty Value for Customs : \$100.00

See Attached Listing for Serial numbers and Faults

Part Number	P/N / Value	Description	Quantity
C43690-R17MED	0110017 (\$22.00 Ea.) Tariff Code: 9019209000 Country of Origin: USA	Teledyne Oxygen sensor R-17MED	10
C44611-R22MED	0110022 (\$25.00 Ea.) Tariff Code: 9019200000 Country of Origin: USA	Teledyne Oxygen sensor R-22MED	1
C61340-R24MED	0110024 (\$25.00) Tariff Code: 9019200000 Country of Origin: USA	Teledyne Oxygen sensor R-24MED	2

Part Number	Serial Number	RMA	Fault Description
C43690-R17MED	566049		High Output
C43690-R17MED	593283		Non Linear
C43690-R17MED	593285		Non Linear
C43690-R17MED	605930		Non Linear
C43690-R17MED	606017		Non Linear
C43690-R17MED	617208		High Output
C43690-R17MED	617216		No Output
C43690-R17MED	617289		Non Linear
C43690-R17MED	646335		Manufacturing Error
C43690-R17MED	646336		Unstable
C44611-R22MED	597204		Unstable
C61340-R24MED	608495		No Output
C61340-R24MED	609211		Non Linear