

Viamed Returns Shipping Invoice

Delivery Address:	Date: 30 Sep 2022	Shipper Address:
Teledyne Analytical Instruments	RMA:	Viamed
Margie Read	Viamed RN: RTS1100863	15 Station Road
16830 Chestnut Street		Cross Hills
City Of Industry		Keighley
California		West Yorkshire
91748-1020		BD20 7DT
		+44 (0) 1535 634542
		+44 (0) 1535 635582
		info@viamed.co.uk

Returns Under Warranty Value for Customs : \$100.00
See Attached Listing for Serial numbers and Faults

Part Number	P/N / Value	Description	Quantity
C43690-R17MED	0110017 (\$22.00 Ea.) Tariff Code: 9019209000 Country of Origin: USA	Teledyne Oxygen sensor R-17MED	6
C44611-R22MED	0110022 (\$25.00 Ea.) Tariff Code: 9019200000 Country of Origin: USA	Teledyne Oxygen sensor R-22MED	3

Part Number	Serial Number	RMA	Fault Description
C43690-R17MED	701320		Output (7mV - 13mV) 00.1 Failed in Q.A. n Did not reach end user
C43690-R17MED	773830		Output (7mV - 13mV) 00.1 Failed in Q.A. n Did not reach end user
C43690-R17MED	773969		Output (7mV - 13mV) 00.1 Failed in Q.A. n Did not reach end user
C43690-R17MED	774192		Output (7mV - 13mV) 00.1 Failed in Q.A. n Did not reach end user
C43690-R17MED	781987		Output (7mV - 13mV) 5.8 Failed in Q.A. n Did not reach end user
C43690-R17MED	782035		Output (7mV - 13mV) 00.1 Failed in Q.A. n Did not reach end user
C44611-R22MED	531596		Output (7mV - 13mV) 0.9 Failed in Q.A. n Did not reach end user
C44611-R22MED	786565		Output (7mV - 13mV) 5.0 Failed in Q.A. n Did not reach end user
C44611-R22MED	786606		Output (7mV - 13mV) 1.6 Failed in Q.A. n Did not reach end user