

Finger Pulse Oximeter

High clinical accuracy and robust

Fast and accurate measurement of SpO_2 and pulse rate, even in cases of low perfusion.

VM-2101
NHS: FBC206
NATO: 6515-99-148-5212



Actual Size



- Display available in 4 rotations.
- Colour coded signal quality and signal strength indicator.
- Adjustable brightness; 6 levels.
- Accuracy $A_{\text{RMS}} = 2.0\%$
- $\pm 1 \text{ S.D.} = 1.86\%$



Clinical Validation



ISO 9919:2009



Colour OLED Display



VIAMED

Viamed Limited · 15 Station Road · Cross Hills
Keighley · West Yorkshire · BD20 7DT
Tel: 01535 633 705 Fax: 01535 635 582
Email: life@viamed.co.uk
Website: www.viamed.co.uk