

Test Report OOM108 Oxygen Sensor (3)

Test Place:	IRI PICU	Tested by:	<i>J. H. Brinser</i>
Ventilator Master Data			
Ventilator type:	Servo-i	Ventilator Serial Number:	13031
Ventilator Software Version:	V 8.00.01	Ventilator Manufacturing Date:	purchased Mar 2015
Sensor Master Data (currently used / original one)			
Sensor type:	Getinge O ₂ cell	Sensor Part Number:	6640044
Sensor Serial Number:	AA125857	Others:	
Sensor Master Data (tested sensor)			
Sensor type:	OOM108	Sensor Part Number:	E01-00-0075
Sensor Serial Number:	AA600017	Others:	

Test Assignment

Check of oxygen sensor OOM108 in MAQUET ventilator regarding to compatibility, usability, installation, calibration and measurement accuracy.

Checklist

Pos.	Check Item	Yes	No	Comments
1	Sensor usability Sensor usability is same as the original one?	✓		
2	Sensor Installation Sensor installation is same as the original one?	✓		
3	Sensor calibration Sensor calibration was successful in the ventilator?	✓		
4	Oxygen Measurement Sensor works without issues in ventilator?	✓		

Final Conclusion:

The oxygen sensor type OOM108 tested in MAQUET ventilator is compatible to the currently used oxygen sensor made by MAQUET.

Yes No

Date: 25/9/23

Signature: *J. H. Brinser*