

**Test Report OOM108 Oxygen Sensor**

Test Place:	CMIAH	Tested by:	Richard
<b>Ventilator Master Data</b>			
Ventilator type:	Servo i	Ventilator Serial Number:	34122
Ventilator Software Version:	4.4	Ventilator Manufacturing Date:	2010
<b>Sensor Master Data (currently used / original one)</b>			
Sensor type:		Sensor Part Number:	
Sensor Serial Number:	A A600005	Others:	
<b>Sensor Master Data (tested sensor)</b>			
Sensor type:	OOM108	Sensor Part Number:	
Sensor Serial Number:		Others:	

**Test Assignment**

Check of oxygen sensor OOM108 in MAQUET ventilator regarding to compatibility, usability, installation, calibration and measurement accuracy.

**Checklist**

Pos.	Check Item	Yes	No	Comments
1	<b>Sensor usability</b> Sensor usability is same as the original one?	X		
2	<b>Sensor installation</b> Sensor installation is same as the original one?	X		
3	<b>Sensor calibration</b> Sensor calibration was successful in the ventilator?	X		
4	<b>Oxygen Measurement</b> Sensor works without issues in ventilator?	X		

**Final Conclusion:**

The oxygen sensor type OOM108 tested in MAQUET ventilator is compatible to the currently used oxygen sensor made by MAQUET.

Yes  No

Date: 9/5/2023

Signature:



	Revision:	Created – Erstellt:	Page – Seite:
C:\Users\E339512\AppData\Local\Microsoft\Windows\NetCache\Content.Outlook\YEPULFMN\OOM108_compatibility_check-0.docx		MS	1 / 1