

Test Report OOM108 Oxygen Sensor

Test Place:	Binh Duong General Hospital	Tested by:	An An Import Export Co. Ltd.
Ventilator Master Data			
Ventilator type:	Servo - Air	Ventilator Serial Number:	15806
Ventilator Software Version:	4.4	Ventilator Manufacturing Date:	2019-10-28.
Sensor Master Data (currently used / original one)			
Sensor type:	Maquet	Sensor Part Number:	6640044
Sensor Serial Number:	AA 438 582	Others:	
Sensor Master Data (tested sensor)			
Sensor type:	OOM108	Sensor Part Number:	0110590
Sensor Serial Number:	AA 6 00008	Others:	

Test Assignment

Check of oxygen sensor OOM108 in MAQUET ventilator regarding to compatibility, usability, installation, calibration and measurement accuracy.

Checklist

Pos.	Check Item	Yes	No	Comments
1	Sensor usability Sensor usability is same as the original one?	✓		
2	Sensor installation Sensor installation is same as the original one?	✓		
3	Sensor calibration Sensor calibration was successful in the ventilator?	✓		
4	Oxygen Measurement Sensor works without issues in ventilator?	✓		

Final Conclusion:

The oxygen sensor type OOM108 tested in MAQUET ventilator is compatible to the currently used oxygen sensor made by MAQUET.

Yes

No

Đỗ Văn Thành
August 2, 2023

Date:


Signature:

File	Revision:	Created – Erstellt:	Page – Seite:
C:\Users\E339512\AppData\Local\Microsoft\Windows\INetCache\Content.Outlook\YEPULFMN\OOM108_compatibility_check-0.docx		MS	1 / 1

Comparison between Original sensor and OOM108 sensor



Sensor Installation



Sensor information is read by the machine

