

Test Report OOM108 Oxygen Sensor

Test Place:	Miah Anh International Hospital		Tested by:	An An Import Export Co. Ltd
Ventilator Master Data				
Ventilator type:	Servo - Air.		Ventilator Serial Number:	22153
Ventilator Software Version:			Ventilator Manufacturing Date:	
Sensor Master Data (currently used / original one)				
Sensor type:	Maquet		Sensor Part Number:	6640044
Sensor Serial Number:	AA450833		Others:	
Sensor Master Data (tested sensor)				
Sensor type:	OOM108		Sensor Part Number:	0110590
Sensor Serial Number:	AA600009		Others:	

Test Assignment

Check of oxygen sensor OOM108 in MAQUET ventilator regarding to compatibility, usability, installation, calibration and measurement accuracy.

Checklist

Pos.	Check Item	Yes	No	Comments
1	Sensor usability Sensor usability is same as the original one?	✓		
2	Sensor installation Sensor installation is same as the original one?	✓		
3	Sensor calibration Sensor calibration was successful in the ventilator?	✓		
4	Oxygen Measurement Sensor works without issues in ventilator?	✓		

Final Conclusion:

The oxygen sensor type OOM108 tested in MAQUET ventilator is compatible to the currently used oxygen sensor made by MAQUET.

☒ Yes ☐ No

Date:

Signature:

Dr. Khanh Thanh
Dr. Khanh Thanh (August, 02, 2023)

	Revision:	Created - Erstellt:	Page - Seite:
C:\Users\E339512\AppData\Local\Microsoft\Windows\NetCache\Content.Outlook\YEPULFMN\OOM108_compatibility_check-0.docx		MS	1 / 1