

Invoice Correction 5263485259

ORIGINAL

Invoice Correction Date: 04.05.2023

Honeywell Healthcare Solutions GmbH
Alter Holzhafen 18
23966 Wismar
Germany

Purchase Order No: RTS3100912

Billed to: 1797448
Vandagraph Sensor Technologies Ltd.
15 Station Road
Cross Hills, West Yorkshire
KEIGHLEY
BD20 7DT
United Kingdom

Sold to: 1797448
Vandagraph Sensor Technologies Ltd.
15 Station Road
Cross Hills, West Yorkshire
KEIGHLEY
BD20 7DT
United Kingdom

Order Information:

Credit Memo Request: 64068317
Sales Office: 2170 DE SPS Wismar
Incoterms: FCA Origin
Payment Terms: Net 30 Days After Invoice Date

Ship to: 1797448
Vandagraph Sensor Technologies Ltd.
15 Station Road
Cross Hills, West Yorkshire
KEIGHLEY
BD20 7DT
United Kingdom

EM NO.	Material No.	QTY	UOM	UNIT PRICE	EXT. PRICE	VAT %
00010	E1002856	4	EA	18,72	-74,88	0%
	Oxygen Sensor OOD103 VAN 8010060					
	Customer Part Nbr: 8010060					

Credit note for SN:
102701
103256
103813
105062

Totals

Total Before VAT -74,88 EUR
VAT-Total 0,00 EUR
Total Amount -74,88 EUR

Tax Information

Tax Exempt Reason: Tax free export - Goods
Fulfilment Date: 04.05.2023

Integrity and Compliance at Honeywell

Honeywell places the highest value on the integrity of its business operations. We expect our employees and business partners to comply with our Code of Business Conduct and all applicable laws and regulations. If you become aware of illegal or improper conduct by a Honeywell employee please contact your Honeywell relationship manager. You can also report misconduct directly to us either via e-mail to our Integrity and Compliance Office at access.integrity.helpline@honeywell.com or via our Access helpline accessible worldwide at 800-237-5982.

Invoice #:

Link Details: No Payment Due

Account Name:

Account No:

Bank Key:

Swift Code:

AN:

Payment method:

Registered No: Schwerin HRB 1838

VAT Reg No: DE137437213

Federal ID #:

Registered Address: Honeywell Healthcare Solutions GmbH, Alter Holzhafen 18, Wismar, 23966, Germany

Credit Analyst: Xenia Musat

Sales Rep: Stefan Dreyer

Customer Care Rep: Jessica Wagner +49 3841 360 206

Unless otherwise agreed in writing by the parties, all sales of products and/or services in this document shall be governed solely by the applicable Honeywell Terms and Conditions in effect at the time a Purchase Order is accepted by Honeywell, a copy of which can be found at <https://sps.honeywell.com/us/en/support/safety/legal-documents/sst>. Honeywell's acknowledgment of receipt of a Purchase Order shall not constitute acceptance. By continuing with this transaction, you (Buyer) agree to those Honeywell Terms and Conditions, which are subject to change at Honeywell's discretion without notice. Honeywell expressly limits its acceptance, fulfillment and performance of this Purchase Order to the terms in the Honeywell Terms and Conditions and expressly rejects any different, conflicting or additional terms in the Purchase Order or any other Buyer documents. This sale shall be governed by the Terms and Conditions of the Honeywell legal entity listed on this document.

PRÜFBERICHT – TEST REPORT

Sauerstoffsensoren/Oxygen Sensors OOA, OOI, OOD

Kunde/ Customer	Vandagaph	Retour/ RMA- No.	Kunde/ Customer	PT 53100912	Envitec	300262953
Artikel/ Article	000103	Anzahl/ Quantity	4		Eing. Datum/ Inc. Date	24.04.23
Bearbeitet/ processed	U. B. M.	Prüfdatum/ Insp. Date	27.04.23		Geprüft/ Inspected	K. B. M.

Produkt/ Identification				Messwert/ Test Result				Bewertung/ Evaluation		
Nr.	Seriennummer/ Serial no.	Lieferlos/ Manufact. code	Luftwert/ Air signal @ 20,9% O ₂ [mV]	Nullsignal/ Offset @ 0% O ₂ [µV]	Linearität/ Linearity [%]	Bemerkung/ Remark	außerhalb Garantie/ out of warranty	kein Fehler festgestellt/ no failure found	Fehlerkode/ Failure- code	Austausch / Replacement
1	102701	05122	3.9	682	i.0.				03	
2	103256	07122	11.1	35	i.0.			X		
3	103813	07122	0						04	
4	105062	09122	11.4	37	i.0.	instab. Elektrolyt			02	
5										
6										
7										
8										
9										
10										
11										
12										
13										
14										
15										
16										
17										
18										
19										
20										

Fehler O₂-Sensor: Signal instabil / Signal not stable

Fehler O2-S001: Luftwert außerhalb Toleranz/ Air signal out of tolerance
 Fehler O2-S003: Offset außerhalb Toleranz/ Offset out of tolerance
 Fehler O2-S005: Gehäuse defekt/ Housing defective
 Fehler O2-S007: Sonstiges (unter Angabe in Bemerkung) / Others (see remark)
 Fehler O2-S002: Signal instabil / Signal not stable
 Fehler O2-S004: Signal 0mV / Signal 0mV
 Fehler O2-S006: Elektrolyt Austritt / Electrolyte leakage