

DVM Calibration record

DVM Calibration Index Number	CE110
DVM Bar code	
Equipment Used	Time 1030 MicroCal
Serial number Test Equipment	6839k6
Calibration Certificate Number	1230925
Test Equipment Calibration Index Number	CE076 (Barcode 102060)
Micro Cal @	DVM Reading
0mV	0.00
5mV	5.0
10mV	10.0
15mV	15.0
20mV	20.0
25mV	25.0
30mV	30.0
35mV	35.0
40mv	40.0
Pass / Fail	Pass
Tested By	Derek Lamb
Tested On	22 / 08 / 2011
Signed	