

Microcal Calibration record

DVM Calibration Index Number	CE059
DVM Bar code	
Equipment Used	Time 1030 MicroCal
Serial number Test Equipment	CE111
Calibration Certificate Number	In House
Test Equipment Calibration Index Number	CE076 (Barcode 102060)
Micro Cal @	DVM Reading
0mV	0.00
5mV	4.9
10mV	9.9
15mV	14.9
20mV	19.9
25mV	25.0
30mV	30.0
35mV	34.9
40mv	39.9
Pass / Fail	Pass
Tested By	Derek Lamb
Tested On	22 / 08 / 2011
Signed	