

Manometer Calibration record

Calibration Index Number	CE149 ✓
Bar code	
Equipment Used	Manometer
Serial number Test Equipment	CE091
Calibration Certificate Number	1231021
Test Equipment Calibration Index Number	CE091 (Barcode 102075)
Manometer CE091 @	CE149
0mbar	0.00
5mbar 5.5	5.5
10mbar	10.0
15mbar	15.0
20mbar	20.0
25mbar	25.0
30mbar	30.1
35mbar	35.1
40mbar	40.1
Pass / Fail	P
Tested By	Derek Lamb
Tested On	22 / 08 / 2011
Signed	

