

Manometer Calibration record

Calibration Index Number	CE149 <i>CE078</i>
Bar code	
Equipment Used	Manometer
Serial number Test Equipment	CE091
Calibration Certificate Number	1231021
Test Equipment Calibration Index Number	CE091 (Barcode 102075)
Manometer CE091 @	CE149
0mbar	0.00
5mbar <i>5.5</i>	<i>5.5</i>
10mbar	<i>10.0</i>
15mbar	<i>15.0</i>
20mbar	<i>20.0</i>
25mbar	<i>25.0</i>
30mbar	<i>30.0</i>
35mbar	<i>35.0</i>
40mbar	<i>40.0</i>
Pass / Fail	<i>P</i>
Tested By	Derek Lamb
Tested On	22 / 08 / 2011
Signed	<i>Derek Lamb</i>