

AM Full, SN Full



SINGLE ITEM 8010006 - VST - Returns Statistics 2024-01-22

Get Data

Export CSV

Start From 22/01/2022

End at 22/01/2024

Honeywell Healthcare Solutions GmbH

 CID5922 Opera VST0004 Type:Company
Supplier

Honeywell Healthcare Solutions GmbH Alter
Holzhafen 18, Wismar, , D-23966, Germany
Tel **0049 3841 360204** Distributor Euro (0004)

 Log Call

Add Issue

Credit Credit Limit 0.00EUR

VAT DE137437213

 Account Home

 Main Details

Account Memos

 Sales History

 Purchase Orders

Stock
References

🔄 Cust./Supp. Returns

Service Visits

 Distributor Agreements

Documents
Files

 **Linked Issues**

Regulatory Agreements

SINGLE ITEM 8010006 - VST

Unique Faults QA , Demo & Customer 22 Jan 2022 - 23 Jan 2024

	Cleaned Fault	8010006
Super Clean	Defective Sensor Housing	1
Super Clean	High Output	1
Super Clean	Low Output	8
Super Clean	No Output	14
Super Clean	Unstable Output Signal	4
Super Clean	Zero Offset Signal Out Of Tolerance	1

Months to Failure From Invoice Customers

22 Jan 2022 - 23 Jan 2024

From Invoice to Customer to Booked into Returns system (Customers)

22/01/2024, 12:31

Stock Group Returns

Clean Code	1	2	3	4	6	7	8	9	10	11	12	15
	Months	Months	Months	Months	Months	Months	Months	Months	Months	Months	Months	Months
Unstable Output Signal	1	1		1					1			
High Output								1				
Zero Offset Signal Out Of Tolerance					1							
No Output	1		1		1	1	3	2	2	1	1	1
Low Output	1				3	1				2		1
Electrolyte Leakage		1										
Defective Sensor Housing			1									

Stock References Months to Failure From Invoice Customers 22 Jan 2022 - 23 Jan 2024

From Invoice to Customer to Booked into Returns system (Customers)

Reference	1	2	3	4	6	7	8	9	10	11	12	15
	Months	Months	Months	Months	Months	Months	Months	Months	Months	Months	Months	Months
8010006	3	2	2	1	5	2	3	3	3	3	1	2

Break Down Parts To Fault Code

Stock Ref	Supplier Ref	Description	Fault Code	Cleaned Fault Code	True Fault?	Qty
8010006	8010006	E1002733, OOD103 Oxygen Sensor	Defective Sensor Housing	Defective Sensor Housing	Fault	1
8010006	8010006	E1002733, OOD103 Oxygen Sensor	High Output	High Output	Fault	1
8010006	8010006	E1002733, OOD103 Oxygen Sensor	Low Output	Low Output	Fault	8
8010006	8010006	E1002733, OOD103 Oxygen Sensor	No Output	No Output	Fault	14
8010006	8010006	E1002733, OOD103 Oxygen Sensor	Out Of Warranty	Out Of Warranty	No Fault	2
8010006	8010006	E1002733, OOD103 Oxygen Sensor	Unstable Output Signal	Unstable Output Signal	Fault	4
8010006	8010006	E1002733, OOD103 Oxygen Sensor	Zero Offset Signal Out Of Tolerance	Zero Offset Signal Out Of Tolerance	Fault	1