

AM Full, SN Full



SINGLE ITEM 8010008 - VST - Returns Statistics 2024-01-22

Get Data    Export CSV

|            |            |
|------------|------------|
| Start From | 22/01/2022 |
|------------|------------|

End at 22/01/2024

# Honeywell Healthcare Solutions GmbH

 CID5922 Opera VST0004 Type:Company  
Supplier

**Honeywell Healthcare Solutions GmbH** Alter  
Holzhafen 18, Wismar, , D-23966, Germany  
Tel **0049 3841 360204** Distributor Euro (0004)

[Log Call](#)
[Add Issue](#)

Log Call Add Issue

Add Issue

Credit Credit Limit 0.00EUR

VAT DE137437213

Account  
Home

 Main Details

## Account Memos

 Sales History

## Purchase Orders

### Stock References

 Cust./Supp. Returns



## Service Visits

## Distributor Agreements

Documents  
Files

 **Linked Issues**

Regulatory Agreements

# SINGLE ITEM 8010008 - VST

# Unique Faults QA , Demo & Customer 22 Jan 2022 - 23 Jan 2024

## Cleaned Fault 8010008

Super Clean Low Output 2

|             |           |   |
|-------------|-----------|---|
| Super Clean | No Output | 3 |
|-------------|-----------|---|

## Months to Failure From Invoice Customers

22 Jan 2022 - 23 Jan 2024

### From Invoice to Customer to Booked into Returns system (Customers)

Clean Code 2 Months 4 Months 6 Months 12 Months

|           |   |   |
|-----------|---|---|
| No Output | 1 | 1 |
|-----------|---|---|

Stock References Months to Failure From

Invoice Customers

22 Jan 2022 - 23 Jan 2024

From Invoice to Customer to Booked into Returns system (Customers)

Reference 2 Months 4 Months 6 Months 12 Months

8010008 1 1 1 1

Break Down Parts To Fault Code

| Stock Ref | Supplier Ref | Description                    | Fault Code | Cleaned Fault Code | True Fault? | Qty |
|-----------|--------------|--------------------------------|------------|--------------------|-------------|-----|
| 8010008   | 8010008      | E1002734, OOD103 Oxygen Sensor | Low Output | Low Output         | Fault       | 2   |
| 8010008   | 8010008      | E1002734, OOD103 Oxygen Sensor | No Output  | No Output          | Fault       | 3   |