

Stock References Months to Failure From

Invoice Customers

22 Jan 2022 - 23 Jan 2024

From Invoice to Customer to Booked into Returns system (Customers)

Reference 6 Months 11 Months

8010007 1 2

Break Down Parts To Fault Code

Stock Ref	Supplier Ref	Description	Fault Code	Cleaned Fault Code	True Fault?	Qty
8010007	8010007	E1001581, OOD103-1V Oxygen Sensor	Low Output	Low Output	Fault	1
8010007	8010007	E1001581, OOD103-1V Oxygen Sensor	No Fault Found	No Fault Found	No Fault	9
8010007	8010007	E1001581, OOD103-1V Oxygen Sensor	Unstable Output Signal	Unstable Output Signal	Fault	2