

Invoice Correction 5266308266

1 message

Schliemann, Andre <Andre.Schliemann@honeywell.com>
To: "cathy.green@viamed.co.uk" <cathy.green@viamed.co.uk>

20 March 2024 at 09:02

Dear Cathy,

attached the credit note, the OxiQuant S is on the way back to you, it needs only a reset.

For the future to safe cost, please see , how you can reset the OxiQuantS.

Reset	
Switch on the device and then press and hold both arrow buttons until the sensor voltage is displayed. Sensor symbol in the display above the number behind the dot.	
Click through the menu to rEs by pressing the On/Off button. Confirm with the down arrow button. Reset performed, use the arrow buttons to set the device to the value of 20.9 % at ambient air.	

Kind regards / Mit freundlichen Gruessen

i. A. André Schliemann, Dipl.-Ing.(FH)

Team Leader Service & Repair
Honeywell | Industrial Automation

Office: +49 (0) 3841 360 423
Fax: +49 (0) 3841 360 222

andre.schliemann@honeywell.com

-

www.honeywell.com

www.envitec.com

Honeywell Healthcare Solutions GmbH

Alter Holzhafen 18

23966 Wismar, Germany

Trade Register: Schwerin, HRB 1838

Managing Director: Charbel Rizkallah

2 attachments



Invoice Correction 5266308266.pdf

25K



300272874.pdf

419K