

Temperature Probes Revision Questions (Edition 1)

Please review the Introduction (DOCID 109087) and Technical (DOCID 111263) training notes in the document index and answer the following questions.

Email *

viamed.robert.connor@gmail.com

What is the purpose of monitoring patient temperature during surgery? *

1 point

- ☒ To detect potential health issues like hypothermia or fever
- ☐ To track blood pressure fluctuations
- ☐ To monitor oxygen saturation
- ☐ To check heart rate during anaesthesia

Which of the following is considered the most accurate site for measuring core body temperature? * 1 point

- ☐ Skin
- ☐ Nasopharynx
- ☐ Oesophagus
- ☒ Rectum

What are the two main types of Bluepoint 400-AC temperature probes? *

1 point

- ☒ Oesophageal/Rectal and Skin/Surface
- ☐ Nasopharyngeal and Surface
- ☐ Disposable and Reusable
- ☐ Handheld and Digital

Which of the following patient sizes are available for the Bluepoint 400-AC temperature probes?

* 1 point

- ☒ Adult, Paediatric, and Neonatal
- ☐ Adult, Infant, and Senior
- ☐ Infant, Child, and Adult
- ☐ Newborn, Child, and Adolescent

What material are Bluepoint 400-AC temperature probes made from to prevent latex allergies? *

1 point

- ☐ Thermoplastic rubber
- ☒ Santoprene
- ☐ Silicone
- ☐ Polycarbonate

How should 400-AC temperature probes be cleaned between uses? *

1 point

- ☐ Hand wash only
- ☒ Steam sterilization using an autoclave
- ☐ Rinsed with alcohol
- ☐ Soaked in bleach

Why is monitoring core temperature critical during anaesthesia? *

1 point

- ☐ Anaesthetised patients can experience rapid temperature fluctuations.
- ☒ Anaesthesia makes it hard for patients to maintain core temperature.
- ☐ Core temperature directly affects heart rate.
- ☐ Both A and B

Which leading medical device manufacturers are Bluepoint 400-AC probes compatible with?

* 1 point

- ☒ YSI, Siemens, GE Healthcare, Dräger
- ☐ YSI, IBM, Samsung, Siemens
- ☐ Dräger, Siemens, Microsoft, Apple
- ☐ GE Healthcare, Apple, Philips, Hewlett Packard

What is the typical warranty period for Bluepoint 400-AC temperature probes? *

1 point

- ☐ 3 months
- ☐ 6 months
- ☒ 12 months
- ☐ Lifetime

What feature of the 400-AC probes minimizes the risk of cross-infection between patients?

* 1 point

- ☐ They are disposable
- ☒ They are steam autoclavable
- ☐ They are coated with antimicrobial material
- ☐ They can be cleaned with alcohol wipes

This content is neither created nor endorsed by Google.

Google Forms