

# INFANT RESUSCITATION CABINET

## OPERATOR'S MANUAL



**CE0086**

## INDEX

<u>Section</u>	<u>Page</u>
Definitions and Symbols	3
Definitions	
Symbols	
Specifications and Technical	4
Introduction	5
Overview	
Wall Mounted Radiant Warmer	6
Precautions	
Front ( <i>diagram</i> )	7
Rear ( <i>diagram</i> )	
Adjusting the Heat Output	8
Using the Halogen Lamp	
LED Heating Indicator	
Alarm and Safety-Power Function	9
Resuscitation Cabinet	10
Low Suction Controller	11
Vacuum Source Connector	
Patient Inlet Connection	
Vacuum Gauge	
Safety Valve	
Setting Suction Levels: Therapy Equipment	12
Setting Suction Levels: Oxylitre	
Suction Receiving Equipment	13
Instructions for Use; Vacsax System	
Instructions for Use; Therapy Equipment System	14
Apgar Timer	15
Description	
Function of the Keys	
Stand-by Mode	
Counting Mode	16
Freeze Mode	
Reset Function	
Specifications	
Battery Replacement	17
Routine Maintenance	
Optional Accessories	
Cleaning and Decontamination	

Tom Thumb Infant Resuscitator	18
Pre-use Set-up TT490-15	
Pre-use Set-up TT480	19
Low-Flow Air/Oxygen Blender	20
Equipment Layout and Interconnections	21
Pre-use Checks	22
Calibrating the Oxygen Analyser	
Consumables and Parts List	24
Parts	
Consumables	25
Operator and Service Manuals	
Servicing	26
Routinely replaced parts	
Service Sheet	27
Apgar Timer Service Sheet	29
Cleaning	30
Clean as per MAC Manual “Part 2; Cleaning (manual) – non-immersion”	
Cleaning Guidelines	
Equipment Required	31
Procedure	
Monitoring and Control	32
Safety Precautions	
Items that can be Steam Sterilized at 121° C as per HTM 2010, Part 3	
Items Intended for Single Patient Use Only	
Warranty	33
Company Details	33

## Definitions and Symbols

### Definitions

**Note:** The remark “Note” is used in the text to indicate procedures or conditions, which might otherwise be overlooked or incorrectly understood. A note may also be used to clarify apparently contradictory or confusing situations.

**Caution:** The remark “Caution” is used to draw attention to a procedure, which must be followed exactly in order to avoid potentially damaging the equipment.

**Warning:** The remark “Warning” is used in the text to draw attention to dangerous situations in connection with the operation, cleaning or maintenance of the equipment if there is a possibility of injury or danger of death to the operator or patient.

### Symbols



Attention: consult accompanying documents



Type B



Attention: hot surface



AC Power



Equipotential plug



Power ON

Power OFF

CE0123

Indicates this device in compliance with EU-Directive 93/42/EEC/Annex II.3. 0123 is the Notified Body Number. All radiators with this signification are with regard to the production supervised from TÜV Product Service.



Date of Manufacture

## SPECIFICATIONS AND TECHNICAL DATA

### Radiant Warmer

Power requirement: 220-240 V AC, 50/60 Hz, 630 W  
Protection class: I  
Degree of protection: B, IP20  
Test provision: IEC 601-2 TUV / CE

### Low Flow Air/Oxygen Blender (model IHC2003)

**Device-Specific Standards:** Complies with ISO 11195 : 1995.

**Oxygen % Range:** 21 to 100%

**Oxygen % Accuracy:**  $\pm 3\%$  of full scale

**Supply Pressure:** Both supplies within range of 30-75 psi (207-517 kPa) and air & oxygen must be within 10 psi (69 kPa) of each other. Do not use on a patient or with a ventilator outside of this range.

**Maximum Flow:**  $\geq 30$  LPM (Low Flow blender) @ 60% setting & 50 psi (345 kPa) inlet pressures.

**Pressure Drop:**  $< 6$  psi (42 kPa) at 50 psi (345 kPa) inlet pressure and 10 LPM flow.

**Alarm/Bypass Reset:** when inlet gas pressure differential is  $\geq 6$  psi (42 kPa).

**Alarm Intensity:** 80 dB at 30 cm (1 ft)

**Input fittings:** Oxygen NIST, Air NIST

**Output Fitting(s):** Male DISS, oxygen type.

**Dimensions:** Height: 8.9 cm (3 1/2")  
Width: 5.7 cm (2 1/4")  
Depth: 7.3 cm (2 7/8")

**Weight:** 1.25 Kg (2 3/4 lbs)

**No electronics incorporated**

**Reverse Gas Flow:** From either gas inlet to the other is zero (complies with clause 6 of ISO 11195).

## INTRODUCTION

This document contains the instructions for the operation, cleaning and basic maintenance of the wall mounted Viamed Infant Resuscitation Cabinet.

Viamed is not liable for the proper functioning of any part of this product if it is not operated according to the instructions, if the maintenance recommendations in this manual are not followed, or if repairs are carried out using non-approved components.

Only qualified personnel should carry out servicing of this equipment. Maintenance documents can be obtained through your local distributor or direct from Viamed.

Servicing personnel must be aware of the potential clinical implications of incorrectly serviced equipment.

The personnel who work with this equipment should read this manual carefully and should fully understand all instructions contained therein. The manual should be kept so that it can be easily inspected; it is advisable to store it in an easily accessible place.

If any function or part of this manual is not clear, please contact Viamed or your distributor in order to obtain further information or clarification.

### Overview

The Viamed Infant Resuscitation Cabinet contains a carefully selected group of products; each main item can be purchased as a single item for use individually.

Viamed have combined these devices into one product in an arrangement ideally suited to Maternity and Neonatal Intensive Care environments.

Devices forming component parts of the Resuscitation Cabinet:

- Cabinet with drop down door, integral bed and mattress
- Wall-mounted Ceratherm 600-2 radiant warmer
- Tom Thumb infant resuscitation unit
- Low suction controller
- Suction receiving jar
- Digital Apgar timer
- Storage bins

The Tom Thumb infant resuscitation unit delivers medical oxygen; if blended gas mixes are required an optional blender and oxygen analyser are available.

All the components carry the correct and relevant CE marking.

## WALL MOUNTED RADIANT WARMER



Radiant Warmer with Double-Hinged Arm

The radiant warmer is mounted on a double-hinged arm, allowing the unit to be folded flat to the wall when not in use.

### **Precautions**

The radiant warmer must not be used in the immediate vicinity of explosive anaesthetic gases or mixtures.

The distance between the surface on which the patient lies, and the lower edge of the radiant warmer, must not be less than 80cm. If this instruction is not followed, prolonged exposure to the heat radiation may cause burns.

**“Danger of burns”:** Contact with the protective guards, element and reflector should be avoided.

**“Danger of fire”:** The grille on the upper side of the radiant warmer must remain unobstructed at all times.

Do not place any items on the grille.

Patients must not be left unattended whilst the radiant warmer is in use.

Do not operate the device if the casing has been removed or is damaged.

Servicing and maintenance should be restricted to qualified technical personnel.

Use of the radiant warmer should be restricted to trained medical personnel.

Use of a radiant warmer can increase fluid loss from a patient.

Handle the warmer with care; the ceramic heating element is fragile.

Do not touch the heating element at any time as this can adversely affect the performance of the element.

The heating element gets very hot when in use: take care if the warmer has recently been used or tested, and allow it to cool before storage. The side handles remain cool to touch.

Each Ceratherm 600-2 Infant Radiant Warmer is safety tested to Class 1, Type B and a test report is supplied. It is recommended that the equipment be retested for safety and function upon receipt, and at least annually thereafter.




### **WARNING!**

The radiant warmer gets hot whilst in use.


Do not obstruct the upper ventilation grille.

Use only the side handles, which remain cool, to reposition the unit.

### **Front**



### **Rear**



## **Adjusting the Heat Output**

Switch radiant warmer On/Off switch to ON. This switch is the master power switch for the device; no electrical functions are active with this switch in the OFF position.

Using the forward and reverse buttons (< and >) set the desired heat output level. The current heat output setting is shown by an illuminated yellow indicator ①, ②, ③ or ④.

The heat output can be adjusted using the four heat output settings (① - ④). The factory-set levels correspond to the following values, the values indicated correspond to a percentage of the maximum power output of the heating element:

Level 1	= 25%	①	②	③	④
Level 2	= 50%	①	②	③	④
Level 3	= 75%	①	②	③	④
Level 4	= 99%	①	②	③	④

The various power levels can be used as follows:

Level ①: To keep the support surface warm for continuous operation.

Level ②: For normal operation on changing and examination areas.

Level ③: Additional warmth for resuscitation, the labour room or for the operating theatre.

Level ④: For increased heat, in the operating theatre, during anaesthesia, or for adults.

## **Using the Halogen Lamp**

The non-dazzling halogen lamp is used for illuminating the patient area.

To turn on the lamp, set the light On/Off switch to ON.

**Note:** the lamp can only be used when the radiant warmer On/Off switch is in the ON position.

If it is required, the halogen lamp can be used without the heating element, to do this adjust the heat output of the radiant warmer to zero. The heat output is set to zero when the reverse button (<) is repeatedly pressed until all yellow indicators (① - ④) are extinguished.

## **LED Heating Indicator**

The Heating Indicator illuminates when the element is being heated, and is off when it is not. This 'duty-cycle' occurs over a period of a second, the percentage of time that the indicator is illuminated during that second represents the power level in terms of a percentage of the maximum output.

### **Alarm and Safety-Power Function**

The radiant warmer incorporates a safety feature, which reduces the output power if the device is left unattended. An audible and visual alarm is triggered a period of 15 minutes after the last key-press; this is to inform the operator that the device is still active.

The safety alarm activates an audible alarm signal, which consists of a series of beeps over a 5 second period, and the red Alarm Indicator on the front panel begins to flash.

The alarm is cancelled by pressing the flashing Alarm Indicator, if this is done during the first 8 seconds after the onset of the alarm condition, the radiant warmer will continue to function at the current power setting, the Alarm Indicator stops flashing and the 15-minute timer begins again.

If, after 8 seconds from the onset of the alarm condition the alarm has not been cancelled, the heat output of the radiant warmer reduces to a pre-set value, which is 20% of the maximum power output of the device.

**Note:** when the radiant warmer enters the reduced safety-power setting, the yellow power level indicator remains illuminated at the previous power level; this is to inform the user of the power level that the device will return to when the alarm is cancelled.

When the alarm is cancelled, the radiant warmer reverts back to the previous power level, the Alarm Indicator stops flashing and the 15-minute timer begins again.



#### **WARNING!**

The radiant warmer must not be left unattended whilst in use.

## RESUSCITATION CABINET



Infant Resuscitation Cabinet in Standard Configuration

The cabinet has been designed to conceal the internal equipment when closed, and provide easy access when open.

The drop down platform has an integral bed, on which the mattress is placed. Additional blanket material may be placed on the mattress, but care should be taken to ensure that this does not become trapped in the space between the platform and the cabinet when raising the platform.

Excessive weight should not be applied to the drop down platform. The maximum loading, including the pressure applied by the operator(s), should be no greater than 25 Kg.

Whilst in use, the patient should be under close observation at all times.

When closing the cabinet, ensure that there are no obstructions; the cabinet should close easily and without significant force. Once the door is in the upright position, turn the catches to secure it in the closed position.



### CAUTION

The door should not present resistance or need to be held closed in order to secure the catches.

Securing the catches whilst forcing the door closed may cause damage to the catches: damage of this type is not covered by the warranty.

## LOW SUCTION CONTROLLER

The Viamed Infant Resuscitation Cabinet is supplied with one of two low suction controllers, designed specifically for medical use: the Oxylitre S714 or the Therapy Equipment 4725-3M.



Therapy Equipment Low Suction Controller

### Vacuum Source Connector

The suction controller is supplied with a 3m vacuum hose assembly, which terminates in a British Standard (BS5682) probe. The device is intended for connection to a wall vacuum outlet.

### Patient Inlet Connection

The inlet connection forms an integral part of a detachable filter cartridge or Disposable Hydrophobic Filtration Unit. This filtration unit has been designed to prevent the ingress of fluids into the controller and the pipeline system.

Once the filter has been used for patient therapy, or if wetted for any reason, it must be replaced. If the filter membrane is wet, the chemical coating on it will prevent fluid passing through it and may restrict the flow of suction.

### Vacuum Gauge

The controller is fitted with an easy to read, dual scale colour coded gauge.

Model	Scale readings
Oxylitre S714	0 to 200 mmHg (0 to 25kpa)
Therapy Equipment 4725-3M	0 to 300 mmHg (0 to 40kpa)

### Safety Valve

The suction controller is fitted with an internal safety valve system. This will protect the suction controller from being damaged if inadvertently connected to a positive pressure outlet.

### **Setting Suction Levels: Therapy Equipment**

To increase the level of suction, rotate the control knob on the top of the suction controller in a clockwise direction.

To decrease the level of suction, rotate the control knob on the top of the suction controller in an anti-clockwise direction.

The Therapy Equipment 4725-3M has a separate ON/OFF control in the form of a 'flag' lever on the right hand side of the device.

To activate suction, move the lever to the vertically downward position, the word 'ON' is displayed on the flag. To turn the suction off, move the lever to the vertically upright position, the word 'OFF' is displayed on the flag.



#### **WARNING**

The suction level on the Therapy Equipment low suction controller will remain set at the level last used.

The suction level must be set before each use to a level determined by guidelines within the customer's own organisation.

### **Setting Suction Levels: Oxylitre**

To increase the level of suction, rotate the control knob on the top of the suction controller in an anti-clockwise direction.

To decrease the level of suction, rotate the control knob on the top of the suction controller in a clockwise direction.

The Oxylitre S714 suction controller does not have a separate ON/OFF control: to turn the device off, decrease the suction level to the minimum setting.



#### **WARNING**

The suction level must be set before each use to a level determined by guidelines within the customer's own organisation.

## SUCTION RECEIVING EQUIPMENT

The Viamed Infant Resuscitation Cabinet is supplied with the VacSax disposable suction receiving system as standard, or a reusable Therapy Equipment suction jar by request.



Vacsax Suction Receiving System

Viamed do not supply suction catheters, however, the system is designed for use with standard 10mm suction tubing compatible with a wide range of commonly used suction catheters.

Whichever receiving system is used, ensure that the suction tubing and catheter is replaced after each use.

### **Instructions for Use: Vacsax Advance System**

Ensure rigid canister is clean: wipe inner rim with damp cloth if necessary. If the canister is contaminated it may be autoclaved at 121° C or cleaned with water-based disinfectants; see instructions later in this document.

Use the correct size canister for the liner.

Remove liner from carton.

Place liner into rigid canister and push down firmly. Ensure patient port is facing forwards for ease of connection.

Push the taper connector on the vacuum tubing from suction controller into the vacuum port with a twisting motion.

Connect the patient tubing firmly to the patient port to ensure a good fit.

Turn the low suction controller on and set to maximum, then set to the desired vacuum level.

**Pre-use check:** Confirm that suction is present at patient tube by manually occluding the patient tube end before proceeding.

**After use:** Ensure that you wear gloves when handling used liners.

**Important:** Do not turn off vacuum source whilst disconnecting the patient tubing, if the system has been overfilled fluids may pass back down the patient tubing.

- Remove patient tubing and fit the stopper located on the rim into the patient port.
- Disconnect the suction tubing from the vacuum port of the liner and switch off the vacuum at the suction controller.
- To remove the liner from the canister, use the handles on the side of the liner to lift the used liner carefully out of the canister.
- Dispose of liner in accordance with hospital policy.
- Dispose of vacuum tubing set if it becomes contaminated during use.

**Note:** the tapered vacuum port connector is reusable and can be cleaned in accordance with the guidelines in the MAC Manual "Part 2; Cleaning (manual) - non-immersion"; see instructions later in this document.

#### **Instructions for Use; Therapy Equipment System (Optional)**

The Therapy Equipment suction jar is a reusable jar manufactured from polysulphone, it may be autoclaved at 121° C or cleaned with water-based disinfectants in accordance with the guidelines in the MAC Manual "Part 2; Cleaning (manual) - non-immersion"; see instructions later in this document. Additionally, it may be pre-washed using an automated disinfectant operating at 90° C.

Connect the suction controller to the receiving jar using 10 mm diameter suction tubing, ensure that the tubing is connected firmly onto the connection marked 'vacuum'.

Connect the patient tubing to the receiving jar using 10 mm diameter suction tubing, ensure that the tubing is connected firmly onto the connection marked 'patient'.

Always use tubing that is a good fit over the connectors of the suction jar.

Always clean the jar thoroughly after use.


The jar cradle requires only external cleaning, which must be done in accordance with the guidelines in the MAC Manual "Part 2; Cleaning (manual) - non-immersion"; see instructions later in this document.

## APGAR TIMER

### Description

The digital Apgar timer is a battery powered digital timer with preset audible timing indications at 1, 5 and 10-minute intervals.

The timer has a large, clear LCD display and is operated using the keys on the front of the instrument.



Viamed Digital Apgar Timer

### The Functions of the Keys

Key	Function
 <b>START</b>	<ul style="list-style-type: none"><li>Starts counting from 00:00.</li><li>Returns to <i>Counting Mode</i> whilst in <i>Freeze Mode</i>.</li></ul>
 <b>FREEZE/ UNFREEZE</b>	<ul style="list-style-type: none"><li>Enters and exits <i>Freeze Mode</i>. Display 'freezes' whilst the actual elapsed time continues counting, but is not displayed. Exiting <i>Freeze Mode</i> returns to <i>Counting Mode</i>, resuming from the ongoing total elapsed time.</li></ul>
 <b>RESET</b>	<ul style="list-style-type: none"><li>Resets counter, returning timer to <i>Standby Mode</i>.</li></ul>

### Standby Mode

The Apgar timer does not have an ON/OFF switch. When not counting, the Apgar timer enters *Standby Mode*.

In *Standby Mode* the display is blank with the exception of a flashing colon (:) to indicate that the timer is functioning correctly.

If the display is completely blank with no flashing colon, the timer may require battery replacement. Replace batteries on an annual basis.

## **Counting Mode**

To enter *Counting Mode*, press the START button.

When the Apgar timer begins counting, the display counts upwards from 00:00 in 1-second intervals up to a maximum of 1 hour (59:59 on display), it will then enter *Standby Mode* automatically.

Whilst counting, the timer will indicate the Apgar scoring intervals with audible beeps:

1 minute:	1 beep
5 minutes:	2 beeps
10 minutes:	3 beeps

## **Freeze Mode**

It is possible to freeze the display whilst the actual time continues to count in the background; this is to allow for recording the time from the start of the counting cycle of a particular event.

To enter *Freeze Mode*, press the FREEZE button.

Whilst the display is frozen, the colon (:) continues to flash.

*Freeze Mode* can be exited and *Counting Mode* resumed by pressing the START or FREEZE button.

Upon unfreezing, the Apgar timer will continue counting, resuming from the ongoing total elapsed time.

Should an Apgar scoring interval be reached whilst the Apgar timer is in *Freeze Mode*, the timer will beep as it would in *Counting Mode* to indicate that interval, but will remain in *Freeze Mode* until exited by the user.

## **Reset Function**

The *Reset Function* allows the timer to be reset and placed back into *Standby Mode*.

Press the RESET button to reset the timer.

## **Specifications**

Dimensions (excluding bracket)	142 mm (W) x 130 mm (H) X 40 mm (D)
Weight	551 g
Display	4 digit LCD. Digit height: 25mm
Power requirements	4 x AA/LR6/MN1500 1.5 V Alkaline Batteries
Controls	Start, Freeze/Unfreeze, Reset
Audio indications	1, 5, 10 minutes

## **Battery Replacement**

For optimal performance, batteries should be replaced on an annual basis.

If the Apgar timer is to be removed from service or stored, the batteries must be removed.

The LCD flashing between the messages “Lo” and “Batt” indicates a low battery condition, when this message appears, the batteries should be replaced at the earliest opportunity.

To replace the batteries, remove the mounting bracket from the rear of the timer using a 3/16<sup>th</sup> inch Allen Key, then using a narrow blade screwdriver release the battery cover. Remove the existing batteries and insert new ones (4 x AA/LR6/MN1500 1.5V Alkaline), observing the correct polarity.

Depleted batteries should be disposed of as per local regulations.

## **Routine Maintenance**

It is recommended that the Apgar timer be checked at least annually to ensure correct performance.

Perform service checks as detailed in accompanying Service Sheet. There are no user serviceable parts inside: if unit does not meet the specification please return it to Viamed for repair.

## **Optional Accessories**

<b>Part Number</b>	<b>Description</b>
0320200	Medirail mounting bracket
0121199	Swivel mount
0121200	Universal mounting clamp

## **Cleaning and Decontamination**

For routine cleaning, the Apgar timer can be cleaned with isopropyl alcohol, or with a mild detergent solution in accordance with the detergent manufacturer's recommendations.

For decontamination, clean in accordance with the guidelines in the MAC Manual “Part 2; Cleaning (manual) - non-immersion”.

## TOM THUMB INFANT RESUSCITATOR



Instructions for Use are attached to the Tom Thumb within the resuscitation cabinet. Please refer to these before using the Tom Thumb.

The Tom Thumb Infant Resuscitator is supplied in one of two variants: the TT490-15, which has a 3m oxygen hose terminated with a BS probe, or the TT480, which is designed to be connected to an existing flowmeter or air/oxygen blender via low-pressure tubing.

The Tom Thumb allows the delivery of oxygen at controlled flow and pressure. A flowmeter is used to control the flow rate of oxygen, and the adjustable pressure relief valve is used to set the maximum pressure delivered.

The Tom Thumb incorporates a secondary precision blow-off valve for added safety.



### **WARNING**

Do not operate the Tom Thumb TT490-15 at flows greater than 15 Litres/min.

Do not operate the Tom Thumb TT480 at flows greater than 15 Litres/min.

### **Pre-use Set-up TT490-15**

The Tom Thumb is supplied with a laminated set of instructions that should be kept near to the device for reference. Refer to these instructions for an illustration of the device and the numbers detailed in the instructions below.

Adjust the flowmeter ① to minimum (fully clockwise) and the adjustable pressure valve control ② to minimum (fully counter clockwise).

Ensure that the flowmeter oxygen hose ③ is connected to the external oxygen supply.

Check that the pressure gauge ④ reads zero. If not, the Tom Thumb requires servicing.

Connect the NeoPEEP patient circuit ⑤ to the Tom Thumb outlet using the connector ⑥ provided. Do not apply to the patient at this stage.

Set the flowmeter to the required flow rate, up to a maximum of 15 L/min.

Occlude the patient opening of the 'T-Piece' ⑦ and the PEEP (Positive End Expiratory Pressure) valve outlet ⑧ to create an airtight seal.

Turn the adjustable pressure valve control ② until the required PIP (Peak Inflation Pressure) is set, as shown by the pressure gauge ④.

Uncover the PEEP valve outlet ⑧. Adjust the PEEP setting by adjusting the PEEP control cap ⑨ on the NeoPEEP patient circuit until the reading on the pressure gauge ④ indicates that the correct PEEP has been achieved.

Connect the 'T-Piece' ⑦ to a suitable resuscitation mask ⑩ or to the patient's E.T. tube.

The Tom Thumb is now ready for use.

### **Pre-use Set-up TT480**

Ensure that the Tom Thumb is connected via low-pressure tubing to a flowmeter, which is in turn connected to a supply of oxygen or mixed gas. If using an air/oxygen blender, refer to the relevant section in this manual to perform the necessary pre-use checks on the blender.

Follow the above pre-use set-up procedures as described for TT490-15.

**Note:** when setting the maximum pressure, the needle on the gauge of the Tom Thumb should be stable and oxygen will be heard hissing from the adjustable valve. If the needle is not stable and no hissing can be heard, it is likely that the mask is not entirely occluded; check the seal by repositioning the mask until a stable reading is achieved.



#### **CAUTION**

For use by qualified and trained personnel only.

Use flow rates within the range of the flowmeter.

Adjust outlet pressure after altering the flow rate, if the flow rate is subsequently altered, always re-set the pressure.

Do not attempt to adjust the safety valve.



#### **WARNING**

Flow rates, Peak Inflation Pressure (PIP), Positive End-Expiratory Pressure (PEEP) and oxygen concentrations for resuscitation are clinical decisions: if in any doubt consult your key trainer or a relevant qualified person.

A set of the appropriate instructions is provided with each instrument, additional copies can be requested from Viamed.

Service manuals are available from Viamed.

## LOW FLOW AIR/OXYGEN BLENDER

The Viamed Infant Resuscitation Cabinet may be used with a low flow air/oxygen blender, providing a model TT480 Tom Thumb is installed. The TT480 has a tapered input to connect via low-pressure tubing to the output flowmeter of an air/oxygen blender.



### CAUTION

If using an air/oxygen blender, please consult the Operator's Manual for that device prior to use.

Viamed supply the Bio-Med Devices model 2003 low flow air/oxygen blender as an option, which can be mounted inside the resuscitation cabinet.

**Note:** for cabinets supplied prior to 2009 the medirail sections may need to be replaced with longer lengths: contact Viamed for further details.



Bio-Med Devices model 2003 Low Flow Air/Oxygen Blender

The Bio-Med Devices model 2003 air/oxygen blender is a precision proportioning device for mixing medical grade air and oxygen to any concentration from 21% to 100% oxygen at a flow-rate up to 30 L/min and delivering it to a variety of respiratory care devices.

The gas inlet fittings are located on the bottom of the blender and are designed such that the air and oxygen connections cannot be reversed.

The blender has an audible alarm built in to detect if either of the gas sources changes by more than 20 psi (138 kPa) from the other. This will warn the user that they are running out of one of the gas sources or that there is a severe pressure drop in one source.

If both gas sources drop or increase together such that a 20 psi (138 kPa) difference cannot be detected, then no alarm will sound.

If the blender is connected but not being used and a 20 psi (138 kPa) difference in gas sources develops, the blender will not alarm.


The blender alarm/bypass function will provide the full 30 L/min upon the loss of air or oxygen if the remaining gas is at 50 psi (345 kPa).



Infant Resuscitation Cabinet with a Low Flow Air/Oxygen Blender and Oxygen Analyser Installed

### Equipment Layout and Interconnections

The equipment should be connected as shown in the image below.



Equipment Layout and Interconnections

## **Pre-use Checks**

Before use, check that a high-pressure air hose is connected to the air fitting and a high-pressure oxygen hose is connected to the oxygen fitting on the bottom of the blender.

Check that the air and oxygen hoses are connected to a supply of air and oxygen via fixed wall outlets.



### **CAUTION**

It is recommended that the infant resuscitation cabinet system operate from a medical gas pipeline system.

If the cabinet has been configured to operate from gas cylinders, ensure that the air and oxygen hoses are connected to the cylinders and that the cylinders are delivering the correct pressure as determined by the engineer(s) who configured the system. If in doubt, consult an engineer.

Ensure that a 0 – 15 L/min flowmeter is connected to the primary outlet of the blender, which on the Bio-Med Devices model 2003 is the outlet on the left of the device when looking from the front.

Ensure that a flowmeter female tubing nipple is connected to the blender flowmeter, and that this is in turn connected using low-pressure tubing to the Tom Thumb infant resuscitator.

Ensure that a paediatric T-adapter is connected to the outlet of the Tom Thumb.

Ensure that a NeoPEEP patient circuit is connected to the outlet of the Tom Thumb using a 15mm I.D. to 15mm I.D. silicone adapter.

Before using the blender to deliver gas to a patient, set the control on the front of the blender to the desired oxygen concentration and use a calibrated oxygen analyser to check the accuracy of the patient gas.

## **Calibrating the Oxygen Analyser**

Viamed supply the Teledyne AX300 oxygen analyser as an option, which can be mounted inside the resuscitation cabinet.

**Note:** for cabinets supplied prior to 2009, the middle and lower medirail sections may need to be replaced with longer lengths in order to accommodate a blender and oxygen analyser: contact Viamed for further details.



### **CAUTION**

If using an oxygen analyser, please consult the Operator's Manual for that device prior to use.

The oxygen sensor is connected via a flow diverter into the paediatric T-adapter, the T-adapter is connected to the outlet of the Tom Thumb, and the patient circuit is connected to the T-adapter.



Oxygen Sensor Connected to T-Adapter via a Flow Diverter

Remove the oxygen sensor from the T-adapter by pulling it straight out of the port.

Unscrew the flow diverter to expose the sensor to air.

Calibrate the oxygen analyser as detailed in the operator's manual for that device.

Once the calibration has been performed and the analyser is displaying an air reading, reconnect the flow diverter and insert the sensor into the T-piece.

To check the accuracy of the patient gas delivered by the blender, set a flow by adjusting the flowmeter on the blender. Set the control on the front of the blender to the desired oxygen concentration and verify that the reading on the oxygen analyser corresponds to the concentration that has been set.



#### NOTE

The accuracy of the blender is  $\pm 3\%$  of full scale

When changing oxygen concentration, wait sixty seconds for the gas mixture to stabilize before checking it against the analyser.



#### WARNING

Determining the flow rates, Peak Inflation Pressure (PIP), Positive End-Expiratory Pressure (PEEP) and oxygen concentration for resuscitation are clinical decisions: follow your own organization's protocols and if in any doubt consult your Key Trainer or a relevant qualified person.

## CONSUMABLES AND PARTS LIST

**Note:** the Viamed resuscitation cabinet is available in different configurations, the parts list contains equipment from all variants; care must be taken when re-ordering replacement parts. If in doubt, please contact Viamed to determine the correct item.

For parts not listed, such as spare parts required to effect repairs, please contact Viamed.

Parts	
Part no.	Description
0310002	Complete resuscitation cabinet including all parts
0310025	Wall mounted radiant warmer
0330120	Element for radiant warmer – 600W ceramic
0330125	Halogen lamp for radiant warmer, 20W
0330105	Cabinet body with shelf and medirail
0332100	Resuscitation cabinet door catch (for cabinets manufactured after January '09)
0320020	Resuscitation bed, 35cm (14") wide
0320043	Mattress, 35cm (14") wide
0310035	Low suction controller (Oxylitre S714)
0330040	Suction hose assembly, 3m (Oxylitre)
0310037	Low suction controller (Therapy Equipment 4725-3M)
0330033	Suction hose assembly, 3m (Therapy Equipment)
3833-132	VacSax reusable suction canister
3833-007	VacSax tapered suction connectors (pack/10)
9910-340	VacSax disposable liners (box/25)
0320010	Universal rail clamp with 'V' for mounting Vacsax suction jar
0310100	Apgar Timer (note: mounting bracket sold separately)
0320200	Mounting bracket with rail clamp for mounting Apgar timer
0310020	Size 3 storage box
0310030	Tom Thumb resuscitation unit, type TT480 without flowmeter
0310034	Tom Thumb resuscitation unit, type TT490-15-3M with flowmeter
0330218	Oxygen hose assembly for Tom Thumb, 3m
0310200	Low-flow air/oxygen blender including flowmeter, hoses and mounting bracket
0310201	Low-flow air/oxygen blender
0320210	Flowmeter, 0 – 15 L/min
0320214	Rail mount bracket for air/oxygen blender
0320216	NIST hose with MKIV oxygen probe, 3m
0320217	NIST hose with MKIV air probe, 3m
0330213	O-ring, thick. Each Tom Thumb uses 1, replaced during service
0330214	O-ring, thin. Each Tom Thumb uses 6, replaced during service

Consumables	
Part no.	Description
0310104	Filter for Oxylitre low suction controller, single unit
0310105	Filters for Oxylitre low suction controller, box of 30
0310109	Filter for Therapy Equipment low suction controller, single unit
0310110	Filters for Therapy Equipment low suction controller, box of 10
0330035	Suction clear tubing, 1m (approx 20cm per system)
3210011	NeoPEEP circuit with variable PEEP, single patient use (without mask)
3210070	Single use silicone face mask, size 00
3210071	Single use silicone face mask, size 01
0120140	Silicone adapter for connecting NeoPEEP circuit to Tom Thumb
0120103	Paediatric T-adapter, 15mm I.D. to 15mm O.D. with 15mm I.D. port <small>Note: only required if connecting an oxygen analyser into the patient circuit</small>
0332000	Annual service kit for model 2003 blender
0332001	Overhaul kit for model 2003 blender – includes annual service kit
Operator and Service Manuals	
Part no.	Description
0391045	Resuscitation cabinet operator's manual
0391050	Resuscitation cabinet installation manual
0390007	Tom Thumb TT490-15 operating instruction card
0390006	Tom Thumb TT480 operating instruction card
0390022	TT480 service manual
0390023	TT490-15 service manual
0390021	Tom Thumb valves service manual (common to all Tom Thumb variants)
2122	BioMed Devices Air Oxygen blender operator's manual

## SERVICING

It is recommended that the resuscitation cabinets, its internal components and the radiant warmer are serviced annually. Service contracts are available; please contact Viamed for further information.

Service training courses are available from Viamed. Service training courses for the air/oxygen blender are available from Inspiration Healthcare.

Service the individual items of equipment using their respective Service or Operator's manuals, copies of which are available from Viamed.

For versions of the Tom Thumb that have an integral flowmeter and hose, Viamed recommends that the hose is inspected for wear or damage every 3 months, and replaced every 4 years regardless of condition of the hose.

If the Viamed Infant Resuscitation Cabinet has been supplied with a Therapy Equipment Low Suction Controller, the manufacturer recommends that the vacuum hose be replaced at least every 4 years, regardless of the condition of the hose.

If the Viamed Infant Resuscitation Cabinet has been supplied with an Oxylitre Low Suction Controller, the manufacturer recommends that the vacuum hose be replaced at least every 5 years, regardless of the condition of the hose.

If the Viamed Infant Resuscitation Cabinet has been supplied with an air/oxygen blender, the manufacturer recommends that the air and oxygen hoses be replaced at least every 4 years, regardless of the condition of the hose. Additionally, the air/oxygen blender should be subject to a complete overhaul every 4 years, please contact Viamed or Inspiration Healthcare for further details on how to achieve this.

The following table details the parts that will be required for service, along with the frequency with which they should be replaced. Which parts are required will depend upon the specific configuration of the resuscitation cabinet in question.

<b>Routinely replaced parts</b>		
<b>Part no.</b>	<b>Description</b>	<b>Replacement frequency in years</b>
0330213	O-ring, thick. Each Tom Thumb uses 1, replaced during service	1
0330214	O-ring, thin. Each Tom Thumb uses 6, replaced during service	1
3833-007	VacSax tapered suction connectors (pack/10)	1
0310104	Filter for Oxylitre low suction controller, single unit	1
0310109	Filter for Therapy Equipment low suction controller, single unit	1
0330035	Suction clear tubing, 1m (approx 20cm per system)	1
0330218	Oxygen hose assembly for Tom Thumb, 3m. Note: only for TT490-15	4
0330040	Suction hose assembly, 3m (Oxylitre)	5
0330033	Suction hose assembly, 3m (Therapy Equipment)	4
0320216	NIST hose with MKIV oxygen probe, 3m, for air/oxygen blender if installed	4
0320217	NIST hose with MKIV air probe, 3m, for air/oxygen blender if installed	4
0332000	Annual service kit for model 2003 blender	1
0332001	Overhaul kit for model 2003 blender – includes annual service kit	4

		Document ref: 0391055
		Version: 1.3
<b>Viamed Infant Resuscitation Cabinet Service Sheet</b>		
Hospital / Organisation:		
Location (department, room number):		
Engineer (print):		
Service Date:		
Action	Notes	OK
<b>Cabinet Body</b> – Serial number:		
Check for overall signs of damage	<input type="checkbox"/>	
Comments:		
Check door catch – swivel catch on door	<input type="checkbox"/>	
Comments:		
Check door catch – retaining clasp on cabinet	<input type="checkbox"/>	
Comments:		
Check integrity of welds on the door hinges	<input type="checkbox"/>	
Comments:		
Check hinges for damage or deformation	<input type="checkbox"/>	
Comments:		
Check the restraining bars for damage or deformation	<input type="checkbox"/>	
Comments:		
Check the cabinet is securely fixed to wall	<input type="checkbox"/>	
Comments:		
Check medirails and tighten any loose rails	<input type="checkbox"/>	
Comments:		
Check grommets on the hose inlets are in place and intact	<input type="checkbox"/>	
Comments:		
Check the bumper stops that prevent door banging during closing	<input type="checkbox"/>	
Comments:		
Check 'caution' sticker is in place on the platform	<input type="checkbox"/>	
Comments:		
Check bed for damage i.e. cracks, chips	<input type="checkbox"/>	
Comments:		
Check all the bed mounting screws are in place	<input type="checkbox"/>	
Comments:		
Check mattress for damage i.e. tears, split seams	<input type="checkbox"/>	
Comments:		
<b>Tom Thumb</b> – Serial Number:		
Follow service procedure as detailed in the Tom Thumb service manual	<input type="checkbox"/>	
Comments:		
Check oxygen hose, if present, for wear, and replace if necessary (at least every 4 years)	<input type="checkbox"/>	
Hose replaced? Yes <input type="checkbox"/> No <input type="checkbox"/>		
Date of last hose replacement:		
Comments:		

Action	Notes	OK
<b>Low Suction Controller</b> – Serial number:		
Follow service procedure as detailed in suction controller operator's or service manual		<input type="checkbox"/>
Comments:		<input type="checkbox"/>
Change filter assembly		<input type="checkbox"/>
Comments:		<input type="checkbox"/>
Check vacuum hose for wear, and replace if necessary (at least every 4 years)		<input type="checkbox"/>
Hose replaced? Yes <input type="checkbox"/> No <input type="checkbox"/>		<input type="checkbox"/>
Date of last hose replacement:		<input type="checkbox"/>
Comments:		<input type="checkbox"/>
Check rail clamp is securely attached to suction controller		<input type="checkbox"/>
Comments:		<input type="checkbox"/>
Replace suction tubing between suction controller and receiving liner		<input type="checkbox"/>
Comments:		<input type="checkbox"/>
Replace tapered connector on suction tubing		<input type="checkbox"/>
Comments:		<input type="checkbox"/>
Check receiving canister for damage, replace if necessary		<input type="checkbox"/>
Comments:		<input type="checkbox"/>
<b>Radiant Warmer</b> – Serial number:		
Note: if the unit is hard wired, a PAT test may not be applicable: consult your organisation's regulations.		
Is unit hard-wired? Yes <input type="checkbox"/> No <input type="checkbox"/>		<input type="checkbox"/>
PAT test performed? Yes <input type="checkbox"/> No <input type="checkbox"/> Pass <input type="checkbox"/> Fail <input type="checkbox"/>		<input type="checkbox"/>
Comments:		<input type="checkbox"/>
Follow service procedure as detailed in radiant warmer service manual		<input type="checkbox"/>
Comments:		<input type="checkbox"/>
<b>APGAR Timer</b> – Serial number:		
Change batteries		<input type="checkbox"/>
Comments:		<input type="checkbox"/>
Follow service procedure as detailed in Apgar timer operator's manual		<input type="checkbox"/>
Comments:		<input type="checkbox"/>
<b>General</b>		
Check storage bins present		<input type="checkbox"/>
Comments:		<input type="checkbox"/>
Tidy any trailing hoses, using tie-wraps if necessary		<input type="checkbox"/>
Comments:		<input type="checkbox"/>
Check Operator's Manuals are in place		<input type="checkbox"/>
Comments:		<input type="checkbox"/>
Clean cabinet with isopropyl alcohol		<input type="checkbox"/>
Comments:		<input type="checkbox"/>
<b>Additional notes or comments:</b>		

			Version: 1.0
<b>Digital Apgar Timer Service Sheet</b>			
Hospital / Organisation:			
Location (department, room no.):			
Engineer (print):			
Service Date:			
Action	Further action required		OK
Serial number:			
Batteries replaced	<input type="checkbox"/>	<input type="checkbox"/>	
Comments:			
Check for overall signs of damage	<input type="checkbox"/>	<input type="checkbox"/>	
Comments:			
Check START button enters <i>Counting Mode</i>	<input type="checkbox"/>	<input type="checkbox"/>	
Comments:			
Check FREEZE button enters <i>Freeze Mode</i>	<input type="checkbox"/>	<input type="checkbox"/>	
Comments:			
Check FREEZE button resumes <i>Counting Mode</i>	<input type="checkbox"/>	<input type="checkbox"/>	
Comments:			
Re-enter <i>Freeze Mode</i> , check START button resumes <i>Counting Mode</i>	<input type="checkbox"/>	<input type="checkbox"/>	
Comments:			
Check RESET button enters <i>Stand-by Mode</i>	<input type="checkbox"/>	<input type="checkbox"/>	
Comments:			
Start a stopwatch whilst simultaneously pressing START			
Check single beep audible indication at 1 minute	<input type="checkbox"/>	<input type="checkbox"/>	
Comments:			
Check two beeps audible indication at 5 minutes	<input type="checkbox"/>	<input type="checkbox"/>	
Comments:			
Check three beeps audible indication at 10 minutes	<input type="checkbox"/>	<input type="checkbox"/>	
Comments:			
Check count is synchronized with the stopwatch at 10 minutes, +/- 1 sec	<input type="checkbox"/>	<input type="checkbox"/>	
Comments:			
Check mounting clamp and bracket are securely fastened	<input type="checkbox"/>	<input type="checkbox"/>	
Comments:			
Clean unit as detailed in the Operator's Manual	<input type="checkbox"/>	<input type="checkbox"/>	
Comments:			

## CLEANING

Cleaning guidelines in this manual complement those given in the Medical Devices Agency document "Sterilization, disinfection and cleaning of medical equipment: guidance on decontamination from the Microbiology Advisory Committee to Department of Health Medical Devices Agency". This document is more commonly known as the 'MAC Manual', and is referred to as such in this document for convenience.

With reference to the MAC Manual, "Introduction to Part 1 - Table 1: Classification of infection risk associated with the decontamination of medical devices"; many of the items of equipment comprising the Viamed Wall-Mounted Resuscitation Cabinet system are classified as low risk, due to being "In contact with healthy skin, or not in contact with the patient".

The recommended decontamination method for low risk items is to clean in accordance with the guidelines in the MAC Manual "Part 2; Cleaning (manual) - non-immersion".

### **Clean as per MAC Manual "Part 2; Cleaning (manual) – non-immersion"**

- Radiant warmer
- Cabinet body
- Mounting rails
- Storage bins
- Integral bed
- Mattress
- Tom Thumb infant resuscitator
- Low suction controller
- VacSax suction receiving canister (can also be steam sterilized if required)
- Apgar timer

### **Cleaning Guidelines**

The items listed above are low risk and not intended to be disinfected or sterilized, but where an alcohol wipe is used, this may have a disinfecting effect. After manual cleaning, an alcohol wipe containing 70% ethanol can be used over all external surfaces, observe safety precautions detailed later in this document.

It is recommended that a neutral detergent solution be used in accordance with the manufacturer's guidelines.

Sodium hypochlorite solutions are not recommended due to their potential corrosive effect on some metals.

The equipment within the cabinet system is mounted on standard mounting rails. To remove the equipment, loosen the rail clamps by rotating the locking nuts until the equipment can be lifted free of the rail.

It is recommended that the bed be cleaned in-situ, should it be necessary, the bed can be removed by removing the 6 screws holding it in place. Care should be taken when re-attaching the bed not to over-tighten the screws.

**NOTE: the following instructions detail the method of manual cleaning (non-immersion) as directed by the MAC Manual. Should these instructions contradict guidelines in place within the customer's own organization, or those of the manufacturer of the cleaning detergent, please contact your Sterile Services Department or other responsible body within your organization for advice.**

### **Equipment Required**

- A warm water/detergent solution at correct dilution.
- A clean, disposable, absorbent, non-shedding cloth for application of detergent solution.
- A clean, disposable, absorbent, non-shedding cloth or mechanic drying facility (e.g. drying cabinet or industrial hot air dryer).
- An appropriate chemical neutraliser, first aid kit and eyewash bottle, in case of splashing with detergent.

### **Procedure**

- If the item is electrical, ensure that it is disconnected from the mains supply before commencing the cleaning procedure.
- Wearing protective clothing, immerse the cleaning cloth in the detergent solution and wring thoroughly.
- Commencing with the upper surface of the item, wipe thoroughly ensuring that detergent solution does not enter electrical components.
- Periodically rinse the cloth in clean water and repeat the above steps.
- Surfaces should be carefully hand-dried using a cloth or industrial hot air dryer or placed into a drying cabinet.

Note: Non-immersion, manual cleaning is not a disinfection process, but where an alcohol wipe is used to dry surfaces, this may have a disinfecting effect.

- Safely dispose of cleaning materials and alcohol wipes, if used.

## **Monitoring and Control**

Owing to the lack of control methods available to the user to test the efficiency of non-immersion cleaning, the user should be aware of the factors that may alter the efficiency of the method:

- staff training
- physical application
- nature of soil
- accessibility of cleaner to item/part of equipment
- detergent concentration

## **Safety Precautions**

- Always wear protective waterproof clothing, robust gloves and eye protection if splashing is likely to occur.
- After removing protective clothing on completion of task, thoroughly wash and dry hands.
- Avoid splashing.
- Precautions should be taken when using alcohol, as it is flammable.
- The 'pooling' of alcohol on equipment should be avoided and alcohol evaporation ensured, if necessary by forced air drying. Care should also be taken to ensure that alcohol does not enter the item e.g. via ventilation slots.

## **Items that can be Steam Sterilized at 121° C as per HTM 2010, Part 3**

- VacSax suction receiving canister
- VacSax suction taper
- Therapy Equipment reusable suction receiving jar

## **Items Intended for Single Patient Use Only**

- Breathing circuit
- Suction tubing
- VacSax suction receiving liners
- Suction catheters



### **WARNING**

Do not attempt to reprocess items intended for single patient use.

## **WARRANTY**

All parts of the Viamed wall mounted infant resuscitation cabinet system are guaranteed for a period of 12 months from the date of purchase.

The best materials and workmanship have been employed throughout every stage of manufacture and every part is thoroughly tested before dispatch.

This warranty covers defects in material and manufacture but excludes damage caused by accident, misuse or neglect.

Should a component develop a defect within the warranty period, it will be repaired or replaced at Viamed's discretion.

In the event of warranty claims or queries regarding this product when purchased outside the UK, please contact your local distributor.

## **COMPANY DETAILS**

All products are CE marked to the requirements of MDD 93/42 EEC and are supplied in accordance with our quality system accreditations:  
BS EN ISO 9001/2000 and ISO 13485/2003.

*Contact details:*

Viamed Ltd.  
15 Station Road  
Cross Hills  
Keighley  
West Yorkshire  
BD20 7DT  
United Kingdom

Tel: +44 (0)1535 634542  
Fax: +44 (0)1535 635582

Email: [info@viamed.co.uk](mailto:info@viamed.co.uk)  
Web: [www.viamed.co.uk](http://www.viamed.co.uk)