

# Radiant Warmer Revision Questions (Edition 1.0)

Please review the Introduction (DOCID

182866 ) and Technical (DOCID 124724) training notes in the document index and answer the following questions.

Email \*

sophie.lines@viamed.co.uk

What is the default power output at **Level 3** on a Ceratherm radiant warmer? \*

- 20%
- 50%
- 75%
- 98%

What type of heating element does the Ceratherm 600-3 use? \*

- Halogen lamp
- Quartz tube
- Ceramic element
- Electric coil

What is the part number for the Ceratherm radiant warmer with mobile floor stand? \*

- 0310302
- 0310303
- 0310304
- 0310314

How long after the last key-press will the safety alarm be triggered? \*

- 5 minutes
- 10 minutes
- 15 minutes
- 20 minutes

What is the wavelength of the infrared radiation emitted by the Ceratherm 600-3? \*

- 1  $\mu\text{m}$
- 2  $\mu\text{m}$
- 3  $\mu\text{m}$
- 5  $\mu\text{m}$

What happens if the safety alarm is not cancelled within 8 seconds? \*

- The warmer turns off
- The power increases to 98% (Level 4)
- The alarm resets itself
- The power drops to 20% (Level 1)

In which of the following settings is the Ceratherm 600-3 typically used? \*

- Dental surgeries
- Delivery suites
- GP offices
- Opticians

What is the minimum clearance required from the underside of the warmer head unit to the patient surface? \*

- 60 cm
- 70 cm
- 80 cm
- 90 cm

What is the standard brightness of the low LED lamp setting on the Ceratherm 600-3? \*

- 10%
- 20%
- 50%
- 75%

What type of temperature control is used in the Ceratherm radiant warmers? \*

- Fully automatic with servo feedback
- Manual control with selectable power levels
- Smartphone app-controlled
- Thermostat-regulated only

This content is neither created nor endorsed by Google.

Google Forms