

Flowsensors Revision Questions (Edition 1.0)

Please review the Introduction (DOCID 190088) and Technical (DOCID 181873) training notes in the document index and answer the following questions.

Please note: Some questions have multiple answers.

Email *

viamed.emily.hanson@gmail.com

Which of the following flowsensors are compatible with Dräger anaesthesia and ventilation equipment? *

☒ 4310007

☒ 4310008

☐ 4310009

What is the maximum number of autoclave cycles that the SpiroTrue A(PC) can withstand at 134°C? *

☐ 50

☐ 75

☒ 100

☐ 150

Flowsensor H utilizes which method for determining flow? *

- ☐ Hot wire anemometric
- ☐ Ultrasonic
- ☒ Pressure differential
- ☐ Optical

How many flowsensors are included in a box of Flowsensor H? *

- ☐ 4
- ☐ 5
- ☒ 6
- ☐ 10

What is the primary function of flowsensors in ventilators? *

- ☐ To filter air
- ☐ To humidify gases
- ☒ To measure gas flow and support tidal volume calculations
- ☐ To sterilize breathing circuits

When placed at the patient's airway, the flowsensor is positioned: *

- ☐ Distally
- ☐ Laterally
- ☒ Proximally
- ☐ Medially

Which flowsensors are described as single-use and not suitable for cleaning or autoclaving? *

- ☒ Flowsensor A
- ☐ Flowsensor A(PC)
- ☒ Flowsensor H
- ☐ All of the Above

What term describes the method used by hot wire flowsensors to measure gas flow? *

- ☐ Thermometric
- ☒ Anemometric
- ☐ Barometric
- ☐ Pneumatic

A customer references OEM part number 53120. Which flowsensor should be suggested? *

- ☐ 4310007
- ☐ 4310008
- ☒ 4310009

What is the recommended cleaning method for Flowsensor A(PC)? *

- ☐ Alcohol wipe only
- ☒ Immersion in water
- ☐ Autoclave at 134°C, not in units with cleaning functions
- ☐ Dishwasher cleaning

This content is neither created nor endorsed by Google.

Google Forms